


EQUIPMENT:**AIR INCUBATOR**

Business Unit	
Laboratory	
Req. #	
SPECIFICATION	REQUIREMENTS
Construction	Chemical and disinfectant resistant, graded steel interior
Dimensions	Not exceeding 1800(w) x 1800 (h) x 780 (d)mm
Footprint	Floor standing
Capacity	600 - 750L
Doors - inner	Glass doors essential
Doors - outer	Double doors (preferably stainless steel)
Temperature range	5 - 70C (requirement for testing 35-40C)
Temperature control accuracy range	+0.3C at 37C and 0.2C accuracy
Shelves	4 detachable
Thermometer	Digital temperature controller
Temperature display	Single or twin digital
Temperature monitoring	Ethernet port for temperature logging onto LIS
Safety thermostat	Yes
Convection method	Natural as opposed to fan forced
Casters for moving	4x lockable
Warranty period	minimum 1 year
Delivery period	6 weeks
Aftersales service	Local
Operator's manual	English
Electronic voltage SA compliant	Yes
Safety compliant	Yes
Laboratory Manager	
Signature	
Date	
Quality Assurance	N Godfrey
Signature	
Date	01 July 2020