



Equipment:	Centrifuge: NON refrigerated
Business Unit	OR Tambo & Chris Hani
Requesting laboratory	Zitulele
Requisition #	1410762
Date	02-Jul-20

Specification item	Specification detail	compliance	nonCompliance
Drive	Brushless induction		
Control	Microprocessor with LCD display		
Speed	2000 rpm as average range provided		
RPM/RCF mode	yes		
Programmability	Pre selection of speed and time		
Balance detection	Electronic with alarm and cut off		
Dimensions	Bench top 600mm(s)x550mm(w) 500mm(h)		
Run Time	Set and continuous timer for 1 hour		
Decibels	≤70 dBA		
non refridgerated	yes		
adjustable Timer	<u>0-30minutes</u>		
Rotar	Swing out 4 bucket rotars with test tube holders		
Tube size	5ml, 7ml		
No of tubes	48 - 78		
Adapters	4 with capacity of 20 or more 15ml conical tubes		
lids	click seal biocontaminant lids		
Buckets have safety covers	screw on covers		
electrical	220V,50- 60Hz		
Ergonomic Design	Bench top		
Delivery time within 6 weeks	yes		
In use in NHLS	yes		
Lab manager	N. Maxhama		
Signature			
Date	23-Jul-20		
QAC	Nokonwabisa Dilinga		
Signature			
Date	23-Jul-20		