



SPECIFICATIONS FOR CAPEX ASSETS 2020/2021

From: Vanessa Pearce
Dept: Bacteriology Port Elizabeth
Cost Code: 21006
Date: 14 July 2020
Equipment Required: MOVABLE LARGE VOLUME FLOOR STANDING REACH IN CO2 INCUBATOR

<u>TECHNICAL PARAMETERS</u>	<u>SPECIFICATION</u>	<u>Comply yes</u>	<u>Comply no</u>
CO2 incubator	Adjustable level CO2 incubator		
SIZE	Floor standing > >750l capacity		
Adjustable Feet	For stability		
CASTERS	4 x lockable wheels/casters		
INTERIOR	Stainless steel		
EXTERIOR	Rolled steel powder coated or Stainless		
DOOR	Glass panel exterior door Front opening Reach in type.		
SHELVES	4-6 stainless steel removable shelves		
RELATIVE HUMIDITY	Water jacketed or reservoir		
TEMPERATURE	Adjustable Temperature controller Range 32-40° C		
ALARM	Visual /Audible alarm for CO2 and water levels		
POWER	Std 220V SABS approved for RSA		
WARRANTY	Min 1year		
TRAINING	Onsite		
OPERATORS MANUAL			
SUPPLY AND INSTALLATION			
DELIVERY	Within 6weeks of order		

Checked by QAD:..... *Pushes*
13/07/2020

W.Pearce