



SPECIFICATIONS FOR CAPEX EQUIPMENT 2020/2021.

From: Vanessa Pearce
Dept: Bacteriology Port Elizabeth
Cost Code: 21006
Date: 14/07/2020

Equipment Required: **PERISTALTIC LAB BLENDER (2 units)**

2 x Small Lab blender for use in Biohazard safety cabinet
Stomaching type action
Double paddle blending system
Removable paddles
small Volume bag system: 5 to <100ml
Stainless steel construction
Adjustable Timer
Variable strength setting
Easy access for cleaning of the blending area
Std power supply 220V
Maximum Dimensions:
✚ 250mm x 400mm
✚ Height less than 200mm

APPROVED BY QAD: *Amhas*

DATE: 15/07/2020

A large, stylized handwritten signature is located in the bottom right corner of the page. The signature is written in black ink and appears to be a cursive or semi-cursive script. A small number "1" is written at the bottom right of the signature.