

## SPECIFICATIONS FOR AUTOCLAVE

Specifications	Requirements	Completed	
		Y	N
Sliding door	Automated, single door		
Capacity	600 liters – rectangular jacket		
Boiler	On board with necessary water softener		
Warranty	Chamber and Jacket must have a 5-year pressure vessel integrity		
Stainless steel Grade	Higher than 316 grade – corrosive and pit resistant		
Blowdown	Automatic with manual blow down option		
Power failure protection	Autoclave cycle continues despite temporary loss of power. UPS so that in the event of power failure autoclave cycle can continue when generator kicks in.		
On board Printer	Print cycle details for real time record keeping		
Internal loading carriage	316L stainless steel		
External floor trolley	316L stainless steel		
Touch screen	Yes		
Able to store different cycles	Both pressurized and non-pressurized		
Fault tracing	Display of error message for troubleshooting		
Operators manual	English		
Onsite training	Included		
Five year service agreement	Quoted		
Delivery, installation and Commissioning	Included		

*GA* RQAM 11/06/2021