

Equipment		Double Door (Glass) Laboratory Fridge	
Business Unit		ORTambo and Chris Hani	
Laboratory		NHLS Queenstown	
Req. #			

Specification:	Detail
Floor Standing	Yes
Doors	2 Glass Doors
	Sliding
	Magnetic Door Gasket.
Shelving	Minimum 4 shelves on each side-adjustable.
Dimensions (approx) L x W x H (mm)	1250 x 1060 x 1850 (approximately).
Capacity	± 1000 L
Temperature Range	Able to maintain +2°C to +8°C
Temperature adjustment	Available
Defrost	Auto-defrost or frost free.
Operators manual	Yes
Warranty Period	Minimum 1 year
Delivery Period	2-4 weeks
Desirable (not essential):	
Self-closing doors	
External digital thermometer	
Lab manager	P Prusent
Signature	<i>P Prusent</i>
Date	12/07/2021
QA	Nokonwabisa Dilinga
Signature	<i>MLD</i>
Date	13/07/2021