

GENERAL SPECIFICATIONS	REQUIREMENTS	TICK REQUIREMENTS
Tissue embedder and cryoconsole	Yes	
Dimensions	560mm(w)X640mm(d)x400mm(h)	
LCD display	yes	
Programmable automatic start and shut down	yes	
Paraffin chamber	min 4L capacity and 50-70C	
Cassette/Molds tray	1.4 L	
Paraffin tank temperature	ambient to 90C	
Cassette/mold tray temperature	ambient to 90C	
Working table temperature	ambient to 90C	
Refrigerating and positioning table/cryoconsole	ambient to -5C	
Heated forceps	Yes	
Forceps holder temperature	65C constant	
Working temperature	1C increments	
Power supply	220V/50Hz	
Maximum power draw	1 000Va	
Net weight	<35Kg	
Operators manual	Yes	
Delivery time within 6 weeks	Yes	
Lab Manager Name:	John Edwards	
Lab Manager Signature:		
Date:	02/06/2022	
QAC Name:	Vicky Kleinhans	
QAC Signature:	<i>[Signature]</i>	
Date:	3/06/2022	

RMA 1726661