



RFQ ..... 0009778 - NON REFRIDGERATED CENTRIFUGE			ST LUCYS	Dr. Malizo
1				
Centrifuge with large visible display, Bio-safe rotor & rotor easy to clean, Anti sample overflow that will prevent accidental leakage during the process & prevent contamination.			Comply	Not Comply
2				
Rotor that can be fully autoclaved for unlimited time, Quick access to samples with snap lid, Active imbalance identification & cut off, Max speed 4000rpm with speed accuracy			Comply	Not Comply
3				
Rotor capacity btwn 2 & 15 ml of total of 64 serum / plasma sample tubes, run time 30 sec-99min/continuous, Noise level <60 DbA at 2130xg.			Comply	Not Comply
4				
Manufactured according to international Std/Regulations, Over speed & Temperature detection safety measures.			Comply	Not Comply
5				
Automatic internal diagnosis safety devices Installation records. Calibration certificate by SANAS accredited supplier.			Comply	Not Comply

NHLS ALFRED NZO & JOE GOUDI  
MANAGEMENT & ADMIN OFFICE

2024-09-05

Name: Buyiswa Ndebe Sign:   
Approved ☒ Not Approved ☐

DR MALIZO MPELE LABORATORY
2024-09-05
TEL: 017-0420413



P.P