

RFQ - NON REFRIDGERATED CENTRIFUGE		Glen Grey	
1			
Centrifuge with large visible display, Bio-safe rotor & rotor easy to clean, Anti sample overflow that will prevent accidental leakage during the process & prevent contamination.		Comply	Not Comply
2			
Rotor that can be fully autoclaved for unlimited time, Quick acces to samples with snap lid, Active imbalance identification & cut off, Max speed 4000rpm with speed accuracy		Comply	Not Comply
3			
Rotor capacity btwn 2 & 15 ml of total of 64 serum / plasma sample tubes, run time 30 sec-99min/continuous, Noise level <60 DbA at 2130xg.		Comply	Not Comply
4			
Manufactured according to international Std/Regulations, Over speed & Teemperature detection safety measures.		Comply	Not Comply
5			
Automatic internal diagnosis safety devices Installation records. Calibration certificate by SANAS accredited supplier.		Comply	Not Comply

W. Apple

24/10/24