



**NATIONAL HEALTH
LABORATORY SERVICE**

Supply Chain Management

1 Modderfontein Road, Sandringham, 2031

Tel: +27 (0)11 555 0398 Fax: +27 (0)11 386 6586

Reference: RFB108/23/24

Ref: RFB108/23/24

Enquiry: Phillip Serage

Tel: (011 555 0595

Date: 20 May 2025

ERRATUM ON TENDER DOCUMENT

RFB108/23/24: SUPPLY, DELIVER, INSTALL AND COMMISSION ONE (1) BENCH TOP INDUCTIVELY COUPLED PLASMA WITH MASS SPECTROMETER (ICP-MS) FOR THE CHROMATOGRAPHY SECTION AT THE DEPARTMENT OF CHEMICAL PATHOLOGY LABORATORY IN IALCH INCLUDING REAGENTS, CONSUMABLES, SERVICE AND MAINTENANCE FOR A PERIOD OF FIVE (5) YEARS.

NHLS went out on a tender for the **Supply, Deliver, Install and Commission one (1) Bench Top Inductively Coupled Plasma with Mass Spectrometer (ICP-MS) for the Chromatography Section at the Department of Chemical Pathology Laboratory in IALCH including reagents, consumables, service and maintenance for a period of five (5) years.**

Published date	15 May 2025
Compulsory briefing session date	23 May 2025
Tender Closing date	13 June 2025

REASON FOR ERRATUM

We hereby request to amend Mandatory requirement No. 5 that read as follows:

The ion source must consist of a water cooled, digital solid state radio frequency (RF) generator at the frequency of 32-40 Megahertz (MHz) operating at powers of up to 1600W for a stable plasma to be maintained.

Substantiation: Bidders are requested to provide a manual/ brochure with a unique reference to locate substantiating evidence. Indicators need to be added for ease of access to the requirement.

Mandatory requirement must now read as follows:

The ion source must consist of a water cooled, digital solid state radio frequency (RF) generator at the frequency of 27-40 Megahertz (MHz) at powers of up to 1600W for a stable plasma to be maintained.

Substantiation: Bidders are requested to provide a manual/ brochure with a unique reference to locate substantiating evidence. Indicators need to be added for ease of access to the requirement.

Requested by:

Phillip Serage
Tender Administrator

Supported by

Tébogo Molefe
Controller: Tenders

Approved by:

Rofhiwa Mammburu

Signed by: Rofhiwa Mammburu,
Rofhiwa.Mammburu@nhls.ac.za
20/05/2025 14:52:58(UTC+02:00)

Rofhiwa Mammburu
Head: Supply Chain Management

