



EQUIPMENT
Centrifuge (Micro high speed)

Business Unit	NMAL	Compliant	Non-compliant
Laboratory	Microbiology		
Req. #			
Specification	Requirement		
Foot print	Benchtop		
Drive	Brushless induction		
Control Panel	LCD display touch pad		
Speed	200 – 14 800 rpm		
RPM/RCF mode	RPM & RCF		
Run time	10 seconds to 99 minutes		
Acceleration	Variable acceleration		
Deceleration	Variable deceleration		
Programmability	Pre selection of speed and time		
Footprint	Benchtop		
Decibels	≤70 dBA		
Refrigerated	No		
Ambient temperature	10 - 35°C		
Internal temperature	Not applicable		
Rotor	Fixed angle		
Tube types & size	1.5 and 2.0 mL tubes		
Number of tubes	Minimum 24 tubes		
Design	Window in lid for speed checking		
	Emergency lid lock release		
Electrical	Compliant to South Africa		
Delivery time	6 weeks		
Operator's manual	English		
Installation and commissioning	By Supplier		
Warranty period	Min 1 year		
After sales service	Annual service for 5 years		
SAHPRA Licence	From supplier		
ISO 13485	From manufacturer		
Laboratory Manager			
Signature			
Date	19/6/2025		
Authorized by			
Signature			
Date	01/07/2025		