

RFQ: 0121308 SUPPLY, DELIVER AND INSTALL A WATER PURIFICATION SYSTEM AT NHLS
PORT ELIZABETH MEDIA SECTION.

Addendum 1. SPECIFICATION

The bid response should specify that the system additional meets the below criteria

Pure (Type II) or EDI
14 – 18 MΩ.cm suitable for clinical analyser feed
pH neutral
<1 cfu/mL