

## **INVITATION FOR BID-A01**

# YOU ARE HEREBY INVITED TO SUBMIT PROPOSAL FOR THE REQUIREMENTS OF NATIONAL HEALTH LABORATORY SERVICE (NHLS)

| BID NUMBER:   | RFB153/25/26   |                                       |  |  |
|---|--|---------------------------------------|--|--|
|   |  |                                       |  |  |
| CLOSING DATE:   | 03 FEBRUARY 2026   |                                       |  |  |
|   |  |                                       |  |  |
| CLOSING TIME:   | 11:00 AM   |                                       |  |  |
|   | DATE OF FERRUARY SOC   |                                       |  |  |
| PUBLIC TENDER   | DATE: 03 FEBRUARY 2026   |                                       |  |  |
| OPENING:  | TIME: 11:30 AM   |                                       |  |  |
|   | VENUE: MAIN CONFERENCE BOARDROOM   | Л                                     |  |  |
|   | NATIONAL HEALTH LABORATORY   | SERVICE                               |  |  |
|   | 1 MODDERFONTEIN ROAD   |                                       |  |  |
|   | SANDRINGHAM  |                                       |  |  |
|   |  |                                       |  |  |
| BID VALIDITY PERIOD:  | 180 days (commencing from the RFB Clos   | ing Date)                             |  |  |
|   |  |                                       |  |  |
| IMPORTANT:  | A COMPULSORY BRIEFING SESSION AND  | THE SITE VISIT WILL BE HELD:          |  |  |
|   | DATE: 16 JANUARY 2026  |                                       |  |  |
|   | TIME: 12:00 PM   |                                       |  |  |
|   | VENUE: 4TH FLOOR, ROOM N431 CLINICAL PATHOLOGY BUILDING SEFAKO MAKGATHO UNIVERSIT MOLOTLEGI STREET GARANKUA                        |                                       |  |  |
|   | SITE VISIT: 16 JANUARY 2026  |                                       |  |  |
|   | PLEASE NOTE THAT LATE COMING WILL NO   | OT BE ACCEPTED                        |  |  |
|   | All questions must be sent per e-mail to <a href="mailto:ntombozuko.mbewu@nhls.ac.za">ntombozuko.mbewu@nhls.ac.za</a> on or before |                                       |  |  |
|   | 21 January 2026.   |                                       |  |  |
| PLACEMENT OF TOTAL AUTOMATION SYSTEM FOR PRE-ANALYTICS, ANALYTIC AND POST-ANALYTICAL SYSTEMS FOR CHEMICAL PATHOLOGY/ STAT LABORATORY FOR NATIONAL HEALTH LABORATORY SERVICE (NHLS) DGM INCLUDING REAGENTS, CONSUMABLES & SERVICE CONTRACT FOR A PERIOD OF FIVE (5) YEARS. |  |                                       |  |  |
|   |  |                                       |  |  |
| BID DOCUMENTS MUST B  | E MARKED WITH THE FOLLOWING: OR  | DEPOSITED IN THE BID BOX SITUATED AT: |  |  |
| NHLS PROCUREMENT TEN  | DER OFFICE   |                                       |  |  |

Confidential

RFB Number: 153/25/26: Placement of Total Automation System for Pre-Analytics, Analytic and Post-Analytical Systems for Chemical Pathology/ STAT Laboratory for National Health Laboratory Service (NHLS) DGM including reagents, consumables & service contract for a period of five (5) years.



| RFB153/25/26 Bidders Name:                       | NHLS MAIN RECEPTION  1 Modderfontein Road, Sandringham, Johannesburg. |
|--|---|
| RFB: Enclosed-Regret (delete N/A)  Closing Date: |   |
|  |   |

Bidders should ensure that Bids are delivered in time to the correct address. If the bid is late, it shall not be accepted for consideration.

ALL BIDS MUST BE SUBMITTED ON THE OFFICIAL FORMS – (Please note that no changes on the content of this document is allowed)

Bidders should ensure that Bids are delivered in time to the correct address. If the bid is late, it shall not be accepted for consideration.

ALL BIDS MUST BE SUBMITTED ON THE OFFICIAL FORMS – (Please note that no changes on the content of this document is allowed).

THIS TENDER IS SUBJECT TO THE GENERAL CONDITIONS OF THE TENDER, THE GENERAL CONDITIONS OF CONTRACT (GCC) AND, IF APPLICABLE, ANY OTHER SPECIAL CONDITIONS OF CONTRACT.

THE FOLLOWING PARTICULARS MUST BE FURNISHED (FAILURE TO DO SO SHALL RESULT IN YOUR BID BEING DISQUALIFIED).

| SUPPLIER INFORMATION       |              |      |     |         |
|----------------------------|--------------|------|-----|---------|
| NAME OF BIDDER             |              |      |     |         |
| POSTAL ADDRESS             |              |      |     |         |
| STREET ADDRESS             |              |      |     |         |
| TELEPHONE NUMBER           | CODE:        | NUMB | ER: |         |
| CELLPHONE NUMBER           |              |      |     |         |
| FACSIMILE NUMBER           | CODE NUMBER: |      |     |         |
| E-MAIL ADDRESS             | ·            |      |     |         |
| VAT REGISTRATION<br>NUMBER |              |      |     |         |
| NOMBER                     | TCS PIN:     |      | OR  | CSD No: |

Confidential Page 2 of 88



| SUPPLIER INFORMATION   |                               |                                     |         |            |         |                                  |  |
|--|-------------------------------|-------------------------------------|---------|------------|---------|----------------------------------|--|
|  |                               |                                     |         |            |         |                                  |  |
| B-BBEE STATUS LEVEL VERIFICATION CERTIFICATE                     | Yes  No [TICK APPLICABLE BOX] | B-BBEE STATUS LEVEL SWORN AFFIDAVIT |         |            |         | ☐ Yes ☐ No [TICK APPLICABLE BOX] |  |
| [A B-BBEE STATUS LEVEL IN ORDER TO CLAIM POIN                    | VERIFICATION CERTIFIC         | •                                   |         | (FOR EMEs& | QSEs) N | MUST BE SUBMITTED                |  |
| SIGNATURE OF BIDDER  |                               |                                     |         |            | Date:   |                                  |  |
| CAPACITY UNDER WHICH proof of authority to sign directors, etc.) | •                             |                                     |         |            |         |                                  |  |
| TOTAL BID PRICE (ALL INCLUSIVE)                                  |                               |                                     |         |            |         |                                  |  |
| BIDDING PROCEDURE AND  | D TECHNICAL ENQUIRIE          | S MAY BE DIREC                      | TED TO: |            |         |                                  |  |
| DEPARTMENT/ PUBLIC ENTITY  |                               |                                     |         |            |         |                                  |  |
| CONTACT PERSON   |                               |                                     |         |            |         |                                  |  |
| TELEPHONE NUMBER   |                               |                                     |         |            |         |                                  |  |
| FACSIMILE NUMBER   |                               |                                     |         |            |         |                                  |  |
| E-MAIL ADDRESS   |                               |                                     |         |            |         |                                  |  |



## Contents

| 1.  | Confidential information disclosure notice                                   | 5  |
|-----|--|----|
| 2.  | Introduction   | 5  |
| 3.  | Definitions  | 5  |
| 4.  | Acronyms and abbreviations   | 9  |
| 5.  | General Rules and Instructions   | 9  |
| 6.  | Response format  | 13 |
| 7.  | Key personnel  | 14 |
| 8.  | Reasons for Disqualification   | 14 |
| 9.  | Bid Preparation  | 14 |
| 10. | Oral presentations and Briefing Sessions                                     | 14 |
| 11. | General Conditions of Bid and Conditions of Contract                         | 15 |
| 12. | Evaluation Criteria and Methodology  | 24 |
| ANN | NEXURE A: Technical Specification  | 29 |
| ANN | NEXURE B: Pricing Schedule   | 46 |
| ANN | NEXURE C: Bidder's Disclosure (SBD4)   | 51 |
| ANN | NEXURE D: Preferential Procurement Claim Form (SBD6.1)                       | 65 |
| swo | ORN AFFIDAVIT: B-BBEE QUALIFYING SMALL ENTERPRISE: GENERAL                   | 71 |
| swo | ORN AFFIDAVIT: B-BBEE QUALIFYING MICRO ENTERPRISE: GENERAL                   | 73 |
| ΔΝΝ | JEXURE F: Government Procurement: General Conditions of Contract – July 2011 | 75 |



## 1. Confidential information disclosure notice

- 1.1 This document may contain confidential information that is the property of the NHLS and the Client.
- 1.2 No part of the contents may be used, copied, disclosed or conveyed in whole or in part to any party in any manner whatsoever other than for preparing a proposal in response to this Bid, without prior written permission from NHLS and the Client.
- 1.3 All Copyright and Intellectual Property herein vests with NHLS and its Client.

#### 2. Introduction

2.1 Based on the Bids submitted and the outcome of the evaluation process according to the set evaluation criteria, NHLS intends to select a preferred bidder with the view of concluding a service level agreement (SLA) with such successful bidder. The Bid shall be evaluated in terms of the Preferential Procurement Policy Framework Act (PPPFA)

#### 2.2 Queries

2.2.1 Should it be necessary for a bidder to obtain clarity on any matter arising from or referred to in this RFB document, please refer queries, in writing, and to the contact person\_email address number listed below on or before **21 January 2026**. Under no circumstances may any other employee within NHLS be approached for any information. Any such action might result in a disqualification of a response submitted in competition to the RFB. NHLS reserves the right to place responses to such queries on the website.

|                           | Telephone | 011 386 6138                |
|---------------------------|-----------|-----------------------------|
| QUERIES: Ntombozuko Mbewu | E-mail    | ntombozuko.mbewu@nhls.ac.za |

#### 3. Definitions

- 3.1 National Health Laboratory Services [hereinafter referred to as NHLS] is a public health laboratory service with laboratories across South Africa. Its activities comprise diagnostic laboratory services, research, teaching and training, and production of sera for anti-snake venom, reagents and media.
- 3.2 NHLS was established in 2001 by an Act of Parliament to provide diagnostic pathology laboratory services to the National and Provincial Health Department.
- 3.3 "Acceptable Bid" means any bid, which, in all respects, complies with the specifications and conditions of the RFB as set out in this document.
- 3.4 **"B-BBEE"** means broad bases black economic empowerment as defined in section 1 of the Broad-Based Black Economic Empowerment Act.

Confidential Page 5 of 88



- 3.5 **"B-BBEE status level of contributor"** means the B-BBEE status received by a measured entity based on its overall performance using the relevant scorecard contained in the Codes of Good Practice on Black Economic Empowerment, issued in terms of section 9(1) of the Broad-Based Black Economic Empowerment Act.
- 3.6 **"Bid"** means a written offer in a prescribed or stipulated form in response to an invitation by an organ of state for the provision of services, works or goods through price quotations, advertised bidding processes or proposals.
- 3.7 **"Bidders"** means any enterprise, consortium or person, partnership, company, close corporation, firm or any other form of enterprise or person, legal or natural, which has been invited by NHLS to submit a bid in response to this bid invitation.
- 3.8 **"Broad-Based Black Economic Empowerment Act"** means the Broad-Based Black Economic Empowerment Act, 2003 (Act No. 53 of 2003).
- 3.9 "Client" means the goods or services requestor.
- 3.10 **"Comparative Price"** Means the price after the factors of a non-firm price and all unconditional discounts that can be utilized have been taken into consideration.
- 3.11 "Consortium" means several entities joining forces as an umbrella entity to gain a strategic collaborative advantage by combining their expertise, capital, efforts, skills and knowledge for the purpose of executing this tender.
- 3.12 "Contractor Agent" means any person mandated by a Prime Contractor or consortium/joint venture to do business for and on behalf of, or to represent in a business transaction, the Prime Contractor and thereby acquire rights for the Prime Contractor or consortium/joint venture against NHLS or an organ of state and incur obligations binding the Prime Contractor or consortium/joint venture in favour of NHLS or an organ of state.
- 3.13 "Disability" means, in respect of a person, a permanent impairment of a physical, intellectual, or sensory function, which results in restricted, or lack of, ability to perform an activity in the manner, or within the range, considered normal for a human being.

## 3.14 Designated group means -

- (a) Black designated groups;
- (b) Black people;
- (c) Women
- (d) People with disabilities; or
- (e) Small enterprises as defined section 1 of the National Small Enterprise Act, 1996 (Act No. 102 of 1996)

Confidential Page **6** of **88** 



- 3.15 "Designated sector" means a sector, sub-sector or industry or product designated by the Department of Trade and Industry.
- 3.16 "EME" means an Exempted Micro Enterprise in terms of a code of good practice on black economic empowerment issued in terms of section 9 (1) of the Broad-Based Black Economic Empowerment Act;
- 3.17 "Firm Price" means the price that is only subject to adjustments in accordance with the actual increase or decrease resulting from the change, imposition or abolition of customs or excise duty and any other duty, levy or tax which, in terms of a law or regulation is binding on the contractor and demonstrably has influence on the price of any supplies or the rendering cost of any service, for the execution of a contract.
- 3.18 "Goods" means any work, equipment, machinery, tools, materials or anything of whatever nature to be rendered to NHLS or NHLS's delegate by the Successful Bidder in terms of this bid.
- 3.19 "Historically Disadvantaged Individual" (HDI) means a South African citizen:
- 3.19.1 Who, due to the apartheid policy that had been in place, had no franchise in national elections prior to the introduction of the Constitution of the Republic of South Africa, 1983, (Act No. 110 of 1983) or the Constitution of the Republic of South Africa, 1993 (Act No. 200 of 1993) (the Interim Constitution); and/or;
- 3.19.2 who is a female; and/or;
- 3.19.3 who has a disability;
  - provided that a person who obtained South African citizenship on or after the coming to effect of the Interim Constitution is deemed not to be an HDI.
- 3.20 "Joint Venture" (Project) means two or more businesses joining together under a contractual agreement to conduct a specific business enterprise with both parties sharing profit and losses. The venture is for one specific project only, rather than for a continuing business relationship as in a strategic alliance. It is about sharing risk with others and providing one or more missing and needed assets and competencies.
- **"Management"** in relation to an enterprise or business, means an activity inclusive of control, and performed on a daily basis, by any person who is a principal executive officer of the company, by whatever name that person may be designated, and whether or not that person is a director.
- 3.22 "Military veteran"- has the meaning assigned to it in section 1 of the Military Veterans Act, 2011 (Act No. 18 of 2011).
- 3.23 "Non-firm Price(s)" means all price(s) other than "firm" price(s).
- "Organ of State" means a National Department or Provincial Administration as stipulated in Schedules 1 and 2 of the Public Service Act, Act 93 of 1994 (as amended).

Confidential Page **7** of **88** 



- 3.25 "Person(s)" refers to a natural and/or juristic person(s).
- 3.26 "Price"- includes all applicable taxes less all unconditional discounts;
- 3.27 **"Prime Contractor"** means any person (natural or juristic) who forwards an acceptable proposal in response to this RFB with the intention of being the main contractor should the proposal be awarded to him/her.
- 3.28 "Proof of B-BBEE status level of contributor" means -
  - (a) B-BBEE Status level certificate issued by an authorized body or person;
  - (b) A sworn affidavit as prescribed by the B-BBEE Codes of Good Practice; and
  - (c) Any other requirement prescribed in terms of the B-BBEE Act.
- 3.29 "QSE" means a qualifying small business enterprise in terms of a code of good practice on black economic empowerment issued in terms of section 9 (1) of the Broad-Based Black Economic Empowerment Act;
- 3.30 **"Rand Value"** means the total estimated value of a contract in South African currency, calculated at the time of invitations and includes all applicable taxes and excise duties.
- 3.31 "Rural Area" means -
  - (a) A sparsely populated area in which people farm or depend on natural resources, including villages and small town that are dispersed through the area; or
  - (b) An area including a large settlement which depends on migratory labour and remittances and government social grants for survival and may have a traditional land tenure system.
- 3.32 "SMME" bears the same meaning assigned to this expression in the National Small Business Act, 1996 (Act No. 102 of 1996).
- 3.33 "Stipulated minimum threshold" means the minimum threshold stipulated for local production and content.
- 3.34 **"Sub-contract"** means the primary contractor's assigning, leasing, making out work to, or employing, another person to support such primary contractor in the execution of part of a project in terms of the contract.
- 3.35 **"Subcontractor"** means any person (natural or juristic) who is subcontracted a portion of an existing contract by a Prime Contractor.
- 3.36 **"Successful Bidder"** means the organization or person with whom the order is placed and who is contracted to execute the work as detailed in the bid.
- 3.37 **"Township"** means an urban living area that any time from late 19<sup>th</sup> century until 27 April 1994, was reserved for black people, including areas developed for historically disadvantage individuals post 27 April 1994.

Confidential Page 8 of 88



3.38 **"Youth"** has the meaning assigned to it in section 1 of the National Youth Development Agency Act, 2008 (Act No. 54 of 2008).

#### 4. Acronyms and abbreviations

4.1 The following acronyms and abbreviations are used in this proposal and must be similarly used in the proposal submitted in response and shall have the meaning ascribed thereto below.

| Abbreviations/Acronyms | Description                                   |
|------------------------|---|
| BBBEE                  | Broad Based Black Economic Empowerment.       |
| СРІ                    | Consumer Price Index.                         |
| DIR                    | Directorate                                   |
| EDMS                   | Electronic Document Management System         |
| HDI                    | Historically Disadvantaged Individuals        |
| ISO                    | International Standard Organization           |
| IT                     | Information Technology                        |
| ITC                    | Information Technology Committee              |
| MISS                   | Minimum Information Security Standard         |
| OEM                    | Original Equipment Manufacturer               |
| PPPFA                  | Preferential Procurement Policy Framework Act |
| RFB                    | Request for Bid                               |
| RSA                    | Republic of South Africa                      |
| NHLS                   | National Health Laboratory Services           |
| SLA                    | Service Level Agreement                       |
| SW                     | Software                                      |
| LIS                    | Laboratory Information system                 |
| 24x7                   | 24 hours a day, 7 days a week                 |

## 5. General Rules and Instructions

## 5.1 Confidentiality

- 5.1.1 The information contained in this document is of a confidential nature and must only be used for purposes of responding to this RFB. This confidentiality clause extends to Bidder partners and/or implementation agents, whom the Bidder may decide to involve in preparing a response to this RFB.
- 5.1.2 For purposes of this process, the term "Confidential Information" shall include all technical and business information, including, without limiting the generality of the foregoing, all secret knowledge and information (including any and all financial, commercial, market, technical, functional and scientific information, and

Confidential Page 9 of 88



information relating to a party's strategic objectives and planning and its past, present and future research and development), technical, functional and scientific requirements and specifications, data concerning business relationships, demonstrations, processes, machinery, know-how, architectural information, information contained in a party's software and associated material and documentation, plans, designs and drawings and all material of whatever description, whether subject to or protected by copyright, patent or trademark, registered or un-registered, or otherwise disclosed or communicated before or after the date of this process.

- 5.1.3 The receiving party shall not, during the period of validity of this process, or at any time thereafter, use or disclose, directly or indirectly, the confidential information of NHLS (even if received before the date of this process) to any person whether in the employment of the receiving party or not, who does not take part in the performance of this process.
- 5.1.4 The receiving party shall take all such steps as may be reasonably necessary to prevent NHLS's confidential information coming into the possession of unauthorised third parties. In protecting the receiving party's confidential information, NHLS shall use the same degree of care, which does not amount to less than a reasonable degree of care, to prevent the unauthorised use or disclosure of the confidential information as the receiving party uses to protect its own confidential information.
- 5.1.5 Any documentation, software or records relating to confidential information of NHLS, which comes into the possession of the receiving party during the period of validity of this process or at any time thereafter or which has so come into its possession before the period of validity of this process:
- 5.1.5.1 shall be deemed to form part of the confidential information of NHLS;
- 5.1.5.2 shall be deemed to be the property of NHLS;
- 5.1.5.3 shall not be copied, reproduced, published or circulated by the receiving party unless and to the extent that such copying is necessary for the performance of this process and all other processes as contemplated in; and
- 5.1.5.4 shall be surrendered to NHLS on demand, and in any event on the termination of the investigations and negotiations, and the receiving party shall not retain any extracts.

## 5.2 News and press releases

5.2.1 Bidders or their agents shall not make any news releases concerning this RFB or the awarding of the same or any resulting agreement(s) without the consent of, and then only in co-ordination with, NHLS and its Client.

# 5.3 Precedence of documents

- 5.3.1 This RFB consists of a number of sections (see list). Where there is a contradiction in terms between the clauses, phrases, words, stipulations or terms and herein referred to generally as stipulations in this RFB and the stipulations in any other document attached hereto, or the RFB submitted hereto, the relevant stipulations in this RFB shall take precedence.
- 5.3.2 Where this RFB is silent on any matter, the relevant stipulations addressing such matter and which appear in the PPPFA shall take precedence. Bidders shall refrain from incorporating any additional stipulations in its

Confidential Page **10** of **88** 



proposal submitted in terms hereof other than in the form of a clearly marked recommendation that NHLS may in its sole discretion elect to import or to ignore. Any such inclusion shall not be used for any purpose of interpretation unless it has been so imported or acknowledged by NHLS.

5.3.3 It is acknowledged that all stipulations in the PPPFA are not equally applicable to all matters addressed in this RFB. It however remains the exclusive domain and election of NHLS as to which of these stipulations are applicable and to what extent. Bidders are hereby acknowledging that the decision of NHLS in this regard is final and binding. The onus to enquire and obtain clarity in this regard rests with the Bidder(s). The Bidder(s) shall take care to restrict its enquiries in this regard to the most reasonable interpretations required to ensure the necessary consensus.

#### 5.4 Preferential Procurement Reform

- 5.4.1 NHLS supports B-BBEE as an essential ingredient of its business. In accordance with government policy, NHLS insists that the private sector demonstrates its commitment and track record to B-BBEE in the areas of ownership (shareholding), skills transfer, employment equity and procurement practices (SMME Development) etc.
- 5.4.2 NHLS shall apply the principles of the Preferential Procurement Policy Framework Act, (Act No. 5 of 2000) to this proposal.
- 5.4.3 Bidders shall complete the preference certificate attached to this proposal. In the case of a consortium and sub-contractors, the preference certificate must be completed for each legal entity.

## 5.5 National Industrial Participation Programme

5.5.1 The Industrial Participation policy, which was endorsed by Cabinet on 30 April 1997, is applicable to contracts that have an imported content. The NIP is obligatory and therefore must be complied with. Bidders are required to sign and submit the Standard Bidding Document (SBD5).

#### 5.6 Language

5.6.1 Bids shall be submitted in English.

### 5.7 Gender

5.7.1 Any word implying any gender shall be interpreted to imply all other genders.

## 5.8 Headings

5.8.1 Headings are incorporated into this proposal and submitted in response thereto, for ease of reference only and shall not form part thereof for any purpose of interpretation or for any other purpose.

## 5.9 Security clearances

5.9.1 Employees and subcontractors of the successful bidder may be required to be in possession of valid security clearances to the level determined by the State Security Agency (SSA) and/or NHLS commensurate with the

Confidential Page 11 of 88



nature of the project activities they are involved in. The cost of obtaining suitable clearances is for the account of the bidders. The bidders shall supply and maintain a list of personnel involved on the project indicating their clearance status.

5.9.1 Employees and subcontractors of the successful bidder will be required to sign a non-disclosure agreement.

#### 5.10 Occupational Injuries and Diseases Act 13 of 1993

5.10.1 The Bidder warrants that all its employees (including the employees of any sub-contractor that may be appointed) are covered in terms of the Compensation for Occupational Injuries and Diseases Act 13 of 1993 and that the cover shall remain in force for the duration of the adjudication of this bid and/ or subsequent agreement. NHLS reserves the right to request the Bidder to submit documentary proof of the Bidder's registration and "good standing" with the Compensation Fund, or similar proof acceptable to NHLS.

#### 5.11 Formal contract

- 5.11.1 This RFB, all the appended documentation and the proposal in response thereto read together, form the basis for a formal contract to be negotiated and finalised between NHLS and/or its clients and the enterprise(s) to whom NHLS awards the bid in whole or in part.
- 5.11.2 Any offer and/or acceptance entered verbally between NHLS and any vendor, such offer shall not constitute a contract and thus not binding on the parties.

#### 5.12 Instructions for submitting a proposal

5.12.1 One (1) original, one (1) hard copy and 1 (one) electronic copy on compact disk (CD)/USB in Portable Document Format (PDF) of the Bid shall be submitted on the date of closure of the Bid.

Pricing: Bid Price must be submitted in a separate envelop and marked clearly as follows: RFB number, RFB description and bidder's name). One (1) original, one (1) hard copy and 1 (one) electronic copy on compact disk (CD)/USB in Portable Document Format (PDF) of the Bid shall be submitted on the date of closure of the Bid.

The original copy must be signed in black ink by an authorised employee, agent or representative of the bidder and each and every page of the proposal shall contain the initials of same signatories.

- 5.12.2 Bidders shall submit proposal responses in accordance with the prescribed manner of submissions as specified above.
- 5.12.3 Bids must be submitted in a prescribed response format herewith reflected as **Response Format**, and be sealed in an envelope clearly marked.
- 5.12.4 Bids that are too large to fit into the tender box must be handed in at the reception desk during office hours from 08:00- 16:30 or before 11:00 on the closing date.
- 5.12.5 All Bids in this regard shall only be accepted if they have been placed in the bid box before or on the closing date, **03 February 2026 and stipulated time, 11h00 am**.
- 5.12.6 Bids received after the time stipulated shall not be considered.

Confidential

Page 12 of 88



- 5.12.7 Bid responses sent by post or courier must reach this office at least **36 hours** before the closing date to be deposited into the proposal box. Failure to comply with this requirement shall result in your proposal being treated as a "late proposal" and shall not be entertained. Such proposal shall be returned to the respective bidders.
- 5.12.8 No proposal shall be accepted by NHLS if submitted in any manner other than as prescribed above.

## 6. Response format

6.1 Bidders shall submit response in accordance with the response format specified below. Failure to do so shall result rejecting vendor's response. No referrals may be made to comment. Failure to comply shall result in the vendor being penalised.

#### 6.2 Schedule Index:

- 6.2.1 Schedule 1: Pages 1 22 of this RFB document
- 6.2.2 Schedule 2: Mandatory Documents
- An original valid Tax Clearance Certificate or a Tax Compliance Status letter (with pin) issued by the South African Revenue Services or a CSD Report reflecting active Tax Clearance Compliance status.
  - If a Consortium, Joint Venture or Subcontractor, an original valid Tax Clearance Certificate or a Tax Compliance Status letter (with pin) issued by the South African Revenue Services or a CSD Report reflecting active Tax Clearance Compliance status must be submitted for each member. (Annexure B)
- 6.2.2.2 National Industrial Participation Programme Certificate from the DTI (read paragraph 5.5 in conjunction with Annex E SBD 5) (If applicable).
- 6.2.2.3 Central Supplier Database (CSD) Registration Report
- 6.2.2.4 General Conditions of Contract (Annexure E).
- 6.2.3 Schedule 3: Executive Summary of proposal
- 6.2.4 Schedule 4: Technical/Functionality
- 6.2.5 **Schedule 5**: Preferential Procurement Claim form and copy of the B-BBEE Verification Certificate(s) issued by an authorised body or person, or a sworn affidavit prescribed by the B-BBEE Codes of Good Practice.
- 6.2.7 **Schedule 6:** Bidder's Disclosure SBD 4 (Annexure C).
- 6.2.8 **Schedule 7**: Bidder Profile:
- 6.2.8.1 Credentials of the company/consortium members etc.
- 6.2.8.2 Structure of the company/ consortium members etc.
- 6.2.8.3 Partnership agreements/contracts
- 6.2.9 Schedule 8: Bid Price (to be submitted in a separate envelop and marked clearly as follows: RFB number, RFB description and bidder's name) (Annexure B).
- 6.3 Bidder background information materials:

Confidential Page 13 of 88

RFB Number 153/25/26: Placement of Total Automation System for Pre-Analytics, Analytic and Post-Analytical Systems for Chemical Pathology



- 6.3.1 <u>Bidder Operating Organisation</u> Provide an overview of the operating structure and geographical locations of the firm at the national, regional, and local levels.
- 6.3.2 <u>Standards</u> Include information regarding your firm's utilization of widely known Industry Standards and guidelines, as they apply to your firm, your firm's proposal and proposed hardware assets.
- 6.3.3 <u>Company Contact(s)</u> Provide the name, title, street address, city, state, telephone and fax numbers and e-mail of the primary company's contact person, and for any sub-Contractors.
- 6.3.4 <u>Corporate Financial Solvency</u> Provide solvency statement signed by a qualified independent auditor that the financial position of the company is sound and that the company will be able to mobilise financial resources to deliver the project.

## 7. Key personnel

7.1 Identify key personnel, by employer (include subcontractor(s), and provide contact information.

#### 8. Reasons for Disqualification

- 8.1 NHLS reserves the right to disqualify any bidder which does any one or more of the following, and such disqualification may take place without prior notice to the offending bidder, however the bidder shall be notified in writing of such disqualification:
- 8.1.1 bidders who submitted did not sign the mandatory documents;
- 8.1.2 bidders who submitted information that is fraudulent, factually untrue or inaccurate, for example memberships that do not exist, B-BBEE credentials, experience, etc.;
- 8.1.3 bidders who received information not available to other vendors through fraudulent means;
- 8.1.4 bidders who do not comply with mandatory requirements as stipulated in this RFB; and
- 8.1.5 bidders who fail to price according to the costing template provided;
- 8.1.6 bidders who failed to attend the compulsory briefing session and/or compulsory site visit.

## 9. Bid Preparation

- 9.1 All additions to the proposal documents i.e. annexes, supporting documentation pamphlets, photographs, technical specifications and other support documentation covering the goods offered etc. shall be neatly bound as part of the schedule concerned.
- 9.2 All responses regarding questions posed in the annex attached herewith shall be answered in accordance with the prescribed **RFB Response Format**.
- 9.3 Telephonic, faxed, e-mailed or oral tenders shall not be accepted.

#### 10. Oral presentations and Briefing Sessions

10.1 Bidders who submit Bids in response to this RFB may be required to give an oral presentation, which may include, but is not limited to, an equipment/service demonstration of their proposal to NHLS. This provides an opportunity for the vendor to clarify or elaborate on the proposal. This is a fact finding and explanation session only and does not include negotiation. NHLS shall schedule the time and location of these presentations. Oral presentations are

Confidential Page 14 of 88



an option of NHLS and may or may not be conducted and must not be construed as being successful in, or, awarded the tender.

- 11. General Conditions of Bid and Conditions of Contract
- Bidders shall provide full and accurate answers to all (including mandatory) questions posed in this document, and are required to explicitly indicate either "Comply/Accept (with a √)" or "Do not comply/Do not accept (with an X)" regarding compliance with the requirements. Where necessary, the bidder shall substantiate their response to a specific question.

NOTE: It is mandatory for bidders to complete or answer this part fully (11.2 to 11.34); otherwise, their bid shall be treated as incomplete and shall be disqualified. Refer to paragraph 8 of this document (reasons for disqualification).

11.2

| This bid is subject to the General Conditions of Contract stipulated in this document. | Accept | Do not Accept |
|--|--------|---------------|
| This bid is subject to the deficial conditions of contract supulated in this document. |        |               |

## 11.3

| The laws of the Republic of South Africa shall govern this RFB, and the Bidders hereby | Accept | Do not Accept |
|--|--------|---------------|
| accept that the courts of the Republic of South Africa shall have the jurisdiction.    |        |               |

## 11.4

| NHLS shall not be liable for any costs incurred by the bidder in the preparation of      | Accept | Do not Accept |
|--|--------|---------------|
| response to this RFB. The preparation of response shall be made without obligation       |        |               |
| to acquire any of the items included in any bidder's proposal or to select any proposal, |        |               |
| or to discuss the reasons why such vendor's or any other proposal was accepted or        |        |               |
| rejected.  |        |               |

## 11.5

| NHLS Procurement Services may request written clarification regarding any aspect of    | Accept | Do not Accept |
|--|--------|---------------|
| this proposal. The bidders must supply the requested information in writing within the |        |               |
| specified time frames after the request has been made, otherwise the proposal shall    |        |               |
| be disqualified.   |        |               |

# 11.6

| In the case of Consortium, Joint Venture or subcontractors, bidders are required to | Accept | Do not Accept |
|---|--------|---------------|
| provide copies of signed agreements stipulating the work split and Rand value.      |        |               |

Confidential Page 15 of 88

RFB Number 153/25/26: Placement of Total Automation System for Pre-Analytics, Analytic and Post-Analytical Systems for Chemical Pathology,



| In the case of Consortium, Joint Venture or subcontractors, all bidders are required to | Accept | Do not Accept |
|---|--------|---------------|
| provide mandatory documents as stipulated in schedule 1 of the Response format.         |        |               |

## 11.8

| NHLS reserves the right to; cancel or reject any proposal and not to award the        | Accept | Do not Accept |
|---|--------|---------------|
| proposal to the lowest bidder or award parts of the proposal to different bidders, or |        |               |
| not to award the proposal at all.   |        |               |

## 11.9

| Where applicable, bidders who are distributors, resellers and installers of network | Accept | Do not Accept |
|---|--------|---------------|
| equipment are required to submit back-to-back agreements and service level          |        |               |
| agreements with their principals.   |        |               |

# 11.10

| By submitting a proposal in response to this RFB, the bidders accept the evaluation | Accept | Do not Accept |
|---|--------|---------------|
| criteria as it stands.  |        |               |

## 11.11

| Where applicable, NHLS reserves the right to conduct benchmarks on | Accept | Do not Accept |
|--|--------|---------------|
| product/services offered during and after the evaluation.          |        |               |

## 11.12

| NHLS reserves the right to conduct a pre-award survey during the source selection   | Accept | Do not Accept |
|---|--------|---------------|
| process to evaluate contractors' capabilities to meet the requirements specified in |        |               |
| the RFB and supporting documents.   |        |               |

# 11.13

| Where the bid calls for commercially available solutions, bidders who offer provide | Accept | Do not Accept |
|---|--------|---------------|
| future based solutions will be disqualified.  |        |               |

## 11.14

| The bidder should not qualify the proposal with own conditions.                      | Accept | Do not Accept |
|--|--------|---------------|
| Caution: If the bidder does not specifically withdraw its own conditions of proposal |        |               |
| when called upon to do so, the proposal response shall be declared invalid.          |        |               |

Confidential Page **16** of **88** RFB Number 153/25/26: Placement of Total Automation System for Pre-Analytics, Analytic and Post-Analytical Systems for Chemical Pathology/



| Should the bidder withdraw the proposal before the proposal validity period expires,  | Accept | Do not Accept |  |
|---|--------|---------------|--|
| NHLS reserves the right to recover any additional expense incurred by NHLS having to  |        |               |  |
| accept any less favourable proposal or the additional expenditure incurred by NHLS in |        |               |  |
| the preparation of a new RFB and by the subsequent acceptance of any less favourable  |        |               |  |
| proposal.   |        |               |  |
|   |        |               |  |

# 11.16

| Delivery of and acceptance of correspondence between NHLS and the bidder sent by       | Accept | Do not Accept |
|--|--------|---------------|
|  |        |               |
| prepaid registered post (by air mail if appropriate) in a correctly addressed envelope |        |               |
| to either party's postal address or address for service of legal documents shall be    |        |               |
| deemed to have been received and accepted after (2) two days from the date of          |        |               |
| postage to the South African Post Office Ltd.  |        |               |
|  |        |               |

## 11.17

|   | Accept | Do not Accept |
|---|--------|---------------|
| Should the parties at any time before and/or after the award of the proposal and prior    | -      |               |
| to, and-or after conclusion of the contract fail to agree on any significant product      |        |               |
| price or service price adjustments, change in technical specification, change in          |        |               |
| services, etc. NHLS shall be entitled within 14 (fourteen) days of such failure to agree, |        |               |
| to recall the letter of award and cancel the proposal by giving the bidder not less than  |        |               |
| 90 (ninety) days written notice of such cancellation, in which event all fees on which    |        |               |
| the parties failed to agree increases or decreases shall, for the duration of such notice |        |               |
| period, remain fixed on those fee/price applicable prior to the negotiations.             |        |               |
|   |        |               |
| Such cancellation shall mean that NHLS reserves the right to award the same proposal      |        |               |
| to next best bidders as it deems fit.   |        |               |
|   |        |               |

#### 11.18

| In the case of a consortium or JV, each of the authorised enterprise's members and/or | Accept | Do not Accept |
|---|--------|---------------|
| partners of the different enterprises must co-sign this document.                     |        |               |

Page **17** of **88** 



| Any amendment or change of any nature made to this RFB shall only be of force and       | Accept | Do not Accept |
|---|--------|---------------|
| effect if it is in writing, and an Amendment to the RFB will be issued. Bidders will be |        |               |
| required to utilise the latest Amendment in preparation of their bid response.          |        |               |

## 11.20

| Failure or neglect by either party to (at any time) enforce any of the provisions of this | Accept | Do not Accept |
|---|--------|---------------|
| proposal shall not, in any manner, be construed to be a waiver of any of that party's     |        |               |
| right in that regard and in terms of this proposal. Such failure or neglect shall not, in |        |               |
| any manner, affect the continued, unaltered validity of this proposal, or prejudice the   |        |               |
| right of that party to institute subsequent action.                                       |        |               |

## 11.21

| Bidders who make use of subcontractors.   | Accept | Do not Accept |
|---|--------|---------------|
| The proposal shall however be awarded to the Vendor as a primary contractor who   |        |               |
| shall be responsible for the management of the awarded proposal. No separate      |        |               |
| contract shall be entered into between NHLS and/or its client and any such        |        |               |
| subcontractors. Copies of the signed agreements between the relevant parties must |        |               |
| be attached to the proposal responses.  |        |               |

## 11.22

| All services supplied in accordance with this proposal must be certified to all legal | Accept | Do not Accept |
|---|--------|---------------|
| requirements as per the South African law.  |        |               |

## 11.23

| No interest shall be payable on accounts due to the successful vendor in an event of a | Accept | Do not Accept |
|--|--------|---------------|
| dispute arising on any stipulation in the contract.                                    |        |               |

# 11.24

| Evaluation of Bids shall be performed by a CFET established by NHLS.                  | Accept | Do not Accept |
|---|--------|---------------|
|   |        |               |
| Bids shall be evaluated on the basis of conformance to the required specifications as |        |               |
| outlined in the RFB. Points shall be allocated to each bidder, on the basis that the  |        |               |
| maximum number of points that may be scored for price is 80/90, and the maximum       |        |               |
| number of preference points that may be claimed for Specific Goals (according to the  |        |               |
| PPPFA) is 20/10.  |        |               |

Confidential Page 18 of 88
RFB Number 153/25/26: Placement of Total Automation System for Pre-Analytics, Analytic and Post-Analytical Systems for Chemical Pathology/



| Prior to the award of any tender or contract the NHLS will check the Prohibition status      | Accept | Do not Accept |
|--|--------|---------------|
| of recommended suppliers/ service providers on the Treasury website                          |        |               |
| ( <u>restricted@treasury.gov.za</u> ) as well as the Treasury Register for Tender Defaulters |        |               |
| (www.treasury.gov.za)  |        |               |

## 11.26

|  | Accept | Do not Accept |  |
|--|--------|---------------|--|
| The NHLS will act against the bidder or person awarded the contract upon detecting |        |               |  |
| that the B-BBEE status level of contribution has been claimed or obtained on a     |        |               |  |
| fraudulent basis or any of the contract conditions have not been fulfilled.        |        |               |  |
|  |        |               |  |

# 11.27

| The NHLS may, in addition to any other remedy that it may have against the bidder             | Accept | Do not Accept |
|---|--------|---------------|
| or person:  |        |               |
| Disqualify the bidder or person from the bidding process;                                     |        |               |
| Recover all costs, losses or damages it has incurred;   |        |               |
| <ul> <li>or suffered as a result of that person's conduct;</li> </ul>                         |        |               |
| Cancel the contract and claim any damages which it;   |        |               |
| <ul> <li>has suffered as a result of having to make less;</li> </ul>                          |        |               |
| favourable arrangements due to such cancellation;   |        |               |
| Restrict the bidder or contractor, its shareholders;  |        |               |
| <ul> <li>and directors, or only the shareholders and directors;</li> </ul>                    |        |               |
| <ul> <li>who acted on a fraudulent basis, from obtaining business;</li> </ul>                 |        |               |
| <ul> <li>from any organ of state for a period not exceeding 10;</li> </ul>                    |        |               |
| <ul> <li>years, after applying the audi alteram partem (hear the other side) rule;</li> </ul> |        |               |
| Forward the matter for Blacklisting by Treasury; and  |        |               |
| Forward the matter for criminal prosecution   |        |               |

## 11.28

| If the successful bidder disregards contractual specifications, this action may result in | Accept | Do not Accept |  |
|---|--------|---------------|--|
| the termination of the contract.  |        |               |  |

## 11.29

| The bidders' response to this Tender, or parts of the response, shall be included as a | Accept | Do not Accept |
|--|--------|---------------|
| whole or by reference in the final contract.   |        |               |



| NHLS has discretion to extend the validity period should the evaluation of this bid not | Accept | Do not Accept |
|---|--------|---------------|
| be completed within the stipulated validity period.                                     |        |               |

#### 11.31

| Upon receipt of the request to extend the validity period of the bid, the bidder must | Accept | Do not Accept |
|---|--------|---------------|
| respond within the required time frames and in writing on whether or not he agrees    |        |               |
| to hold his original bid response valid under the same terms and conditions for a     |        |               |
| further period.   |        |               |

## 11.32

| Should the bidder change any wording or phrase in this document, the bid shall be | Accept | Do not Accept |
|---|--------|---------------|
| deemed unresponsive and may lead to the disqualification of the bid response.     |        |               |

#### 11.33

| The cost validation for Analysers and reagents for the validation will be borne by the supplier and NHLS will not be charged for this. | Accept | Do not Accept |
|--|--------|---------------|
|  |        |               |

#### 11.34

| No alternative tander offers will be considered  | Accept | Do not Accept |
|--|--------|---------------|
| No alternative tender offers will be considered. |        |               |

# 12. NHLS Special Conditions of Contract

12.1 Bidders shall provide full and accurate answers to all (including mandatory) questions posed in this document, and are required to explicitly indicate either "Comply/Accept (with a  $\sqrt{\ }$ " or "Do not comply/Do not accept (with an X)" regarding compliance with the requirements. Where necessary, the bidder shall substantiate their response to a specific question.

NOTE: It is mandatory for bidders to complete or answer this part fully (12.1.1 to 12.1.25); otherwise, their bid shall be treated as incomplete and shall be disqualified. Refer to paragraph 8 of this document (reasons for disqualification).

| 12.1.1 Applicable Hardware   | Accept | Do Not Accept |
|--|--------|---------------|
| There must be upgradeable or up scalable at the supplier's cost in the event |        |               |
| of new technologies, capabilities, changes in work volume and instrument     |        |               |
| suboptimal performance.  |        |               |

| 12.1.2 | Any software updates within the five years of warranty period       | Accept | Do Not Accept |
|--------|---|--------|---------------|
|        | should be at bidders' cost, this is to ensure no additional cost is |        |               |
|        | charged by the bidder.  |        |               |

Confidential Page 20 of 88



| 12.1.3  | Downtime   | Accept | Do Not Accept |
|---------|--|--------|---------------|
|         | oplier to provide alternative testing platform if instrument me is outside the laboratory and analyte specific turnaround time.    |        |               |
| <u></u> |  |        |               |
| 12.1.4  | The supplier must state all user-replaceable parts and   | Accept | Do Not Accept |
|         | consumables required for the duration of the contract and the  |        |               |
|         | appropriate replacement frequency.   |        |               |
|         |  |        | •             |
| 12.1.5  | Supply unlimited initial and continual technical training of lab staff   | Accept | Do Not Accept |
|         | on-site for the duration of the contract. This includes appropriate  |        |               |
|         |  |        |               |
|         | testing (both written and witnessing) immediately after training   |        |               |
|         | testing (both written and witnessing) immediately after training as well as an on-going basis for technical competency assessment. |        |               |

| 12.1.6 | Contract to include cost of performing mandatory analyte/parameter                    | Accept | Do Not Accept |
|--------|---|--------|---------------|
|        | verification as per protocol. Supplier to provide all reagents, controls, calibrators |        |               |
|        | and consumables required to perform the verification, including EP15 and              |        |               |
|        | linearity control material. Control material supplied must be from two different      |        |               |
|        | lot numbers: one lot number to control the analyser and one lot number to             |        |               |
|        | perform verification.   |        |               |

| 12.1.7 | The quotation must be for the total cost of the solution, including costs of      | Accept | Do Not Accept |
|--------|---|--------|---------------|
|        | reagents, all consumables including those in the kits, controls, calibrators,     |        |               |
|        | leasing of the equipment, insurance, repairs, maintenance, calibration of all     |        |               |
|        | centrifuges including certification and upgrades. Must include any other          |        |               |
|        | equipment or services or installations that will be required for complete         |        |               |
|        | functionality of the proposed system, e.g., installation of water purification    |        |               |
|        | systems, pumps or lines, installation of additional IT infrastructure, additional |        |               |
|        | software programs or software updates, middleware licences, etc.                  |        |               |

| 12.1.8 | The middleware functionality (review of results, review of QC, management of | Accept | Do Not Accept |
|--------|--|--------|---------------|
|        | rules, etc.) must be installed in a minimum of 20 computers/ laptops. All    |        |               |
|        | authorised laboratory staff must be allowed login access on any of the       |        |               |
|        | computers/ laptops.  |        |               |

Confidential Page **21** of **88** 



| 12.1.9 | The cost of renovations required to place the configuration in the space         | Accept | Do Not Accept |
|--------|--|--------|---------------|
|        | available must be considered as part of the cost of this solution. Supplier to   |        |               |
|        | obtain a quotation, which will be evaluated as part of the financial evaluation. |        |               |

| 12.1.10 | The implementation plan must take account of the physical constraints of the   | Accept | Do Not Accept |
|---------|--|--------|---------------|
|         | laboratory space available. The system must be compatible with the site's      |        |               |
|         | existing power, safety, water, plumbing. The new system must fit into the      |        |               |
|         | current floor space. Any changes should be at the supplier's cost. Structural  |        |               |
|         | pillars cannot be moved. The plan must permit the laboratory to continue to be |        |               |
|         | fully functional throughout implementation. Turn-around times must remain      |        |               |
|         | intact.  |        |               |
| 12.1.11 | Analytical liquid waste must comply with municipal by-laws, and environmental  | Accept | Do Not Accept |
|         | and waste management laws regarding toxic or biohazardous waste disposal       |        |               |
|         | and provide relevant certification/letter.                                     |        |               |

| 12.1.12 T | he solution must be tailored to the laboratory size, scope of work, sample flow   | Accept | Do Not Accept |
|-----------|---|--------|---------------|
| а         | nd turnaround requirements. It must accommodate all current pre-analytic          |        |               |
| а         | and analytic functionality, 98% of the current test repertoire for chemistry,     |        |               |
| 1         | .00% for coagulation and serology as attached in the tender document (use of      |        |               |
| tl        | hird-party reagents to fulfil test repertoire is acceptable), sample arrival flow |        |               |
| р         | peaks, test volumes, and required turnaround times according to NHLS agreed-      |        |               |
| u         | pon targets as well as client Service Level Agreements (SLA). The solution must   |        |               |
| h         | nave sufficient space for an estimated 10% increase in volume per annum.          |        |               |

| 12.1.13 Provide a computer aided diagram of the layout of the solution, demonstrating | Accept | Do Not Accept |
|---|--------|---------------|
| how the proposed solution will fit in the available space including the following     | :      |               |
| Individual analyser weights, and total weight   |        |               |
| Floor loading, anti-vibration and anti-static requirements                            |        |               |
| Heating, ventilation and cooling requirements   |        |               |
| Drainage and liquid waste removal requirements  |        |               |
| Optimal operating temperature requirements  |        |               |
| Availability of instrument malfunction and reagent critical level alarms              |        |               |
|   |        |               |

Confidential Page 22 of 88



| 12.1.14 System must interface with NHLS LIS prior to acceptance of system. Interface | Accept | Do Not Accept |
|--|--------|---------------|
| development must be at supplier cost.  |        |               |

| 12.1.15 Provide the following detailed information/evidence per test:                 | Accept | Do Not Accept |
|---|--------|---------------|
| The International Standards to which each test method is traceable                    |        |               |
| The methodology or methodologies employed per test per analyser                       |        |               |
| Performance characteristics per test (measuring range, LOD, linearity,                |        |               |
| sensitivity, specificity, interferences, reproducibility, % equivocal data (Virology- |        |               |
| serology tests)   |        |               |
| Recommended reference intervals per test, and the study data these were               |        |               |
| derived from.   |        |               |
| For hs-cardiac troponin testing, provide the 99th percentile of the Upper             |        |               |
| Reference Limit (URL), and provide the analytical CV at the 99th percentile of        |        |               |
| the URL   |        |               |

| 12.1.16 Provide an electronic package insert per test. | Accept | Do Not Accept |
|--|--------|---------------|
|  |        |               |

| 12.1.17 | It will be expected of the supplier to work closely with the laboratory IT          | Accept | Do Not Accept |
|---------|---|--------|---------------|
|         | personnel to identify required interfaces and to design an effective interface      |        |               |
|         | system. All the required hardware and software must be provided and installed.      |        |               |
|         | All the cabling from the host computers to the laboratory information system        |        |               |
|         | (LIS) must be provided. It will be expected of the supplier to have a specialist on |        |               |
|         | site during interfacing until the system is fully functional and validated          |        |               |
|         | according to specifications. All the interfaces must be tested and fully            |        |               |
|         | operational before implementation.  |        |               |

| 12.1.18 An instrument technician / engineer, and an application specialist, must at all | Accept | Do Not Accept |
|---|--------|---------------|
| times be available to NHLS at Dr George Mukhari Academic Laboratory to                  |        |               |
| provide an uninterrupted service. If an on-site application specialist is not           |        |               |
| available, then remote access for troubleshooting is required. Instrument               |        |               |

Confidential Page 23 of 88



|         | engineers must be available on a 24-hour basis with on-site technical           |        |               |
|---------|---|--------|---------------|
|         | assistance, when required, and backup to be available from supplier. Response   |        |               |
|         | time from the time a call is logged for analyser breakdown or malfunction until |        |               |
|         | the arrival of the technician or engineer on-site must be less than 2 hours.    |        |               |
|         |   |        |               |
| 12.1.19 | Preventive maintenance of instrument by supplier with agreed-upon schedule      | Accept | Do Not Accept |
|         | and defined timeframes for downtime.  |        |               |
|         |   |        |               |
| 12.1.20 | Where instrument calibration and controls are required to be performed by the   | Accept | Do Not Accept |
|         | engineer after servicing or replacement of a major part, this calibration and   |        |               |
|         | controls must be included in the service.                                       |        |               |
|         |   |        |               |
| 12.1.21 | Time interval for supply and delivery of reagents and consumables from receipt  | Accept | Do Not Accept |
|         | of order must be delivered on/or before the need by date specified on the       | -      | -             |
|         | order.  |        |               |
|         |   |        |               |
| 12.1.22 | The supplier must provide at least two operational platforms (2 chemistry, 2    | Accept | Do Not Accept |
|         | endocrinology), two Coagulation platforms and 1 ESR for Core Lab, 1             |        |               |
|         | operational platform for STAT Lab (Chemistry and Endocrinology) and 1           |        |               |
|         | Coagulation platform  |        |               |
|         |   |        |               |
| 12.1.23 | The system should accommodate all commonly used tube sizes and cap              | Accept | Do Not Accept |
|         | types/colours from multiple suppliers on all the instruments.                   |        |               |
|         |   |        |               |
| 12.1.24 | Prior analyser evaluation by NHLS HTA and/or peer-reviewed publications of the  | Accept | Do Not Accept |
| instrum | ent's analytical performance evaluation.  |        |               |
|         |   |        | <u> </u>      |
| 12.1.25 | Provision of reagent and calibrator performance monitoring and updates          | Accept | Do Not Accept |
|         | timeously and recall of malfunctioning reagent when required.                   | •      | •             |
|         |   |        |               |

## 13. Evaluation Criteria and Methodology

## 13.1 Evaluation of tenders and selection of contractors'/service providers

The NHLS is a Schedule 3A Government Institution subjected to the Public Finance Management Act (PFMA), the Public Preferential Framework Act (PPPFA) and Treasury Regulations/ Instructions. Bidders must assist the NHLS to eliminate corruption and fraud by completing and submitting form SBD4.

Confidential Page 24 of 88



- 13.1.1. Any tender closing is followed by a Public Opening where the names and pricing of all bids received are read out to the bidders attending. NHLS tender opening officials sign the pages where pricing is indicated to prevent any alterations.
- 13.1.2 Next steps of evaluation is administrative pre-qualification verification and the "technical" or so called "functional" evaluation which is purely based on NHLS specifications (Annexure 2) and Scope of Work. NHLS end-user department (who requested the bid), Procurement Services, Finance and subject specialists are part of the Cross Functional Evaluation Team (CFET) meeting which is chaired by Quality Assurance (QA). All the members of the CFET must complete Declaration of Interest forms and must recuse themselves in case of any conflict of interest.
- 13.1.3 The final stage of evaluation is done after the CFET has reached their verdict and is done by NHLS Procurement Services and separately from the CFET meeting. Points for Specific Goals (commercial evaluation) is being added in order to get the final order of merit for the bidders being evaluated.
- 13.1.4 Bidders that score the minimum threshold are recommended and submitted to the NHLS Tender Evaluation Committee (TAC) for adjudication and the bid MUST be awarded to the bidder who scored the highest points (Merit 1) during the CFET and Commercial evaluation(B-BBEE). All the members of the CFET must complete Declaration of Interest forms and must recuse themselves in case of any conflict of interest. Should the TAC decide on a bidder other than Merit 1, this decision must be motivated as a Deviation from NHLS Policy & procedure and Treasury must be advised accordingly.
- 13.1.5 The CEO of the NHLS must finally approve the recommendation by the TAC, in his capacity as the Accounting Officer.
- 13.1.6 Details of the successful bidder to be advertised in the Government Tender Bulletin.
- 13.1.7 Suppliers must accept the Terms & Conditions of our contract(s) which will result from the RFB document".

  RFB conditions and pricing shall be fixed and firm from RFB closing date to the end of contract.

## 13.2 BID EVALUATION STAGES

The bid evaluation process consists of several stages that are applicable according to the nature of the bid as defined below:

- Stage 1: Administrative Compliance verification.
- Stage 2: Technical Mandatory requirement evaluation.
- Stage 3: Technical Functionality requirement evaluation.
- Stage 4: Price / Specific Goals evaluation.

NOTE: The bidder must qualify for each stage to be eligible to proceed to the next stage of the evaluations.

## 13.3 ADMINISTRATIVE COMPLIANCE REQUIREMENTS

Administrative compliance/responsiveness will be tested based on returnable documents submitted and signatures on the Bid documents.

Confidential Page 25 of 88



♣ At this stage, it must be determined what documents are required to be returned by Bidders. Returnable documents are categorised as follows:

## **Mandatory Returnable documents**

(NOTE: Failure to provide the below listed documents WILL lead to disqualification).

| 1. | Proof of Attendance of Compulsory Site Briefing session to measure designated | Comply        | Do Not Comply      |
|----|---|---------------|--------------------|
|    | space and weight restrictions.  |               |                    |
| Sı | bstantiation: The bidder is to indicate whether they attended the Compulsory  | site briefing | session to measure |
| de | signated space and weight restrictions.                                       |               |                    |

| 2.The Service Providers have to agree with NHLS General Conditions of Bid and | Comply | Do Not Comply |
|---|--------|---------------|
| Conditions of Contract.   |        |               |

Substantiation: The bidder must submit and attach to the bid response the signed and accepted NHLS General Conditions of Bid and Conditions of Contract.

| 3. The Service Providers have to agree with NHLS Special Conditions of Contract. |             | Do Not Comply       |
|--|-------------|---------------------|
|  |             |                     |
| Substantiation: The bidder must submit and attach to the bid response accep      | ted NHLS Sp | ecial Conditions of |
| Contract.  |             |                     |

## **Essential Returnable documents**

(NOTE: Failure to provide the below listed documents MAY lead to disqualification).

| . Fully completed and signed Bidder's Disclosure SBD 4.   | Comply | Do Not Comply |  |  |
|---|--------|---------------|--|--|
|   |        |               |  |  |
| Substantiation: The bidder must submit and attach to the bid response the signed Bidder's Disclosure. |        |               |  |  |

| 2. Preferential Procurement Claim form and copy of the B-BBEE Verification                           | Comply | Do Not Comply |
|--|--------|---------------|
| Certificate(s) issued by an authorised body or person or a sworn affidavit prescribed                |        |               |
| by the B-BBEE Codes of Good Practice.  |        |               |
| Substantiation: The bidder must submit and attach to the bid response a copy of a valid certificate. |        |               |

| 3. Submission of original valid Tax Clearance Certificate, a Tax Compliance Status letter            | Comply | Do Not Comply |
|--|--------|---------------|
| (with pin) issued by the South African Revenue Services, or a CSD Report reflecting                  |        |               |
| active Tax Clearance Compliance status.  |        |               |
| Substantiation: The bidder must submit and attach to the bid response a copy of a valid certificate. |        |               |

| 4. Proof of Central Supplier Database (CSD) Registration.                  | Comply | Do Not Comply |
|--|--------|---------------|
|  |        |               |
| Substantiation: The bidder must submit a CSD Report with the bid response. |        |               |

Confidential Page 26 of 88



of conformity.

5. Audited Financial Statement not older than two (2) years (if applicable).

| response.  |                |               |  |  |
|--|----------------|---------------|--|--|
|  |                |               |  |  |
| 6. The product must comply with the following:   | Comply         | Do Not Comply |  |  |
| (a) Environmental Safety compliant (Provide proof by means of VALID  |                |               |  |  |
| Certificates/letter of conformity from the regulator).   |                |               |  |  |
| (b) Occupation Health and Safety (OHS) (Provide proof by means of  |                |               |  |  |
| letter/Certificates).  |                |               |  |  |
| Substantiation: The bidder must submit and attach to the bid response, for (a) proof   | by means of \  | /ALID         |  |  |
| Certificates/letters of conformity from the regulator, for (b) proof by means of letter  | /Certificates. |               |  |  |
|  |                |               |  |  |
| 7. The product must be approved by any of the IMDRF regulatory authorities   | Comply         | Do Not Comply |  |  |
| listed below. (Note: Approval are at the bidder's cost).   |                |               |  |  |
| Substantiation: The bidder is to provide at least one certificate of the IMDRF Regula  | tory Authority | below:        |  |  |
| Australia: Therapeutic Goods Administration.   |                |               |  |  |
| Brazil: National Health Surveillance Agency (ANVISA).  |                |               |  |  |
| • Canada: Health Canada.   |                |               |  |  |
| China: China Food and Drug Administration.   |                |               |  |  |
| <ul> <li>European Union: European Commission Directorate-General for Internal Market, Industry,</li> <li>Entrepreneurship and SMEs.</li> </ul> |                |               |  |  |
| Japan: Pharmaceuticals and Medical Devices Agency and the Ministry of He   | alth, Labour a | nd Welfare.   |  |  |
| Russia: Russian Ministry of Health.  |                |               |  |  |
| Singapore: Health Sciences Authority.  |                |               |  |  |
| South Korea: Ministry of Food and Drug Safety.   |                |               |  |  |
| <ul> <li>United States of America: US Food and Drug Administration (FDA).</li> </ul>   |                |               |  |  |
|  |                |               |  |  |
| 8. The product must be ISO 13485 compliant.  | Comply         | Do Not Comply |  |  |
|  |                | 1             |  |  |

Substantiation: The bidder must submit and attach a financial statement not older than two (2) years with the bid

**Do Not Comply** 

Do Not Comply

Comply

Comply

Confidential Page 27 of 88

9. Supplier to provide calibration and calibration certificate by an accredited service

provider as per supplier's recommendations for all centrifuges.

Substantiation: The bidder must submit and attach to the bid response a proof by means of a valid certificate/ letter



Substantiation: Supplier to provide commitment letter (to outsource supplier for calibration if required).

# 13.4 The 80/20 or 90/10 PPPFA principle and the points for evaluation criteria are as follows:

| Price points   | 80/90      |
|----------------|------------|
| Specific goals | 20/10      |
| Total          | 100 points |

Confidential Page 28 of 88



#### **ANNEXURE A: Technical Specification**

#### 1 SPECIAL INSTRUCTIONS TO VENDORS

- 1.1 Should a Bidder have reasons to believe that the Technical Specification is not open and/or is written for a particular brand or product; the Bidder shall notify Procurement Services within ten (10) days after publication of the bid.
- 1.2 Bidders shall provide full and accurate answers to the mandatory questions posed in this document, and, where required explicitly state either "Comply/Not Comply" regarding compliance with the requirements. Bidders **must** substantiate their response to all questions, including full details on how their proposal/solution will address specific functional requirements. All documents as indicated must be supplied as part of the submission.
- 1.3 Bidders are encouraged to promote the growth and development of SMME's, and will be assessed on their efforts in this regard during the evaluation of this Tender.

#### 2 ACRONYMS AND ABBREVIATIONS

| Term | Definition                         |
|------|------------------------------------|
| EBS  | Oracle e-Business Suite            |
| DR   | Disaster Recovery                  |
| DB   | Database                           |
| NHLS | National Health Laboratory Service |
| PMO  | Project Management Office          |
| SLA  | Service Level Agreement            |

## 3 BACKGROUND

**Chemical Pathology/Stat Laboratory /Coagulation/ ESR**: The automated systems at Dr George Mukhari Academic Hospital (DGM) have reached the end of the contract and are therefore, due for replacement.

The advantages of total automation, including the pre-analytical automation and analysers, are as follows: Reduction in pre-analytical errors, standardisation, quality improvement, removal of many manual steps, automated repeat and reflex testing, improved specimen tracking, improved turnaround time, increased throughput, improved efficiency, increase in productivity, improvement in safety and reduction in overall cost.

The challenges with ageing instruments include but are not limited to deteriorating analytical performance, frequent downtimes due to breakdowns and increased costs due to the replacement of parts. All of these have an adverse impact on workflow and turnaround time and lead to our customers' overall loss of confidence

Confidential
Page 29 of 88

REP Number 153/25/26: Placement of Total Automation System for Dro Analytics Analytic and Bost Analytical Systems for Chamical Pathology



#### 4 SCOPE OF WORK

**Chemistry/ virology**: Fully Automated Track system with a pre-analytical module for online Centrifugation, sorting and aliquoting, Middleware and Inventory system. Analysers should accommodate Chemistry, Endocrinology, and Immunology, and provide an online refrigerated Storage Facility. In addition, the contract should include and support a standalone chemistry and endocrine analyser for operations at the Stat laboratory similar to the Main Chemistry lab platform.

- 4.2. **A fully automated stand-alone ESR analyser**. The ESR analyser must be able to accommodate both paediatric and adult EDTA tubes. The method of analysis should be by the Westergren method.
- **4.3. Coagulation**: Two fully automated Coagulation analysers with cap piercing capability that must be able to accommodate the full existing test volume and repertoire in the Core Lab and 1 in STAT lab (see test list attached).

The following tables Test Volumes are inserted:

- Table 1: DGM Chemical Pathology Laboratory Volumes
- Table 2: DGM ESR and Coagulation Volumes
  - Table 3: DGM Virology (Serology) Laboratory Volumes

#### 5 MANDATORY REQUIREMENTS

If a bidder does not comply fully with each of the mandatory requirements, it Shall be regarded as mandatory non-performance/non-compliance and the bid Shall be disqualified. No "unanswered" questions will be allowed. If a response to a question has been indicated as comply but not elaborated upon or substantiated it shall be regarded as mandatory non-performance/non-compliance, and the bid shall be disqualified.

Bidders shall provide full and accurate answers to the mandatory questions posed in this document, and, where required, explicitly state either "Comply/Accept (with a "Yes")" or "Do not comply/do not accept (with a "No")" regarding compliance to the requirements. Bidders must substantiate their responses to all mandatory questions. PLEASE NOTE: If the response does not substantiate any of the points or requirements in the body of the tender, it will be deemed to not comply, even if the 'Comply' field has been marked. Please note: All documentation to substantiate the mandatory requirements has to be supplied.

Confidential
Page 30 of 88
RFB Number 153/25/26: Placement of Total Automation System for Pre-Analytics. Analytic and Post-Analytical Systems for Chemical Patholog



## **TECHNICAL MANDATORY REQUIREMENTS:**

The National Health Laboratory Service (NHLS) – Dr George Mukhari Academic Laboratories seeks to appoint a single Service Provider to supply, install, integrate, validate, and support a Comprehensive Full Automation Solution covering multiple diagnostic testing disciplines. The solution must streamline workflow, reduce manual handling, improve TAT (turnaround time), and ensure high reliability and quality across all high-volume and specialised testing areas. Scope of Automation

The bidder must provide a unified automation platform and connected systems capable of supporting the following test subcategories: Chemical Pathology / Clinical Chemistry, Coagulation, ESR (Erythrocyte Sedimentation Rate), STAT / Emergency Laboratory and Serology.

The solution must be fully scalable and handle the submitted annual test volumes provided by NHLS (to be inserted or referenced by the bidder). The bidder should ensure that proposed systems can manage peak workflows and support future growth.

#### All departments technical mandatory requirements

| 1. | Fully automated system with pre-analytical, analytical, and post-analytical components. This includes a fully automated track system, online Centrifugation with bulk loader, sorting and aliquotting, Middleware, decapper and recapper/sealer, automatic disposal system, analysers to accommodate Chemistry, Endocrinology, Coagulation, ESR and Virology (Serology), and an online refrigerated Storage Facility to store total work volume for 7 days. The track should include an aliquoter that can be assigned to specific target racks for sorting. | Comply      | Do Not Comply  |
|----|--|-------------|----------------|
|    | Substantiate: Provide proof using a brochure/s indicating each of the requirements.  |             |                |
| 2. | Provide a Chemistry, Coagulation and Endocrine analyser similar to the Main Laboratory Chemical Pathology for the Stat lab for standardization.  | Comply      | Do Not Comply  |
|    | Substantiation: The bidder must submit a commitment letter.  |             |                |
| 3. | Supply of water purification system with electronic readings, monitoring system, alarm error notification, with linked UPS and reservoir. Capable of fully operating under a low-pressure environment. Provision of the water purification and maintenance/ repairs thereof must be at the bidders' cost for the duration of the contract if needed. It should include an additional >10 000 litres Water Reservoir system e.g. JOJO tank.   | Comply      | Do Not Comply  |
|    | Substantiate: Provide proof using a brochure/specification and a commitment letter for system.   | or the wate | r reservoir    |
| 4. | The system should accommodate all tube sizes and all sample types (Serum, plasma, whole blood, urine, CSF and other body fluids) on all the instruments, alternatively, the bidder should provide off-line centrifuges and paediatric sample cups as well as storage solutions for low-volume samples.   | Comply      | Do Not Comply  |
|    | Substantiate: Provide proof using a brochure and signed commitment letter.   | <u>I</u>    |                |
| 5. | The automated Track system should be capable of connecting all analysers tendered and should also be able to accommodate third-party analysers.  | Comply      | Do Not Comply  |
|    | Substantiation: The bidder must submit and attach to the bid response a catalogue/br provide a list of third-party analysers that can be connected to the Track.   | ochure. Th  | e bidder is to |

Confidential
Page 31 of 88

PER Number 153/25/26: Placement of Total Automation System for Pro Analytics Analytics and Post Analytical Systems for Chemical Pathology

RFB Number 153/25/26: Placement of Total Automation System for Pre-Analytics, Analytic and Post-Analytical Systems for Chemical Pathology/ STAT Laboratory for National Health Laboratory Service DGM including reagents, consumables & service contract for a period of five (5) years.



| 6.  | The track system should make use of a multi-lane system or design that will  | Comply       | Do Not Comply   |  |  |
|-----|--|--------------|-----------------|--|--|
|     | accommodate bi-directional movement and limit delays to analysers on the track.  |              |                 |  |  |
|     | Substantiation: Provide proof using a brochure/catalogue.  |              |                 |  |  |
| 7.  | Supplier to provide the ability for planned purchase orders for the duration of the  | Comply       | Do Not Comply   |  |  |
|     | tender and QC and Reagent Lot reservations of at least 6 months. IQC must be   |              |                 |  |  |
|     | delivered directly to the site.  |              |                 |  |  |
|     | Substantiation: The supplier must provide details and a signed commitment letter.  |              |                 |  |  |
| 8.  | The facility must be able to set up and view patient-based quality control monitoring                                      | Comply       | Do Not Comply   |  |  |
|     | for every analyte (for additional quality control). Provide documented evidence and a tracking system.                     |              |                 |  |  |
|     | Substantiation: The bidder must submit and attach to the bid response a catalogue/br                                       | ochure.      |                 |  |  |
| 9.  | Provision of an application specialist dedicated to our laboratory and an on-standby                                       | Comply       | Do Not Comply   |  |  |
|     | engineer 24/7.   |              |                 |  |  |
|     |  |              |                 |  |  |
|     | Substantiation: The bidder must submit a commitment letter.  |              |                 |  |  |
| 10. | The entire system must be capable of fulfilling peak workflow volumes to meet  | Comply       | Do Not Comply   |  |  |
|     | turnaround time as per laboratory TAT SOP. Chemical pathology-1000 tests/hour.   |              |                 |  |  |
|     | Coagulation – 50-75 tests/hour, virology (Serology) - 250 tests per hour and ESR - >10                                     |              |                 |  |  |
|     | tests/hour and accommodate at least a 10% increase per year for the duration of the tender.                                |              |                 |  |  |
|     | terider.   |              |                 |  |  |
|     | Substantiation: The bidder must submit and attach to the bid response a relevant app or attachments.                       | licable cata | alogue/brochure |  |  |
| 11. | System to accommodate all chemistry, and coagulation (including Stat lab and third   | Comply       | Do Not Comply   |  |  |
|     | party) analysers on the common Middleware programme with ability for offsite   |              |                 |  |  |
|     | monitoring/reviewing. If the middleware requires Wi-Fi, it should be provided by the                                       |              |                 |  |  |
|     | supplier. In addition, the provision of an adequate number of licences to support 15                                       |              |                 |  |  |
|     | Middleware access points for Chem Path and 2 Middleware access points for Stat lab.  | ochuro ac    | woll as a       |  |  |
|     | Substantiation: The bidder must attach and submit to the bid response a catalogue/brochure as well as a commitment letter. |              |                 |  |  |
| 12. | Ability to set up user-defined rules on the Middleware for IQC, Reruns, Repeats,   | Comply       | Do Not Comply   |  |  |
|     | Dilutions, Blocking Results, and Reflex Testing.   |              |                 |  |  |
|     | Substantiation: Provide evidence- by brochure, catalogue, and a screenshot of the facility.                                |              |                 |  |  |
| 13. | The supplier must have a system for remote proactive monitoring of technical   | Comply       | Do Not Comply   |  |  |
| 13. | parameters to detect deteriorating machine components and performance to ensure  | Comply       | Do Not Comply   |  |  |
|     | proactive interventions before a breakdown occurs that disrupts operations. This   |              |                 |  |  |
|     | should be in addition to the annual predicted and scheduled preventative   |              |                 |  |  |
|     | maintenance.   |              |                 |  |  |
|     | Substantiation: Provide proof using a brochure/catalogue.  |              |                 |  |  |
| 14. | The system should be capable of level sensing, clot detection and short sample   | Comply       | Do Not Comply   |  |  |
|     | (sample volume) detection.   |              |                 |  |  |
|     | Substantiation: Provide proof using a brochure/catalogue.  | <u> </u>     | Γ=              |  |  |
| 15. | Quantitative Spectrophotometric Detection of Haemolysis, Lipaemia and Icterus.   | Comply       | Do Not Comply   |  |  |
|     | Cubetantistian, Dunnida una et universitate a horachuna (antali  |              |                 |  |  |
|     | Substantiation: Provide proof using a brochure/catalogue.  |              |                 |  |  |



| 16. | The supplier should provide, support and troubleshoot all IQC including 3rd party IQC  | Comply    | Do Not Comply |
|-----|--|-----------|---------------|
|     | material with assigned target QC Ranges if required. The transfer of the IQC ranges  |           |               |
|     | from package inserts to each platform type should have  a) The facility is to be downloaded onto chemistry, coagulation and endocrine  |           |               |
|     | platforms using direct scanning of a coded insert/electronic download from an  |           |               |
|     | external source b) The facility is to be transferred to middleware directly from   |           |               |
|     | Chemistry, Coagulation and Endocrine platforms without the need for manual entry of  | of        |               |
|     | data.  |           |               |
|     | Substantiation: Provide proof using a brochure/catalogue as well as a screenshot.  |           |               |
| 17. | The noise level produced by the system must be below =<65db / If an air compressor   | Comply    | Do Not Comply |
|     | is used, it must be enclosed.  |           |               |
|     | Substantiation: Provide proof using a brochure/catalogue.  |           |               |
| 18. | Applicable hardware and software must be upgradeable or up scalable at the   | Comply    | Do Not Comply |
|     | supplier's cost in the event of new technologies, capabilities, changes in the test men  | u         |               |
|     | or changes in work volume.   |           |               |
|     | Substantiation: The supplier must provide details and a signed commitment letter.  |           |               |
|     |  |           |               |
| 19. | The supplier must supply all reagents, calibrators and quality control material for  | Comply    | Do Not        |
|     | the validation/verification studies until all analytes pass the acceptable criteria.   |           | Comply        |
|     | Onsite support should be provided for the duration of the verification.  |           |               |
|     |  |           |               |
|     |  |           |               |
|     | Substantiation: The bidder must submit a commitment letter.  |           | l .           |
|     |  |           |               |
| 20  | Tree of the control o |           |               |
| 20. | Liquid waste must be compliant to be directly plumbed to municipality lines i.e. it  | Comply    | Do Not        |
|     | must be suitable to be disposed of in this manner.   |           | Comply        |
|     | Substantiation: The bidder must submit and attach to the bid response, proof by n  | neans of  |               |
|     | brochure/specifications.   |           |               |
|     |  |           |               |
| 21. | Analyzer/System must be able to fit in the current designated space. If Cor  | nply      | Do Not Comply |
| 21. | any modifications are required, they must be at the supplier's cost (e.g.  | прту      | Do Not Comply |
|     | Renovation/Floor Reinforcement).   |           |               |
|     | Substantiation: The bidder must submit and attach to the bid response, proof by n  | neans of  |               |
|     | brochure/specifications.   |           |               |
|     |  |           |               |
| 22. | Supply initial and unlimited continual technical training of lab staff for the duration  | Comply    | Do Not        |
|     | of the contract. This includes appropriate testing (both written and witnessed)  |           | Comply        |
|     | immediately after training. Test/witnessing records and certificates are to be   |           |               |
|     | provided.  |           |               |
|     | Substantiation: The hidder must submit and attach to the hid response proof by   | noons of  |               |
|     | Substantiation: The bidder must submit and attach to the bid response, proof by n brochure/specifications. A commitment letter must be provided.   | icalis UI |               |
|     |  |           |               |
| 23. |  | Comply    | Do Not Comply |



|     | Availability of bi-directional communication with host query mode. Systems   |  |   |  |  |
|-----|--|--|---|--|--|
|     | must be interfaceable with the NHLS LIS. If it is a new development, this must   |  |   |  |  |
|     | be at the supplier's cost. Interface to be validated by the supplier during  |  |   |  |  |
|     | installation.  |  |   |  |  |
|     | Substantiation: The bidder must submit and attach to the bid response, proof by  |  |   |  |  |
|     | brochure/specifications as well as letter of commitment for interface validation.  |  |   |  |  |
|     |  |  |   |  |  |
| 24. | All items for instrument use e.g. consumables, reagents, equipment parts   | Comply                                     | Do Not                                      |  |  |
|     | including, printers, scanners and software upgrades/licenses should be   |  | Comply                                      |  |  |
|     | included in the bidding costs.   |  |   |  |  |
|     | Substantiation: The bidder must submit and attach to the bid response, proof by  | means of                                   |   |  |  |
|     | brochure/specifications. A commitment letter must be provided.   |  |   |  |  |
|     |  |  |   |  |  |
| 25. | Reagent or IQC material used by engineers/application specialist for   | Comply                                     | Do Not Comply                               |  |  |
|     | troubleshooting should be covered by the supplier.   |  |   |  |  |
|     | Substantiation: The bidder must submit and attach to the bid response, proof by  | means of                                   |   |  |  |
|     | brochure/specifications. A commitment letter must be provided.   |  |   |  |  |
|     |  |  |   |  |  |
| 26. | Reagents and calibrators should be able to load automatically while the  | Comply                                     | Do Not Comply                               |  |  |
|     | instruments are in operation ("on the fly").   |  |   |  |  |
|     | Substantiation: Provide proof by means of a brochure/ signed commitment letter. Failure to provide the   |  |   |  |  |
|     | · · · · · · · · · · · · · · · · · · ·  | er. Failure to pi                          | ovide the                                   |  |  |
|     | Substantiation: Provide proof by means of a brochure/ signed commitment letter information will lead to disqualification.  | er. Failure to pi                          | ovide the                                   |  |  |
| 27  | information will lead to disqualification.   |  |   |  |  |
| 27. | Information will lead to disqualification.  Test methods, calibrators and controls should be traceable to an international   | er. Failure to pi                          |   |  |  |
| 27. | Test methods, calibrators and controls should be traceable to an international standard.   |  |   |  |  |
| 27. | Information will lead to disqualification.  Test methods, calibrators and controls should be traceable to an international   |  |   |  |  |
| 27. | Test methods, calibrators and controls should be traceable to an international standard.   |  | Do Not Comply                               |  |  |
|     | Test methods, calibrators and controls should be traceable to an international standard.   |  |   |  |  |
|     | Test methods, calibrators and controls should be traceable to an international standard.  Substantiation: Provide evidence by means of a table.  | Comply                                     | Do Not Comply                               |  |  |
| 27. | Test methods, calibrators and controls should be traceable to an international standard.  Substantiation: Provide evidence by means of a table.  IQC and patient data must be retrievable in an accessible format independent  | Comply                                     | Do Not Comply                               |  |  |
|     | IQC and patient data must be retrievable in an accessible format independent of the analyser software (e.g. CSV) using a USB or Hard drive.  | Comply                                     | Do Not Comply                               |  |  |
|     | Information will lead to disqualification.  Test methods, calibrators and controls should be traceable to an international standard.  Substantiation: Provide evidence by means of a table.  IQC and patient data must be retrievable in an accessible format independent of the analyser software (e.g. CSV) using a USB or Hard drive.  Substantiation: The bidder must submit and attach to the bid response, proof by  | Comply                                     | Do Not Comply                               |  |  |
| 28. | Information will lead to disqualification.  Test methods, calibrators and controls should be traceable to an international standard.  Substantiation: Provide evidence by means of a table.  IQC and patient data must be retrievable in an accessible format independent of the analyser software (e.g. CSV) using a USB or Hard drive.  Substantiation: The bidder must submit and attach to the bid response, proof by  | Comply                                     | Do Not Comply  Do Not Comply                |  |  |
| 28. | Information will lead to disqualification.  Test methods, calibrators and controls should be traceable to an international standard.  Substantiation: Provide evidence by means of a table.  IQC and patient data must be retrievable in an accessible format independent of the analyser software (e.g. CSV) using a USB or Hard drive.  Substantiation: The bidder must submit and attach to the bid response, proof by brochure/specifications and evidence of extracted data.  | Comply  Comply  means of                   | Do Not Comply  Do Not Comply                |  |  |
|     | Test methods, calibrators and controls should be traceable to an international standard.  Substantiation: Provide evidence by means of a table.  IQC and patient data must be retrievable in an accessible format independent of the analyser software (e.g. CSV) using a USB or Hard drive.  Substantiation: The bidder must submit and attach to the bid response, proof by brochure/specifications and evidence of extracted data.  Current and previous LOT numbers of IQC data and L-J charts are to be available on the analyser or retrievable from storage.  Substantiation: The bidder must submit and attach to the bid response, proof by substantiation: The bidder must submit and attach to the bid response, proof by substantiation: The bidder must submit and attach to the bid response, proof by | Comply Comply means of Comply              | Do Not Comply  Do Not Comply                |  |  |
| 28. | Test methods, calibrators and controls should be traceable to an international standard.  Substantiation: Provide evidence by means of a table.  IQC and patient data must be retrievable in an accessible format independent of the analyser software (e.g. CSV) using a USB or Hard drive.  Substantiation: The bidder must submit and attach to the bid response, proof by brochure/specifications and evidence of extracted data.  Current and previous LOT numbers of IQC data and L-J charts are to be available on the analyser or retrievable from storage.  | Comply Comply means of Comply              | Do Not Comply  Do Not Comply                |  |  |
| 28. | IQC and patient data must be retrievable in an accessible format independent of the analyser software (e.g. CSV) using a USB or Hard drive.  Substantiation: The bidder must submit and attach to the bid response, proof by brochure/specifications and evidence of extracted data.  Current and previous LOT numbers of IQC data and L-J charts are to be available on the analyser or retrievable from storage.  Substantiation: The bidder must submit and attach to the bid response, proof by brochure/specifications and evidence of extracted data.  | Comply Comply means of Comply              | Do Not Comply  Do Not Comply                |  |  |
| 28. | Information will lead to disqualification.  Test methods, calibrators and controls should be traceable to an international standard.  Substantiation: Provide evidence by means of a table.  IQC and patient data must be retrievable in an accessible format independent of the analyser software (e.g. CSV) using a USB or Hard drive.  Substantiation: The bidder must submit and attach to the bid response, proof by brochure/specifications and evidence of extracted data.  Current and previous LOT numbers of IQC data and L-J charts are to be available on the analyser or retrievable from storage.  Substantiation: The bidder must submit and attach to the bid response, proof by brochure/specifications.  Analysers should be able to run offline if the LIS system is unavailable and allow        | Comply Comply means of Comply              | Do Not Comply  Do Not Comply  Do Not Comply |  |  |
| 28. | IQC and patient data must be retrievable in an accessible format independent of the analyser software (e.g. CSV) using a USB or Hard drive.  Substantiation: The bidder must submit and attach to the bid response, proof by brochure/specifications and evidence of extracted data.  Current and previous LOT numbers of IQC data and L-J charts are to be available on the analyser or retrievable from storage.  Substantiation: The bidder must submit and attach to the bid response, proof by brochure/specifications and evidence of extracted data.  | Comply  Comply  means of  Comply  means of | Do Not Comply  Do Not Comply  Do Not Comply |  |  |
| 28. | Information will lead to disqualification.  Test methods, calibrators and controls should be traceable to an international standard.  Substantiation: Provide evidence by means of a table.  IQC and patient data must be retrievable in an accessible format independent of the analyser software (e.g. CSV) using a USB or Hard drive.  Substantiation: The bidder must submit and attach to the bid response, proof by brochure/specifications and evidence of extracted data.  Current and previous LOT numbers of IQC data and L-J charts are to be available on the analyser or retrievable from storage.  Substantiation: The bidder must submit and attach to the bid response, proof by brochure/specifications.  Analysers should be able to run offline if the LIS system is unavailable and allow        | Comply Comply means of Comply means of     | Do Not Comply  Do Not Comply  Do Not Comply |  |  |



|     | Supplier-maintained 24-hour remote middleware backup facility in the event of  |                 |                     |
|-----|--|-----------------|---------------------|
|     | a server crash (Not Applicable to ESR).  |                 |                     |
|     | Substantiation: The bidder must submit and attach to the bid response, proof us commitment letter.   | sing a brochure | e/specifications or |
| 32. | Access to peer comparison data for IQC should be provided.   | Comply          | Do Not Comply       |
|     | Substantiation: Provide proof by means of a signed commitment letter. Failure t lead to disqualification.  | o provide the i | nformation will     |
| 33. | Prior analysers evaluation by NHLS HTA and/or peer-reviewed publications of the instrument's analytical performance evaluation. (Evaluations are at the Bidder's costs).       | Comply          | Do Not Comply       |
|     | Substantiate: Provide documentation/evidence (HTA certificate/letter from the NHLS evaluating) Provide HTA certificate and/or peer reviewed articles on instrument validation. |                 |                     |

Bidders who fail to comply with the Technical Mandatory Requirements WILL be disqualified.

#### **TECHNICAL FUNCTIONALITY:**

- 1.1 The bidder must complete in full of all the TECHNICAL FUNCTIONALITY requirements.
- 1.2 The bidder **must provide a unique reference number** (e.g. binder/folio, chapter, section, page) to locate substantiating evidence in the bid response. During evaluation, NHLS reserves the right to treat substantiation evidence that cannot be located in the bid response as "NOT COMPLY".

**Evaluation per requirement**. The evaluation (scoring) of bidders' responses to the requirements will be determined by the completeness, relevance and accuracy of substantiating evidence.

Each technical functionality requirement will be evaluated using on the following generic point scale

## **FUNCTIONAL EVALUATION CRITERIA**

| Functionality Requirements for<br>Chemical Pathology and Stat<br>Laboratory Analysers Section | SPECIFICATIONS                  | Weighting |
|---|---------------------------------|-----------|
| А   | Methods and Test Repertoire     | 10%       |
| В   | Service and Maintenance         | 20%       |
| С   | QC and Calibrators              | 13%       |
| D   | Reagents and Stock control      | 15%       |
| E   | IT/Data Recovery/Middleware     | 10%       |
| F   | Track (Sample Management,       | 32%       |
|   | Centrifugation, Sample storage) |           |
| Total   |                                 | 100%      |

Confidential Page **35** of **88** 



## TECHNICAL FUNCTIONALITY REQUIREMENTS FOR GENERAL CHEMISTRY ENDOCRINE AND VIROLOGY ANALYSERS.

The bidder must achieve a minimum score of 80% (Per discipline) to be eligible to proceed to the next stage of evaluation.

## **Functionality Requirements for Automated Chemical Pathology Analysers**

## Bidder must meet a minimum threshold of 80%.

| Weighting | Section   | Specification  | Score  |
|-----------|-----------|--|--|
| 10%       | Section A | Methods and Test Repertoire  | Score Rating   |
| 0%        |           | Provide a list (Tabulate) detailing the International Standards to which each test method is traceable and provide validation study data for each test on Chemistry, Endocrine, and Serology, platform types.  | No Score - for information only  |
| 6%        |           | 1. Detailing full test menu per analyser.  2. Indicate the methodology employed per test per analyser, e.g. MEIA for TSH on the immunoassay analyser.  3. Indicate which tests can be accommodated as user-defined methods (provide evidence of acceptable   | Score 1 point per evidence. [6 x 1= 6]  Score an additional 1 points for variable reagent pack sizes.          |
|           |           | <ul> <li>performance in an NHLS or any other lab).</li> <li>4. Include Cardiac Troponin T or I testing (indicate the 99 percentile cut-off value) with references.</li> <li>5. State the on-board stability period of the reagents test repertoire.</li> <li>6. Indicate whether platforms have sufficient reagent capacity to accommodate the lab's full test repertoire</li> </ul> |  |
| 1%        |           | and volume and standby reagent.  Must be able to recognise and retrieve spare samples for making aliquots if the primary sample is insufficient.   | Score 1: Able to recognise and retrieve spare samples. Score 0: Unable to recognise and retrieve spare samples |
| 1%        |           | State the reagent pack sizes that will accommodate the slow-moving tests indicated in the test repertoire and volume list.   | Score 1: Variable reagent pack sizes to cater for low volume tests.  Score 0: No variable reagent pack sizes.  |
| 1%        |           | ISE reagents, including wash, buffers and detergents should be sufficient to run uninterrupted for twenty-four hours according to the laboratory's current ISE workload.   | Score 1: Ability to run uninterrupted for > 24-hours. Score 0: Ability to run uninterrupted for < 24 hours     |

Confidential Page **36** of **88** 



| Weighting | Weighting Section Specification |   | Score  |
|-----------|---------------------------------|---|--|
| 1%        |                                 | Provide validation study data for each test on both Chemistry and Endocrine platform types.         | Score 1: Precision, interference, method comparison studies data provided.   |
|           |                                 |   | Score 0: Any one Precision, interference, method comparison studies data provided, or No validation data provided. |
| 20%       | Section B                       | Service and Maintenance   | Score Rating   |
| 8%        |                                 | Commitment to a total resolution time or alternatives   | Score 8: <2 hours' arrival time and  |
|           |                                 | should be provided to meet the TAT e.g.) external referrals to private                              | less than 6 hours' downtime.   |
|           |                                 | The total duration of downtime should not exceed 6  | Score 6: > 2 hours but < 4 hours'  |
|           |                                 | hours.  Bidders are to provide a commitment letter with details.                                    | arrival time and < 6 hours' downtime.  |
|           |                                 | details.  | Score 2: > 4 but <6 hours response   |
|           |                                 |   | and < 6 hours' downtime  |
|           |                                 |   |  |
|           |                                 |   | Score 0: Downtime >6 hours.  |
| 8%        |                                 | Describe the daily, weekly, and monthly maintenance   | Score 8: If daily <=30min, weekly  |
|           |                                 | procedures, including time taken.   | <=60 min and monthly < 2hr.  |
|           |                                 |   | Score 6: Do not meet any 1 of the  |
|           |                                 |   | above.   |
|           |                                 |   | Score 0: If do not meet any 2 of the   |
| 20/       |                                 |   | above.   |
| 2%        |                                 | Warm-up time after instrument shutdown.   | Score 2: < 15 minutes.   |
|           |                                 |   | <b>Score 1:</b> 15 – 20 minutes. <b>Score 0:</b> > 20 minutes.   |
| 2%        |                                 | Indicate if an integrated system/ monitoring tools are  | Score 2: Capability to monitor if  |
| 270       |                                 | available to monitor if adequate daily clean up and/or  | adequate daily instrument  |
|           |                                 | maintenance has been done.  | interventions have been done.  |
|           |                                 |   | Score 0: No capability to monitor if   |
|           |                                 |   | daily interventions are adequate.  |
| 13%       | Section C                       | QC and Calibrators  | Score Rating   |
| 3%        |                                 | Facility to scan a barcode or transfer Calibrator and QC  | Score 3: Barcode Scan and  |
|           |                                 | data electronically from an external source directly onto Chemistry and immunoassay platform types. | electronic transfer.   |
|           |                                 |   | Score 1: If > 1 platform types   |
| 40/       |                                 | Calibration and OC array datastics. floration and   | require manual entry.  |
| 4%        |                                 | Calibration and QC error detection, flagging, and automatic blocking of subsequent QC and patient   | Score 4: If available.   |
|           |                                 | testing.  | Score 0: No calibration error  |
|           |                                 | _   | detection and blocking of further  |
|           |                                 |   | analysis.  |
| 2%        |                                 | Peer group data: Access to third-party and supplier IQC   | Score 2: If the IQC peer group   |
|           |                                 | peer group database/management programme.   | database is available in real-time   |
|           |                                 |   | and documented evidence of a lab   |



| Weighting | Section   | Specification  | Score  |
|-----------|-----------|--|--|
|           |           | The provision of an adequate number of licences to support at least 15 Peer Group Programme user access points.  | using the named programme is available.  |
|           |           | Provide details of the workings of the recommended programme, as well as show evidence of a local lab using this programme and proof of the functionality  | Score 1: Programme available and evidence provided but not realtime.   |
|           |           | that can be expected.  | <b>Score 0:</b> If the programme is not available for recommended 3rd  |
|           |           | Timeous IQC troubleshooting assistance and conclusion in <24 hours.  | party IQC material.  |
| 4%        |           | Supply and support 3rd party IQC material with assigned QC Ranges.  The transfer of the IQC Ranges from Package inserts to each platform type should have the:  A. Facility to be downloaded onto Chemistry, Immunology and Endocrine platforms using direct scanning of a coded insert / electronic download from an external source. | Transfer IQC Range onto Analysers: Score 2: If direct scanning/transfer of QC data is available for each Chemistry, and the Endocrine platform type. Score 0: If manual lot number entry of QC data is available for each Chem, and Endo platform. |
|           |           | B. Facility to be transferred to Middleware directly from Chemistry and Endocrine - platforms / scanned from coded insert / downloaded from external electronic source without the need for manual entry of data/lot number.  (Indicate which platform this facility is not available).  | Transfer IQC Range onto Middleware: Score 2: If able to obtain QC data by scanning from insert / electronic download from an external source/transfer back from Chem, and Endo analyser to Middleware for each platform type. Score 0: If unable.  |
| 15%       | Section D | Reagents and Stock Control   | Score Rating   |
| 4%        |           | Provision of an electronic Reagent and Consumable stock monitoring system on the instrument. To provide all the details, i.e. level sensor of the volume of tests remaining and alert the operator via an alarm on low volume test.  | Score 4: Electronic System available.  Score 0: Electronic System not available.   |
| 2%        |           | On-board reagent stability monitoring with countdown alerts to on board expiry.  | Score 2: On-board stability reagent monitoring.  Score 0: No on-board reagent stability monitoring.  |
| 4%        |           | Facility for emergency stock delivery if the laboratory unexpectedly runs out of stock or has to take on extra work from other labs, including a contingency plan for after-hours or public holidays.  | Score 4: Emergency stock plan.  Score 0: No Plan.  |
| 2%        |           | Chem-Immuno-Sero test repertoire (see the full list of tests).  Haematology, Coagulation and ESR test repertoire (see the full list of tests).   | Score 2: Ready-for-use reagents without workflow interruption on both Chem and Immuno platforms. Ability to load two lots of the same reagent at the same time.>10 spare   |
|           |           | Indicate which tests require prior preparation and or mixing before loading.   | channels to accommodate UDR per analyser.  |

Confidential Page **38** of **88** 



| Weighting   | Section   | Specification  | Score  |
|-------------|-----------|--|--|
|             |           |  | Score 1: All available but less than                 |
|             |           | It should be able to load two different lots of the same   | 5% analytes require reagent                          |
|             |           | reagents simultaneously for lot-to-lot comparison  | mixing.  |
|             |           | purposes.  | Sang C. Natavailahla                                 |
|             |           | >=5 spare channels available to accommodate future   | Score 0: Not available.                              |
|             |           | user-defined reagents (UDR).   |  |
| 3%          |           | Allows continuous loading of reagents without the  | Score 3: If available.                               |
|             |           | downtime of the instrument. Loading of reagents  |  |
|             |           | without interruption of sample processing required.  | Score 0: No continuous loading on all modules.       |
| 10%         | Section E | IT/Data Recovery/Middleware  | Score Rating   |
| 2%          |           | Audit Trail: Ability to show operator name/code for all  | Score 2: Audit trail functionality                   |
|             |           | events, including the configuration of calibrator,   | available.   |
|             |           | controls, and reagents and be able to retain this  | Sang C. Na franchismality                            |
|             |           | information.   | Score 0: No functionality.                           |
|             |           | Middleware = 8%  |  |
| 2%          |           | User-defined reports for test count, QC's, calibrations,   | Score 2: All reports available.                      |
|             |           | patient tests, turnaround time, efficiencies and key performance indicator monitoring.             | Score 1: On user-defined KPI                         |
|             |           | performance maleutor morntoring.   | monitoring.  |
| 2%          |           | Real-time communication between middleware and   | Score 2: If functionality is present.                |
|             |           | the instrument. e.g. updating of QC ranges and error   |  |
|             |           | exclusions.  | Score 0: If not available                            |
| 4%          |           | Ability to calculate, report and monitor uncertainty of measurement.                               | <b>Score 4:</b> If functionality is present.         |
|             |           | measurement.   | Score 0: If not available.                           |
| 32%         | Section F | TRACK (Sample Management, Online Centrifugation,   | Score Rating   |
|             |           | Sample Storage).   |  |
| ==-/        |           | F (a) Sample Management = 22%  |  |
| 5%          |           | Automatic reintroduction of resealed/recapped samples from the Storage Unit onto the Track with no | <b>Score 5:</b> If no manual steps are required.     |
|             |           | manual intervention for repeat testing, reflex testing,  | required.  |
|             |           | or automated dilutions/decapping / unsealing   | Score 0: For any other response.                     |
|             |           | processes. (Provide details of this process).  |  |
| 4%          |           | The track can reroute samples in real time based on  | <b>Score 4:</b> If able to do it in real-time.       |
|             |           | delays/availability of tests on the analyser.  | Score 0: If requires manual                          |
|             |           |  | intervention.  |
| 2%          |           | Provision of an online sample mixer.   | Score 2: If the system has an online                 |
|             |           | · ·  | mixer  |
|             |           |  | Searc O. If the system decrees                       |
|             |           |  | Score 0: If the system does not have an online mixer |
| 5%          |           | Ability to track, monitor and regulate unprocessed   | Score 5: If the system can perform                   |
|             |           | samples transit and circulation time on the Track  | all requirements.                                    |
|             |           | (From entry to storage).   |  |
| <b>5</b> 0/ |           |  | Score 0: No Tracking capabilities.                   |
| 5%          |           | Provision of: Decapper with recapper / re-sealer   | Score 5: If all options are available.               |
|             |           | facility or Direct tube puncture/cap piercer.  |  |



| Weighting    | Section | Specification  | Score   |
|--------------|---------|--|---|
|              |         |  | Score 2: If either option is available.   |
|              |         |  | Score 0: If no recapping or resealing, or direct tube puncture is available.  |
| 1%           |         | Real time directing and blocking of samples to available analysers.  | Score 1: Real time directing of samples to analyser that is available to ensure timeous analyses.   |
|              |         |  | Score 0: Use of predefined destinations with no ability to change sample route without manual intervention.   |
|              | T       | F(b) On-Line Centrifugation = 3%   |   |
| 2%           |         | Indicate whether the centrifugation speed and duration of spin can be adjusted to accommodate the requirements for 3rd party Chemistry and Coagulation samples that will be placed on the Track.  Please state the Maximum centrifuge capacity/hour.                               | Score 3: If Centrifuge speed and duration can be adjusted and capacity adequate for our volumes.  Score 2: If Centrifuge speed and duration can be adjusted and no capacity details are provided.   |
|              |         |  | Score 0: If speed and duration are fixed.   |
|              |         | F (c) Sample Storage Unit / Module / Stockyard =   | 7%  |
| 2%           |         | An online refrigerated storage facility is required.  Facility to store minimum 7 days' total work volume (see volume print outs) in 1 or more storage units/ modules (Provide details of the number of samples that can be stored per unit as well as the foot-print per module). | Score 2: Refrigerator space can fulfil laboratory needs for Chemistry and Endocrine samples for stipulated time frames.  Score 1: If refrigerator capacity can meet either Chemistry or Endocrine requirements.  Score 0: If a storage facility is not available. |
| 4%           |         | Ability to manually store samples in the storage module e.g.) samples run off the track, low volume samples, 3rd party analysers.  | Score 4: If capable. Score 0: Is not available  |
| 1%           |         | Ability to manually retrieve samples from the storage unit if needed.  | Score 1: If present.  |
| Total = 100% |         |  | Score 0: If not present.  Total = 100   |

Page **40** of **88** 



# **Functionality Requirements for Automated Stand Alone ESR Analyser**

# Bidder must meet a minimum threshold of 80%.

| Section A: Sample Management  | 35%   | Scoring Rate  |
|---|-------|---|
| The analyser to operate without requiring the user to be present throughout (walkaway functionality).   | 20%   | Score 20: Capable. Score 0: Not capable.  |
| (Provide proof by means of specification/brochure).   |       |   |
| Instrument must have a hand-held sample barcode reader for manual and stat samples.  (Provide proof by means of specification/brochure).  | 15%   | Score 15: Available. Score 0: Not available.  |
| Section B: Service and Maintenance  | 20%   | Scoring Rate  |
| Daily, Weekly and Monthly Maintenance Average time spent by staff in daily maintenance - (< 15min). Total down time for maintenance should not exceed 30 minutes.   | 20%   | Score 20: If daily <=15min, weekly <=30min and monthly <=30min.  Score 10: Do not meet any 1 of the above.  Score 0: If do not meet |
| (Provide proof by means of specification/brochure).   |       | any 2 of the above.   |
| Section C: QC and Calibrators   | 15%   | Scoring Rate  |
| The ability to scan a barcode or transfer Calibrator/ QC data electronically from an external source directly onto or download the data to the analyser.  (Provide proof by means of specification/brochure).                             | 15%   | Score 15: Available. Score 0: Not available.  |
| Section D: Reagents and Stock Control   | 30%   | Scoring Rate  |
| Provision of an electronic Reagent and Consumable stock monitoring system on the instrument to monitor reagent levels, lot number expiry dates and user traceability, (Audit trail).  (Provide proof by means of specification/brochure). | 15%   | Score 15: Available. Score 0: Not available.  |
| Reagents management mode (efficiency report) to minimise wastage.   | 5%    | Score 5: Available.   |
| (Provide proof by means of specification/brochure).   |       | Score 0: Not available.   |
| Reagent tracking available on analyser with printable and downloadable option.  | 10%   | Score 10: Available.  |
| (Provide proof by means of specification/brochure).   | 4000/ | Score 0: Not available.   |
| Total   | 100%  |   |

Page 41 of 88



# Functionality Requirements for two (2) Automated Coagulation Analysers

# Bidder must meet a minimum threshold of 80%.

| Section A: Sample Management   | 20%  | Scoring Rate                      |
|--|------|-----------------------------------|
| Ability to interrupt processing of non-urgent samples to run 2-5 stat specimens.                                       | 10%  | Score 10: Available. Score 0: Not |
| (Provide proof by means of specification/brochure).  | 100/ | available.                        |
| Capacity to accommodate all reagents required to accommodate our test repertoire on board.                             | 10%  | Score 10: Available. Score 0: Not |
| (Provide proof by means of specification/brochure).  |      | available.                        |
| Section B: Service and Maintenance   | 20%  | Scoring Rate                      |
| Daily, Weekly and Monthly Maintenance  | 20%  | Score 20: If daily                |
| Average time spent by staff in daily maintenance - (< 15min). Total down time for                                      |      | <=15min, weekly                   |
| maintenance should not exceed 30 minutes.  |      | <=30min and                       |
|  |      | monthly <=30min.                  |
|  |      | Score 10: Do not                  |
|  |      | meet any 1 of the                 |
|  |      | above.                            |
|  |      | Score 0: If do not                |
|  |      | meet any 2 of the                 |
| (Provide proof by means of specification/brochure).  |      | above.                            |
| Section C: QC and Calibrators  | 20%  | Scoring Rate                      |
| The analyser to store >2 calibrations curves per test parameter.   | 20%  | Score 20: Available.              |
|  |      | Score 0: Not                      |
| (Provide proof by means of specification/brochure).  |      | available.                        |
| Section D: Reagents and Stock Control  | 30%  | Scoring Rate                      |
| Provision of an electronic Reagent and Consumable stock monitoring system on the                                       | 10%  | Score 10: Available.              |
| instrument to monitor reagent levels, lot number expiry dates and user traceability,                                   |      | Score 0: Not                      |
| (Audit trail).   |      | available.                        |
| (Provide proof by means of specification/brochure).  | F0/  | Coons E. Austlahla                |
| Reagents management mode (efficiency report) to minimise wastage.  (Provide proof by means of specification/brochure). | 5%   | Score 5: Available. Score 0: Not  |
| (Provide proof by means of specification/brochure).  |      | available.                        |
|  |      | available.                        |
| Reagent tracking available on analyser with printable and downloadable option.   | 10%  | Score 10: Available.              |
|  |      | Score 0: Not                      |
| (Provide proof by means of specification/brochure).  |      | available.                        |
| Haematology, Coagulation and ESR test repertoire (see the full list of tests).   | 5%   | Score 5: Ability to               |
| macmatology, coagulation and Lon test repertolle (see the full list of tests).   | 3/0  | load two lots of the              |
| Indicate which tests require prior preparation and or mixing before loading.   |      | same reagent at the               |
|  |      | same time. Spare                  |
| It should be able to load two different lots of the same reagents simultaneously for lot-to-                           |      | channels to                       |
| lot comparison purposes.   |      | accommodate UDR                   |
|  |      | per analyser.                     |
| Spare channels available to accommodate future user-defined reagents (UDR).  |      |                                   |



|   |      | Score 0: Not available                             |
|---|------|--|
| Section E: IT Middleware  | 10%  |  |
| Provision of a middleware programme, including tablets (with applicable software) for offsite monitoring/reviewing. In addition, the provision of an adequate number of licences to support 3 middleware access points.  (Provide proof by means of specification/brochure as well as commitment letter). | 10%  | Score 10: Available.<br>Score 0: Not<br>available. |
| Total   | 100% |  |

# **Functionality Requirements for Automated Virology (Serology) Analysers**

# Bidder must meet a minimum threshold of 80%.

| Section A: Methods and Test  | 14% | Scoring Rate                         |
|--|-----|--------------------------------------|
| Tabulate the following points of evidence:                                 | 14% | Score 12: If all 6 points available. |
| 1. Detailing full test menu per analyser and stating whether all the tests |     | Score10: If 5 points                 |
| requested in this tender are included.                                     |     | available.                           |
| 2. Indicate which tests can be accommodated as user-defined methods        |     | Score 8: If 4 points                 |
| (provide evidence of acceptable performance in an NHLS or any other        |     | available.                           |
| lab).  |     | Score 6: If 3 points                 |
| 3. Indicate the methodology employed per test per analyser, e.g. MEIA      |     | available.                           |
| 4. State the on-board stability period of the reagents test repertoire.    |     | Score 4: If 1 point                  |
| 5. Indicate whether platforms have sufficient reagent capacity to          |     | available.                           |
| accommodate the lab's full test repertoire and volume and standby          |     | Score 2: If 1 point                  |
| reagent.   |     | available or none is                 |
| 6. State the reagent pack sizes that will accommodate the slow-moving      |     | available.                           |
| tests indicated in the test repertoire and volume list.                    |     |                                      |
| Section B: Service and Maintenance   | 20% |                                      |
| Provision of 1 dedicated Instrument technician / engineers and 1           | 10% | Score 8: If all available.           |
| Application Specialist (AS) available for the Gauteng region to provide an |     | Score 4: If only one of the          |
| uninterrupted service with back up. The same team cannot be shared         |     | two is available.                    |
| with other provinces.  |     | Score 0: If not available.           |
| (Provide a commitment letter indicating allocation).                       |     |                                      |
| < 2 hour Response time (arrival of technician / engineer on site) from     | 6%  | Score 6: If available <2hrs.         |
| time of call log for analyser breakdown / malfunction.                     |     | Score 4: If available                |
|  |     | between 2-4hrs.                      |
|  |     | Score 2: If available                |
|  |     | between 4-6.                         |
|  |     | Score 0: If >6rs.                    |
| Tabulate the full daily, full weekly, and full monthly maintenance         | 4%  | Score 4: If daily≤ 30 min,           |
| procedures, including time taken.  |     | weekly≤ 60 min and                   |
|  |     | monthly ≤3hr.                        |
|  |     | Score 0: Do not meet                 |
|  |     | above criteria.                      |
|  |     |                                      |
|  |     |                                      |
|  |     |                                      |
|  |     |                                      |

Confidential Page **43** of **88** 



| Section C: QC and Calibrators   | 18% |  |
|---|-----|--|
| RFID Reagents and calibrators.  | 6%  | Score 5: RFID reagents and calibrators. Score 0: If nothing. Score 3: If either reagents or calibrators have RFID functionality. Score 0: If not provided.           |
| Calibration error detection and flagging.   | 6%  | Score 5: If available. Score 0: No calibration error detection and flagging.   |
| Ability to block patient testing due to calibrator or IQC failure.  | 6%  | Score 5: Functionality available for both calibrator and IQC failure. Score 3: Functionality available for either calibrator or IQC failure. Score 0: Not available. |
| Section D: Reagents and Stock Control   | 14% |  |
| Provision of an electronic Reagent and Consumable stock monitoring system on the instrument. To provide all the details, i.e. level sensor of the volume of tests remaining and alert the operator via an alarm on low volume test. | 4%  | Score 3: Electronic System available. Score 0: Electronic if system not available or suboptimal.   |
| The inventory Management system that can be interphased with the NHLS procurement system support/assist the laboratory stock handling, including timeous backorder, reports on delivery, stock-outs, short deliveries.              | 3%  | Score 2: Available. Score 0: Not available.  |
| Facility for emergency stock delivery if the laboratory unexpectedly runs out of stock or has to take on extra work from other labs, including a contingency plan for after-hours or public holidays.                               | 2%  | Score 2: Emergency stock plan. Score 0: No Plan.   |
| Allows continuous loading of reagents without the downtime of the instrument. Loading of reagents without interruption of sample processing required.   | 2%  | Score 2: If available. Score 0: No continuous loading on all modules.  |
| Ready to load reagents: Tabulate as per test repertoire provided.   | 3%  | score 3: If all reagents are ready to load. Score 2: If half of the reagents are ready to load. Score 0: If less than half of reagents are ready to load.            |
| Section E: IT / Middleware  | 10% |  |
| Audit Trail: Ability to show operators name/code for all events, including the configuration of calibrator, controls, and reagents and be able to retain this information.  | 5%  | Score 3: Audit trail functionality available. Score 0: No functionality.   |
| Ability to download audit trail information into a CSV file or similar.   | 5%  | Score 3: Available. Score 0: Not available.  |



| F (a) Sample Management = 24%   | 24%  |   |
|---|------|---|
| Automatic reintroduction of resealed/recapped samples from the Storage Unit onto the Track with no manual intervention for repeat testing, reflex testing, or automated dilutions/decapping / unsealing processes. (Provide details of this process). | 5%   | Score 5: If no manual steps are required.  Score 0: For any other |
|   |      | response.   |
| The track can reroute samples in real time based on delays/availability of tests on the analyser.   | 5%   | <b>Score 5:</b> If able to do it in real-time.                    |
|   |      | Score 0: If requires manual intervention.                         |
| Provision of an online sample mixer.  | 4%   | Score 2: If the system has an online mixer.                       |
|   |      | Score 0: If the system does not have an online mixer.             |
| Ability to track, monitor and regulate unprocessed samples transit and circulation time on the Track (From entry to storage).   | 5%   | Score 5: If the system can perform all requirements.              |
|   |      | <b>Score 0:</b> No Tracking capabilities.                         |
| Provision of: Decapper with recapper / re-sealer facility or Direct tube puncture/cap piercer.  | 5%   | Score 5: If all options are available.                            |
|   |      | <b>Score 2:</b> If either option is available.                    |
|   |      | <b>Score 0:</b> If no recapping or resealing, or direct tube      |
| Total   | 100% | puncture is available.  |

# **ANNEXURE B: Pricing Schedule**

| Plea | se indicate your total bid price here: R(inclu                                       | sive of all ap | plicable taxes, e.g. |  |  |  |
|------|--|----------------|----------------------|--|--|--|
| VAT  | 7)   |                |                      |  |  |  |
| mp   | ortant:  |                |                      |  |  |  |
| t is | mandatory to indicate your total bid price as requested above. This price m          | ust be the sa  | me as the total bid  |  |  |  |
| oric | e you submit in your pricing schedule. Should the total bid prices differ, the       | total bid prid | ce indicated above   |  |  |  |
| shal | ll be considered the correct price.  |                |                      |  |  |  |
| Γhe  | following must be noted:   |                |                      |  |  |  |
| 1.   | All prices must be VAT inclusive of all applicable taxes and must be quoted in       | n South Africa | ın Rand (ZAR).       |  |  |  |
| 2.   | All prices must be firm and fixed from the tender closing date and for the du        | ration of the  | contract             |  |  |  |
| 3.   | All the consortium or joint venture partners must submit a complete set              | of the lates   | t audited financial  |  |  |  |
|      | statements.  |                |                      |  |  |  |
| 4.   | All bidders must cost according to the costing template provided or this will        | lead to disqu  | alification.         |  |  |  |
|      |  |                |                      |  |  |  |
| 5.   |  |                |                      |  |  |  |
|      | e cost of installation, delivery, site preparation etc. Must be included in this     | Comply         | Do Not comply        |  |  |  |
| þr   | oposal.  |                |                      |  |  |  |
| Su   | bstantiate / Comments.   |                |                      |  |  |  |
|      |  |                |                      |  |  |  |
| 6.   |  |                |                      |  |  |  |
| No   | price adjustments that are 100% linked to exchange rate variations shall be          | Comply         | Do Not comply        |  |  |  |
| all  | owed.  |                |                      |  |  |  |
| Su   | bstantiate / Comments.   |                | - <b>L</b>           |  |  |  |
|      |  |                |                      |  |  |  |
| 7.   |  |                |                      |  |  |  |
| Th   | e bidder must indicate <b>clearly</b> which portion of the purchase price as well as | Comply         | Do Not comply        |  |  |  |
| the  | the monthly costs is linked to the exchange rate.                                    |                |                      |  |  |  |
| Su   | bstantiate / Comments.   | 1              | ,                    |  |  |  |
|      |  |                |                      |  |  |  |
| 8.   |  |                |                      |  |  |  |
| ΔII  | All additional costs must be clearly specified.                                      |                |                      |  |  |  |
|      | , ·  |                |                      |  |  |  |
| Su   | bstantiate / Comments.   |                |                      |  |  |  |

Confidential Page **46** of **88** 





# PRICING SCHEDULE – FIRM PRICES (PURCHASES)

NOTE: ONLY FIRM PRICES WILL BE ACCEPTED. NON-FIRM PRICES (INCLUDING PRICES SUBJECT TO RATES OF EXCHANGE VARIATIONS) WILL NOT BE CONSIDERED.

IN CASES WHERE DIFFERENT DELIVERY POINTS INFLUENCE THE PRICING, A SEPARATE PRICING SCHEDULE MUST BE SUBMITTED FOR EACH DELIVERY POINT.

| Nam    | Name of bidder:  |                               |  |  |  |  |
|--------|--|-------------------------------|--|--|--|--|
| Bid n  | Bid number: RFB153/25/26 Closing Time 11:00 am Closing date: 03 February 2026                        |                               |  |  |  |  |
| Bid F  | Price (Vat incl.) R  |                               |  |  |  |  |
|        |  | DAYS FROM THE CLOSING D       |  |  |  |  |
| ITEM   | QUANTITY   | DESCRIPTION                   | BID PRICE IN RSA CURRENCY                            |  |  |  |
| NO.    |  |                               | ** (ALL APPLICABLE TAXES INCLUDED)                   |  |  |  |
| -      | Required by:   |                               |  |  |  |  |
| -      | At:  |                               |  |  |  |  |
|        |  |                               |  |  |  |  |
| -      | Brand and model  |                               |  |  |  |  |
| -      | Country of origin  |                               | <del></del>  |  |  |  |
| _      | Does the offer compl   | y with the specification(s)?  | *YES/NO  |  |  |  |
| -      | If not to specification  | , indicate deviation(s)       |  |  |  |  |
|        | 5 . 1 . 16 .   |                               |  |  |  |  |
| -      | Period required for d  | elivery                       | *Delivery: Firm/not firm                             |  |  |  |
| -      | Delivery basis   |                               |  |  |  |  |
| Note:  | e: All delivery costs must be included in the bid price, for delivery at the prescribed destination. |                               |  |  |  |  |
| ** "al | l applicable taxes" inc  | ludes value- added tax, pay a | as you earn, income tax, unemployment insurance fund |  |  |  |

\*\* "all applicable taxes" includes value- added tax, pay as you earn, income tax, unemployment insurance fund contributions and skills development levies.



#### PRICE DECLARATION FORM-

#### Dear Madam /Sir,

Having read through and examined the Tender Document, RFB NO:153/25/26, General Conditions, the requirement and all other Annexures to the Tender Document, we offer to provide Placement of Total Automation System for Pre-Analytics, Analytic and Post-Analytical Systems for Chemical Pathology/STAT Laboratory for National Health Laboratory Service DGM including reagents, consumables & service contract for a period of five (5) years as detailed in the bid document, for the total Tendered Contract Sum of in:

|   | (VAT Incl.) Amount in Words   |
|---|-------------------------------|
|   |                               |
| 3 | (VAT Incl.) Amount in Numbers |

We confirm that this price covers all activities associated with RFB153/25/26: Placement of Total Automation System for Pre-Analytics, Analytic and Post-Analytical Systems for Chemical Pathology/STAT Laboratory for National Health Laboratory Service DGM including reagents, consumables & service contract for a period of five (5) years. We confirm that NHLS will incur no additional costs whatsoever over and above this amount in connection with the supply of this solution.

We further confirm that all licences required for complete implementation of the solution, and the costs associated therewith, as well as any licences that may be required for future expansion have been fully described and disclosed in this document.

We undertake to hold this offer open for acceptance for a period of **180 days** from the date of submission of offers. We further undertake that upon final acceptance of our offer; we will commence with delivery when required to do so by the Client.

Moreover, we agree that until formal Contract Documents have been prepared and executed, this Form of Tender, together with a written acceptance from the Client shall constitute a binding agreement between us, governed by the terms and conditions set out in this Request for Proposals.

We understand that you are not bound to accept the lowest or any offer and that we must bear all costs which we have incurred in connection with preparing and submitting this tender.

We hereby undertake for the period during which this tender remains open for acceptance not to divulge to any persons, other than the persons to which the tender is submitted, any information relating to the submission of this tender or the details therein except where such is necessary for the submission of this tender.

Confidential

Page 48 of 88

PER Number 152/25/26: Placement of Total Automation System for Pro Analytics Analytic and Post Analytical Systems for Chamical



| SIGNED:                  |              | DATE: |
|--------------------------|--------------|-------|
|                          |              |       |
|                          |              |       |
| Print name of signatory) |              |       |
|                          |              |       |
|                          |              |       |
| Designation              |              |       |
|                          |              |       |
| FOR AND ON BEHALF OF:    | COMPANY NAME |       |
|                          | Tel No       |       |
|                          | Fax No       |       |
|                          | Cell No      |       |



Bidders *must* provide the NHLS with costing information for a 5 years' contract duration. The bid price quoted must be inclusive as per the scope of work.

#### Note:

- a) Bidder must complete the pricing as per tables below.
- b) Prices must be provided in South African Rand (R).
- c) Line Prices are all VAT EXCLUSIVE and TOTAL PRICE is VAT INCLUSIVE.
- d) Bidder to ensure that the Prices listed below are included on the Total Declared Price.
- e) Bidders who fail to price according to the costing template provided will be disqualified.



**Costing Table: DGM Laboratory** 

| PLACEMENT   | Quantity | Monthly<br>Cost in Year<br>1<br>(VAT Excl.) | Annual Cost<br>Year 1<br>(VAT Excl.) | Monthly Cost<br>in Year 2<br>(VAT Excl.) | Annual<br>Cost<br>Year 2<br>(VAT Excl.) | Monthly Cost in Year 3 (VAT Excl.) | Annual<br>Cost<br>Year 3<br>(VAT<br>Excl.) | Monthly<br>Cost in<br>Year 4<br>(VAT<br>Excl.) | Annual<br>Cost<br>Year 4<br>(VAT<br>Excl.) | Monthly<br>Cost in<br>Year 5<br>(VAT<br>Excl.) | Annual<br>Cost<br>Year 5<br>(VAT<br>Excl.) | Total Annual<br>Cost Year 1 to<br>5<br>(VAT Excl.) |
|---|----------|---|--------------------------------------|--|---|------------------------------------|--|--|--|--|--|--|
| Placement Fee   |          | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |
| Kit/Reagents  |          | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |
| Test Consumables                                      |          | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |
| Controls  |          | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |
| Calibrators   |          | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |
| Service and Maintenance<br>Costs                      | 1        | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |
| Software/ IT maintenance                              |          | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |
| Consumables Needed During<br>Preventative Maintenance |          | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |
| Insurance   |          | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |
| Training  |          | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |
| Subtotal (VAT Excl.)                                  |          | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |
| VAT (15%)   |          | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |
| Total Price (VAT Incl.)                               |          | R   | R                                    | R  | R                                       | R                                  | R  | R  | R  | R  | R  | R  |

Confidentia

Page **51** of **88** 



| PLACEMENT | Quantity | Monthly<br>Cost in Year<br>1<br>(VAT Excl.) | Annual Cost<br>Year 1<br>(VAT Excl.) | Monthly Cost<br>in Year 2<br>(VAT Excl.) | Annual<br>Cost<br>Year 2<br>(VAT Excl.) | Monthly<br>Cost in<br>Year 3<br>(VAT<br>Excl.) | Annual<br>Cost<br>Year 3<br>(VAT<br>Excl.) | Monthly<br>Cost in<br>Year 4<br>(VAT<br>Excl.) | Annual<br>Cost<br>Year 4<br>(VAT<br>Excl.) | Monthly<br>Cost in<br>Year 5<br>(VAT<br>Excl.) | Annual<br>Cost<br>Year 5<br>(VAT<br>Excl.) | Total Annual<br>Cost Year 1 to<br>5<br>(VAT Excl.) |
|-----------|----------|---|--------------------------------------|--|---|--|--|--|--|--|--|--|
|           |          |   |                                      |  |   |  |  |  |  |  |  |  |

Please indicate the summary cost per test for the following items: -

| Item             | Cost per Test | Monthly Cost (Rand ) |
|------------------|---------------|----------------------|
| Test Consumables |               |                      |
| Controls         |               |                      |
| Calibrators      |               |                      |

Training

| Description | Total cost Vat Excl. | Total cost Vat Incl. |
|-------------|----------------------|----------------------|
|             |                      |                      |

Please any additional comments in the box below to further clarify any details about the all-in cost per test for your assay: -

List content of reagent kit for consumables (is column for analysis included as consumables in reagent kit).



Please provide a detailed bill of materials for the assays included in the proposal specifications per NHLS laboratory:

| Test | Test Volumes per month | Test per kit | Unit Cost | Cost per billable |
|------|------------------------|--------------|-----------|-------------------|
|      |                        |              |           |                   |
|      |                        |              |           |                   |
|      |                        |              |           |                   |
|      |                        |              |           |                   |
|      |                        |              |           |                   |
|      |                        |              |           |                   |
|      |                        |              |           |                   |
|      |                        |              |           |                   |
|      |                        |              |           |                   |
|      |                        |              |           |                   |
|      |                        |              |           |                   |
|      |                        |              |           |                   |
|      |                        |              |           |                   |



**Table 1: DGM Chemical Pathology Laboratory Volumes:** 

| Depa                                  | Department of Chemical Pathology |        |        |        |        |        |        |  |  |  |  |  |
|---------------------------------------|----------------------------------|--------|--------|--------|--------|--------|--------|--|--|--|--|--|
| Test                                  | FY23_24<br>Test                  | Year 1 | Year 2 | Year 3 | Year 4 | Year 5 | Total  |  |  |  |  |  |
| 17-Alpha-Hydroxy-Progesterone         | 31                               | 33     | 35     | 37     | 39     | 41     | 185    |  |  |  |  |  |
| Acetyl choline Receptor Antibody      | 28                               | 30     | 32     | 34     | 36     | 38     | 170    |  |  |  |  |  |
| Adenosine Deaminase                   | 2276                             | 2278   | 2280   | 2282   | 2284   | 2286   | 11410  |  |  |  |  |  |
| Adenosine Deaminase - CSF             | 89                               | 99     | 109    | 119    | 129    | 139    | 595    |  |  |  |  |  |
| Adenosine Deaminase - Fluid           | 638                              | 640    | 642    | 644    | 646    | 648    | 3220   |  |  |  |  |  |
| Adreno-Corticotrophic Hormone (ACTH   | 2                                | 4      | 6      | 8      | 10     | 12     | 40     |  |  |  |  |  |
| Alanine Transaminase                  | 68185                            | 68191  | 68196  | 68202  | 68207  | 68213  | 341008 |  |  |  |  |  |
| Albumin - Serum                       | 48267                            | 48285  | 48303  | 48321  | 48339  | 48357  | 241605 |  |  |  |  |  |
| Alcohol (In Blood )                   | 9                                | 11     | 13     | 15     | 17     | 19     | 75     |  |  |  |  |  |
| Alpha 1 Antitrypsin                   | 30                               | 32     | 34     | 36     | 38     | 40     | 180    |  |  |  |  |  |
| Alpha Feto-Protein (AFP)              | 7864                             | 7879   | 7894   | 7909   | 7924   | 7939   | 39545  |  |  |  |  |  |
| Amikacin                              | 50                               | 52     | 54     | 56     | 58     | 60     | 280    |  |  |  |  |  |
| Ammonia                               | 220                              | 235    | 250    | 265    | 280    | 295    | 1325   |  |  |  |  |  |
| Amylase -Fluid                        | 399                              | 414    | 429    | 444    | 459    | 474    | 2220   |  |  |  |  |  |
| Amylase Total - Serum                 | 4937                             | 4942   | 4947   | 4952   | 4957   | 4962   | 24760  |  |  |  |  |  |
| Angiotensin Convert Enzyme (ACE)      | 14273                            | 14283  | 14293  | 14303  | 14313  | 14323  | 71515  |  |  |  |  |  |
| Anti-Mullerian Hormone                | 126                              | 131    | 136    | 141    | 146    | 151    | 705    |  |  |  |  |  |
| Anti-TG                               | 240                              | 242    | 244    | 246    | 248    | 250    | 1230   |  |  |  |  |  |
| Antithyroidperoxidase Antibodiesb(Ant | 399                              | 409    | 419    | 429    | 439    | 449    | 2145   |  |  |  |  |  |
| Aspartate Transaminase                | 38266                            | 38281  | 38296  | 38311  | 38326  | 38341  | 191555 |  |  |  |  |  |
| Barbiturates                          | 0                                | 2      | 4      | 6      | 8      | 10     | 30     |  |  |  |  |  |
| Benzodiazepines                       | 5                                | 7      | 9      | 11     | 13     | 15     | 55     |  |  |  |  |  |
| Beta-2 Microglobulin                  | 720                              | 727    | 734    | 741    | 748    | 755    | 3706   |  |  |  |  |  |
| BHCG (human chorionic gonadotropin)   | 8839                             | 8852   | 8864   | 8877   | 8889   | 8902   | 44383  |  |  |  |  |  |
| Bilirubin Direct                      | 39864                            | 39880  | 39895  | 39911  | 39926  | 39942  | 199553 |  |  |  |  |  |
| Bilirubin Total                       | 59228                            | 59238  | 59247  | 59257  | 59266  | 59276  | 296284 |  |  |  |  |  |

Confidential Page **54** of **88** 



| Ca-125 Tumour Marker                   | 3626   | 3641   | 3656   | 3671   | 3686   | 3701   | 18355   |
|--|--------|--------|--------|--------|--------|--------|---------|
| Ca-15.3 Tumour Marker                  | 20     | 25     | 30     | 35     | 40     | 45     | 175     |
| Ca-19.9 Tumour Marker                  | 4257   | 4272   | 4287   | 4302   | 4317   | 4332   | 21510   |
| Caeruloplasmin                         | 42     | 44     | 47     | 49     | 52     | 54     | 246     |
| Calcium - Serum                        | 68040  | 68052  | 68064  | 68076  | 68089  | 68101  | 340382  |
| Calcium - Urine                        | 24     | 26     | 28     | 30     | 32     | 34     | 150     |
| Carbamazepine Level                    | 931    | 937    | 943    | 948    | 954    | 960    | 4742    |
| Carcino Embryonic Antigen (CEA)        | 7503   | 7518   | 7533   | 7548   | 7563   | 7578   | 37740   |
| Chloride                               | 139247 | 139257 | 139267 | 139277 | 139287 | 139297 | 696385  |
| Chloride - Fluid                       | 869    | 876    | 883    | 890    | 897    | 904    | 4450    |
| Cholesterol HDL                        | 24810  | 24825  | 24840  | 24855  | 24870  | 24885  | 124275  |
| Cholesterol Total                      | 86915  | 86930  | 86945  | 86960  | 86975  | 86990  | 434800  |
| Cholestrol LDL - Measured              | 42856  | 42871  | 42886  | 42901  | 42916  | 42931  | 214505  |
| Cholinesterase Total                   | 696    | 709    | 722    | 736    | 749    | 762    | 3678    |
| Chromogranin A                         | 19     | 21     | 23     | 25     | 27     | 29     | 125     |
| CO2 Total                              | 139247 | 139257 | 139267 | 139277 | 139287 | 139297 | 696385  |
| CO2 Content -Fluid                     | 943    | 958    | 973    | 988    | 1003   | 1018   | 4940    |
| Cortisol                               | 1584   | 1599   | 1614   | 1629   | 1644   | 1659   | 8145    |
| C-Peptide                              | 394    | 409    | 424    | 439    | 454    | 469    | 2195    |
| C-Reactive Protein(wide range & High S | 75468  | 75477  | 75486  | 75496  | 75505  | 75514  | 377479  |
| Creatine Kinase                        | 6914   | 6919   | 6924   | 6929   | 6934   | 6939   | 34645   |
| Creatine Kinase MB                     | 11209  | 11212  | 11214  | 11217  | 11220  | 11222  | 56085   |
| Creatinine                             | 303656 | 303666 | 303676 | 303685 | 303695 | 303705 | 1518427 |
| Creatinine - Urine                     | 4790   | 4799   | 4808   | 4817   | 4825   | 4834   | 24083   |
| Digoxin Level                          | 106    | 111    | 116    | 121    | 126    | 131    | 605     |
| Dihydroepianrostene Sulphate (DHEAS    | 22     | 27     | 32     | 37     | 42     | 47     | 185     |
| Follicle-Stimulating Hormone (FSH)     | 3164   | 3179   | 3194   | 3209   | 3224   | 3239   | 16045   |
| Triiodothyronine Free (FT3)            | 28391  | 28406  | 28421  | 28436  | 28451  | 28466  | 142180  |
| Free light chains                      | 920    | 928    | 935    | 943    | 950    | 958    | 4713    |
| Free PSA                               | 3745   | 3753   | 3760   | 3768   | 3775   | 3783   | 18838   |
| Fructosamine                           | 90     | 95     | 100    | 105    | 110    | 115    | 525     |
| Gentamicin                             | 50     | 52     | 54     | 56     | 58     | 60     | 280     |
| Glucose - Plasma                       | 10252  | 10257  | 10262  | 10267  | 10272  | 10277  | 51335   |
| Glucose -CSF                           | 167    | 177    | 187    | 197    | 207    | 217    | 985     |
| Glucose Fluid & Urine                  | 1915   | 1916   | 1917   | 1918   | 1919   | 1920   | 9590    |
|  |        |        |        |        |        |        |         |

Confidential Page 55 of 88



| Glycated Haemoglobin (HbA1c)        | 80058  | 80073  | 80088  | 80103  | 80118  | 80133  | 400515 |
|-------------------------------------|--------|--------|--------|--------|--------|--------|--------|
| Haptoglobin                         | 7841   | 7846   | 7851   | 7856   | 7861   | 7866   | 39280  |
| Human Growth Hormone - Hgh          | 70     | 72     | 74     | 76     | 78     | 80     | 380    |
| Human IgG Subclasses                | 83     | 88     | 93     | 98     | 103    | 108    | 490    |
| Immunoglobulins IgA                 | 1422   | 1427   | 1432   | 1437   | 1442   | 1447   | 7185   |
| Immunoglobulins IgG                 | 1589   | 1594   | 1599   | 1604   | 1609   | 1614   | 8020   |
| Immunoglobulins IgM                 | 1518   | 1523   | 1528   | 1533   | 1538   | 1543   | 7665   |
| Insulin                             | 558    | 563    | 568    | 573    | 578    | 583    | 2865   |
| Iron                                | 19791  | 19801  | 19811  | 19821  | 19831  | 19841  | 99105  |
| Lactate                             | 8      | 10     | 12     | 14     | 16     | 18     | 70     |
| Lactate Dehydrogenase Total         | 39605  | 39620  | 39635  | 39650  | 39665  | 39680  | 198250 |
| Lipase                              | 6199   | 6209   | 6219   | 6229   | 6239   | 6249   | 31145  |
| Lithium                             | 118    | 120    | 122    | 124    | 126    | 128    | 620    |
| Luteinizing Hormone (LH)            | 2866   | 2876   | 2886   | 2896   | 2906   | 2916   | 14480  |
| Magnesium                           | 66776  | 66791  | 66806  | 66821  | 66836  | 66851  | 334105 |
| Magnesium Urine                     | 14     | 16     | 18     | 20     | 22     | 24     | 100    |
| Urine Albumin (Microalbumin)        | 768    | 779    | 790    | 801    | 812    | 823    | 4005   |
| N-terminal proBNP (NT-proBNP)       | 3873   | 3888   | 3903   | 3918   | 3933   | 3948   | 19590  |
| Oestradiol Total                    | 2530   | 2545   | 2560   | 2575   | 2590   | 2605   | 12875  |
| Paracetamol/Acetaminophen           | 996    | 1006   | 1016   | 1026   | 1036   | 1046   | 5130   |
| Parathyroid Hormone                 | 284    | 289    | 294    | 299    | 304    | 309    | 1495   |
| Pentobarbital                       | 473    | 478    | 483    | 488    | 493    | 498    | 2440   |
| Phenobarbitone                      | 733    | 738    | 743    | 748    | 753    | 758    | 3740   |
| Phenytoin                           | 300    | 305    | 310    | 315    | 320    | 325    | 1575   |
| Phosphatase Alkaline                | 55991  | 56006  | 56021  | 56036  | 56051  | 56066  | 280180 |
| Phosphorus - Urine                  | 16     | 22     | 29     | 35     | 41     | 47     | 174    |
| Phosphorus                          | 66801  | 66816  | 66831  | 66846  | 66861  | 66876  | 334230 |
| Potassium                           | 139247 | 139262 | 139277 | 139292 | 139307 | 139322 | 696460 |
| Potassium - Fluid                   | 7230   | 7235   | 7240   | 7245   | 7250   | 7255   | 36225  |
| Potassium -Urine                    | 336    | 341    | 346    | 351    | 356    | 361    | 1755   |
| Procalcitonin                       | 18357  | 18377  | 18397  | 18417  | 18437  | 18457  | 92085  |
| Progesterone                        | 4047   | 4057   | 4067   | 4077   | 4087   | 4097   | 20385  |
| Prolactin                           | 2683   | 2693   | 2703   | 2713   | 2723   | 2733   | 13565  |
| Prostatic Specific Antigen (PSA)    | 68742  | 68762  | 68782  | 68802  | 68822  | 68842  | 344010 |
| Protein Total (Fluid)               | 614    | 619    | 624    | 629    | 634    | 639    | 3145   |
| Protein Total (CSF)                 | 4263   | 4275   | 4287   | 4299   | 4311   | 4323   | 21495  |
| Protein Total Urine                 | 2010   | 2015   | 2020   | 2025   | 2030   | 2035   | 10125  |
| Protein Total (serum)               | 33999  | 34017  | 34035  | 34053  | 34071  | 34089  | 170264 |
| Sex Hormone Binding Globulin (SHBG) | 494    | 499    | 504    | 509    | 514    | 519    | 2545   |

Confidential Page 56 of 88



|                                   |         |         | 1       | 1       | 1       | 1       | 1        |
|-----------------------------------|---------|---------|---------|---------|---------|---------|----------|
| Salicylates                       | 739     | 744     | 749     | 754     | 759     | 764     | 3770     |
| Serum Ferritin                    | 19530   | 19545   | 19560   | 19575   | 19591   | 19606   | 97877    |
| Serum Folate                      | 19861   | 19876   | 19891   | 19906   | 19921   | 19936   | 99530    |
| Serum Vitamin B12                 | 31282   | 31297   | 31312   | 31327   | 31342   | 31357   | 156635   |
| Sodium                            | 139247  | 139257  | 139267  | 139277  | 139287  | 139297  | 696385   |
| Sodium Fluid & Urine              | 5788    | 5798    | 5808    | 5818    | 5828    | 5838    | 29090    |
| Testosterone Total                | 1847    | 1857    | 1867    | 1877    | 1887    | 1897    | 9385     |
| Theophylline                      | 70      | 75      | 80      | 85      | 90      | 95      | 425      |
| Thyrogobulin                      | 29      | 34      | 39      | 44      | 49      | 54      | 220      |
| Thyroid Stimulating Hormone (TSH) | 47371   | 47391   | 47411   | 47431   | 47451   | 47471   | 237155   |
| Thyroxine Free (FT4)              | 36001   | 36021   | 36041   | 36061   | 36081   | 36101   | 180305   |
| Transferrin                       | 19403   | 19418   | 19433   | 19448   | 19463   | 19478   | 97240    |
| Triglyceride                      | 42613   | 42628   | 42643   | 42658   | 42673   | 42688   | 213290   |
| cTroponin I (high sensitivity)    | 13382   | 13392   | 13402   | 13412   | 13422   | 13432   | 67060    |
| TSH Receptor Antibodies           | 200     | 202     | 204     | 206     | 208     | 210     | 1030     |
| Urea                              | 139247  | 139262  | 139277  | 139292  | 139307  | 139322  | 696460   |
| Urea - Fluid                      | 338     | 343     | 348     | 353     | 358     | 363     | 1765     |
| Urea - Urine                      | 3304    | 3309    | 3314    | 3319    | 3324    | 3329    | 16595    |
| Uric Acid - Fluid & Urine         | 85      | 90      | 95      | 100     | 105     | 110     | 500      |
| Uric Acid                         | 16485   | 16500   | 16515   | 16530   | 16545   | 16560   | 82650    |
| Urinary Free Cortisol             | 24      | 29      | 34      | 39      | 44      | 49      | 195      |
| Valproic Acid Levels              | 2258    | 2263    | 2268    | 2273    | 2278    | 2283    | 11365    |
| Vancomycin Levels                 | 126     | 131     | 136     | 141     | 146     | 151     | 705      |
| Vitamin D                         | 331     | 336     | 341     | 346     | 351     | 356     | 1730     |
| Holotranscobalamin                | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Everolimis                        | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Sirolimis                         | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Mycophenolic acid                 | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Methotraxate                      | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Cystatin C                        | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| IL-6                              | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Total Bile Acid                   | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| ApoLipoprotein B                  | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Lipoprotein (a)                   | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Soluble transferrin receptors     | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Glycated Albumin                  | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Calcitonin                        | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Cyclosporin                       | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Tricyclic antidepressants (TCA)   | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Enhanced Liver fibrosis(ELF)      | 25      | 30      | 35      | 40      | 45      | 50      | 200      |
| Total                             | 2485170 | 2486361 | 2487552 | 2488742 | 2489933 | 2491124 | 12443712 |

Confidential Page 57 of 88



**Department: HAEMATOLOGY** 

**Table 2: DGM Coagulation and ESR Analyzer: Volumes** 

| TEST TEST                               | FY 23/24 | FY 24/25 | FY 25/26 | FY 26/27 | FY27/28 | Test<br>Volumes/5<br>years |
|---|----------|----------|----------|----------|---------|----------------------------|
| COAGULATION ANALYSER                    |          |          |          |          |         |                            |
| Prothrombin Time / Int Normalised Ratio | 17289    | 19017    | 20918    | 23009    | 25309   | 105542                     |
| Activated Partial Thromboplastin Time   | 10832    | 11915    | 13106    | 14416    | 15857   | 66126                      |
| D-Dimer                                 | 9108     | 10018    | 11019    | 12120    | 13332   | 55597                      |
| Fibrinogen                              | 4015     | 4416     | 4857     | 5342     | 5876    | 24506                      |
| Thrombin Time                           | 394      | 433      | 476      | 523      | 575     | 2401                       |
| Lupus Anticoagulant                     | 1069     | 1175     | 1292     | 1421     | 1563    | 6520                       |
| Lupus Sensitive APTT                    | 1063     | 1169     | 1285     | 1413     | 1554    | 6484                       |
| Protein C                               | 937      | 1030     | 1133     | 1246     | 1370    | 5716                       |
| Free Protein S                          | 944      | 1038     | 1141     | 1255     | 1380    | 5758                       |
| Anti-Thrombin III                       | 428      | 470      | 517      | 568      | 624     | 2607                       |
| Factor VIII                             | 69       | 75       | 82       | 90       | 99      | 415                        |
| Factor VIII Inhibitor                   | 72       | 79       | 86       | 94       | 103     | 434                        |
| Factor IX                               | 50       | 55       | 60       | 66       | 72      | 303                        |
| Factor IX Inhibitor                     | 5        | 6        | 7        | 8        | 9       | 35                         |
| Anti Factor Xa Assay                    | 29       | 31       | 34       | 37       | 40      | 171                        |
| von Willebrand Factor Antigen           |          |          |          |          |         |                            |
| von Willebrand Factor Activity          |          |          |          |          |         |                            |
| Total                                   | 46304    | 50927    | 56013    | 61608    | 67763   | 282615                     |
| AUTOMATED ESR ANALYSER                  |          |          |          |          |         |                            |
| Erythrocyte Sedimentation Rate          | 8767     | 9643     | 10607    | 11667    | 12833   | 53517                      |
| Total                                   | 8767     | 9643     | 10607    | 11667    | 12833   | 53517                      |
| OTHER                                   |          |          |          |          |         |                            |
| Erythropoetin                           | 17       | 18       | 19       | 20       | 22      | 96                         |

Confidential Page 58 of 88



| TEST  | FY 23/24 | FY 24/25 | FY 25/26 | FY 26/27 | FY27/28 | Test<br>Volumes/5<br>years |
|-------|----------|----------|----------|----------|---------|----------------------------|
| Total | 17       | 18       | 19       | 20       | 22      | 96                         |

Confidential Page **59** of **88** 



**Table 3: DGM Stat Laboratory Volumes** 

| D                                       | CDA CT |        |       |       |       |         |
|---|--------|--------|-------|-------|-------|---------|
| Test                                    | FY25   | AT Lab | FY27  | FY28  | FY29  | Total   |
| Alanine Transaminase                    | 19329  | 19334  | 19340 | 19345 | 19351 | 136 477 |
| Albumin - Serum                         | 11890  | 11908  | 11926 | 11944 | 11962 | 85 983  |
| Amylase Total - Serum                   | 1596   | 1601   | 1606  | 1611  | 1616  | 11 100  |
| Aspartate Transaminase                  | 9476   | 9491   | 9506  | 9521  | 9536  | 68 470  |
| BHCG (human chorionic gonadotropin)     | 2332   | 2345   | 2357  | 2370  | 2382  | 16 759  |
| Bilirubin Direct                        | 10116  | 10131  | 10147 | 10162 | 10178 | 72 793  |
| Bilirubin Total                         | 16075  | 16084  | 16094 | 16104 | 16113 | 114 304 |
| Calcium - Serum                         | 17941  | 17954  | 17966 | 17978 | 17990 | 128 170 |
| Chloride                                | 41955  | 41965  | 41975 | 41985 | 41995 | 293 597 |
| Cholesterol HDL                         | 5876   | 5891   | 5906  | 5921  | 5936  | 42 831  |
| Cholesterol Total                       | 22390  | 22405  | 22420 | 22435 | 22450 | 160 551 |
| Cholestrol LDL - Measured               | 10153  | 10168  | 10183 | 10198 | 10213 | 73 911  |
| Cholinesterase Total                    | 194    | 208    | 221   | 234   | 247   | 1 494   |
| CO2 Total                               | 41955  | 41965  | 41975 | 41985 | 41995 | 293 597 |
| Cortisol                                | 366    | 381    | 396   | 411   | 426   | 2 804   |
| C-Reactive Protein(wide range & High Se | 20557  | 20566  | 20575 | 20585 | 20594 | 146 065 |
| Creatine Kinase MB                      | 3276   | 3278   | 3281  | 3284  | 3286  | 23 041  |
| Creatinine                              | 82184  | 82194  | 82204 | 82214 | 82223 | 584 291 |
| Triiodothyronine Free (FT3)             | 6441   | 6456   | 6471  | 6486  | 6501  | 47 298  |
| Glucose - Plasma                        | 3124   | 3129   | 3134  | 3139  | 3144  | 21 863  |
| Glucose -CSF                            | 48     | 58     | 68    | 78    | 88    | 431     |
| Glutamyl Transpeptidase                 | 8794   | 8809   | 8824  | 8839  | 8854  | 63 108  |
| Lactate Dehydrogenase Total             | 10710  | 10725  | 10740 | 10755 | 10770 | 76 275  |
| Lipase                                  | 1759   | 1769   | 1779  | 1789  | 1799  | 12 504  |
| Magnesium                               | 17515  | 17530  | 17545 | 17560 | 17575 | 125 255 |
| N-terminal proBNP (NT-proBNP)           | 994    | 1009   | 1024  | 1039  | 1054  | 7 262   |
| Phosphatase Alkaline                    | 16004  | 16019  | 16034 | 16049 | 16064 | 112 953 |
| Phosphorus                              | 17743  | 17758  | 17773 | 17788 | 17803 | 126 631 |
| Potassium                               | 41960  | 41975  | 41990 | 42005 | 42020 | 293 672 |
| Procalcitonin                           | 7515   | 7535   | 7555  | 7575  | 7595  | 50 778  |
| Protein Total (CSF)                     | 1135   | 1147   | 1159  | 1171  | 1183  | 8 196   |
| Protein Total (serum)                   | 8387   | 8405   | 8423  | 8441  | 8459  | 60 682  |
| Sodium                                  | 41955  | 41965  | 41975 | 41985 | 41995 | 293 597 |
| Thyroid Stimulating Hormone (TSH)       | 10125  | 10145  | 10165 | 10185 | 10205 | 75 141  |
| Thyroxine Free (FT4)                    | 7897   | 7917   | 7937  | 7957  | 7977  | 58 361  |
| Triglyceride                            | 11189  | 11204  | 11219 | 11234 | 11249 | 80 052  |
| cTroponin I (high sensitivity)          | 4024   | 4034   | 4044  | 4054  | 4064  | 28 250  |
| Urea                                    | 41960  | 41975  | 41990 | 42005 | 42020 | 293 672 |
| Uric Acid                               | 4110   | 4125   | 4140  | 4155  | 4170  | 29 742  |



**Department: Virology (Serology)** 

**Table 4: DGM Serology Laboratory Volumes** 

| Serology Tests_(Virology) |                         |        |        |        |        |        |       |
|---------------------------|-------------------------|--------|--------|--------|--------|--------|-------|
| Test                      | FY23_24 Test<br>Volumes | Year 1 | Year 2 | Year 3 | Year 4 | Year 5 | Total |
| Anti-HBc                  | 700                     | 753    | 809    | 870    | 935    | 1005   | 4371  |
| Anti-HBc IgM              | 200                     | 215    | 231    | 248    | 267    | 287    | 1249  |
| Anti-HBs                  | 1000                    | 1075   | 1156   | 1242   | 1335   | 1436   | 6244  |
| Anti-HCV                  | 600                     | 645    | 693    | 745    | 801    | 861    | 3746  |
| HBsAg                     | 5100                    | 5483   | 5894   | 6336   | 6811   | 7322   | 31845 |
| HBeAg                     | 925                     | 994    | 1069   | 1149   | 1235   | 1328   | 5776  |
| HBeAb                     | 100                     | 108    | 116    | 124    | 134    | 144    | 624   |
| HIV Ag/Ab                 | 900                     | 968    | 1040   | 1118   | 1202   | 1292   | 5620  |
| HTLV I/II                 | 100                     | 108    | 116    | 124    | 134    | 144    | 624   |
| Hepatitis A IgM           | 700                     | 753    | 809    | 870    | 935    | 1005   | 4371  |
| Rubella IgG               | 600                     | 645    | 693    | 745    | 801    | 861    | 3746  |
| Rubella IgM               | 600                     | 645    | 693    | 745    | 801    | 861    | 3746  |
| CMV IgG                   | 400                     | 430    | 462    | 497    | 534    | 574    | 2498  |
| CMV IgM                   | 400                     | 430    | 462    | 497    | 534    | 574    | 2498  |
| EBV NA IgG                | 350                     | 376    | 404    | 435    | 467    | 502    | 2185  |
| EBV VCA IgG               |                         | 0      | 0      | 0      | 100    | 108    | 208   |
| EBV VCA IgM               | 350                     | 376    | 404    | 435    | 467    | 502    | 2185  |
| Parvovirus IgG            |                         | 0      | 0      | 0      | 100    | 108    | 208   |
| Parvovirus IgM            |                         | 0      | 0      | 0      | 100    | 108    | 208   |
| Toxoplasma IgG            | 300                     | 323    | 347    | 373    | 401    | 431    | 1873  |
| Toxoplasma IgM            | 320                     | 344    | 370    | 398    | 427    | 459    | 1998  |



#### ANNEXURE C: Bidder's Disclosure (SBD4)

#### 1. PURPOSE OF THE FORM

Any person (natural or juristic) may make an offer or offers in terms of this invitation to bid. In line with the principles of transparency, accountability, impartiality, and ethics as enshrined in the Constitution of the Republic of South Africa and further expressed in various pieces of legislation, it is required for the bidder to make this declaration in respect of the details required hereunder.

Where a person/s are listed in the Register for Tender Defaulters and / or the List of Restricted Suppliers, that person will automatically be disqualified from the bid process.

#### 2. Bidder's declaration

2.1 Is the bidder, or any of its directors / trustees / shareholders / members / partners or any person having a controlling interest<sup>1</sup> in the enterprise,

employed by the state?

YES/NO

2.1.1 If so, furnish particulars of the names, individual identity numbers, and, if applicable, state employee numbers of sole proprietor/ directors / trustees / shareholders / members/ partners or any person having a controlling interest in the enterprise, in table below.

| Full Name | Identity Number | Name of State institution |
|-----------|-----------------|---------------------------|
|           |                 |                           |
|           |                 |                           |
|           |                 |                           |
|           |                 |                           |
|           |                 |                           |
|           |                 |                           |
|           |                 |                           |
|           |                 |                           |
|           |                 |                           |

| 2.2   | Do you, or any person connected with the bidder, have a relationship with any person who is employed |
|-------|--|
|       | by the procuring institution? YES/NO   |
| 2.2.1 | If so, furnish particulars:  |
|       |  |
|       |  |

<sup>&</sup>lt;sup>1</sup> the power, by one person or a group of persons holding the majority of the equity of an enterprise, alternatively, the person/s having the deciding vote or power to influence or to direct the course and decisions of the enterprise.



2.3 Does the bidder or any of its directors / trustees / shareholders / members / partners or any person having a controlling interest in the enterprise have any interest in any other related enterprise whether or not they are bidding for this contract?

YES/NO

|           | or not they are bidding for this contract? YES/NO  |
|-----------|--|
| 2.3.1     | If so, furnish particulars:  |
| 3 DE      | ECLARATION   |
| I, the ur | ndersigned, (name) in submitting the accompanying bid, do hereb  |
| make th   | ne following statements that I certify to be true and complete in every respect:   |
| 3.1       | I have read and I understand the contents of this disclosure;  |
| 3.2       | I understand that the accompanying bid will be disqualified if this disclosure is found not to be true and complete in every respect;  |
| 3.3       | The bidder has arrived at the accompanying bid independently from, and without consultation communication, agreement or arrangement with any competitor. However, communication between partner in a joint venture or consortium <sup>2</sup> will not be construed as collusive bidding.  |
| 3.4       | In addition, there have been no consultations, communications, agreements or arrangements with an competitor regarding the quality, quantity, specifications, prices, including methods, factors or formulas used to calculate prices, market allocation, the intention or decision to submit or not to submit the bid, bidding with the intention not to win the bid and conditions or delivery particulars of the products or services to which this bid invitation relates. |
| 3.4       | The terms of the accompanying bid have not been, and will not be, disclosed by the bidder, directly or indirectly to any competitor, prior to the date and time of the official bid opening or of the awarding of the contract.  |

- 3.5 There have been no consultations, communications, agreements or arrangements made by the bidder with any official of the procuring institution in relation to this procurement process prior to and during the bidding process except to provide clarification on the bid submitted where so required by the institution; and the bidder was not involved in the drafting of the specifications or terms of reference for this bid.
- 3.6 I am aware that, in addition and without prejudice to any other remedy provided to combat any restrictive practices related to bids and contracts, bids that are suspicious will be reported to the Competition Commission

<sup>&</sup>lt;sup>2</sup> Joint venture or Consortium means an association of persons for the purpose of combining their expertise, property, capital, efforts, skill and knowledge in an activity for the execution of a contract.



for investigation and possible imposition of administrative penalties in terms of section 59 of the Competition Act No 89 of 1998 and or may be reported to the National Prosecuting Authority (NPA) for criminal investigation and or may be restricted from conducting business with the public sector for a period not exceeding ten (10) years in terms of the Prevention and Combating of Corrupt Activities Act No 12 of 2004 or any other applicable legislation.

I CERTIFY THAT THE INFORMATION FURNISHED IN PARAGRAPHS 1, 2 and 3 ABOVE IS CORRECT.

I ACCEPT THAT THE STATE MAY REJECT THE BID OR ACT AGAINST ME IN TERMS OF PARAGRAPH 6 OF PFMA SCM INSTRUCTION 03 OF 2021/22 ON PREVENTING AND COMBATING ABUSE IN THE SUPPLY CHAIN MANAGEMENT SYSTEM SHOULD THIS DECLARATION PROVE TO BE FALSE.

| Signature | Date           |
|-----------|----------------|
|           |                |
|           |                |
| Position  | Name of bidder |



#### ANNEXURE D: Preferential Procurement Claim Form (SBD6.1)

#### PREFERENCE POINTS CLAIM FORM IN TERMS OF THE PREFERENTIAL PROCUREMENT REGULATIONS 2022

This preference form must form part of all tenders invited. It contains general information and serves as a claim form for preference points for specific goals.

NB: BEFORE COMPLETING THIS FORM, TENDERERS MUST STUDY THE GENERAL CONDITIONS, DEFINITIONS AND DIRECTIVES APPLICABLE IN RESPECT OF THE TENDER AND PREFERENTIAL PROCUREMENT REGULATIONS, 2022.

#### 1 GENERAL CONDITIONS

- 1.1 The following preference point systems are applicable to invitations to tender:
  - the 80/20 or 90/10 system for requirements with a Rand value of up to R50 000 000 (all applicable taxes included); and

#### 2.1 To be completed by the organ of state

(delete whichever is not applicable for this tender).

- a) The applicable preference point system for this tender is the 80/20 preference point system.
- b) The 80/20 or 90/10 preference point system will be applicable in this tender. The lowest/ highest acceptable tender will be used to determine the accurate system once tenders are received.
- 2.2 Points for this tender (even in the case of a tender for income-generating contracts) shall be awarded for:

a) Price; and

b)Specific Goals.

#### 2.3 To be completed by the organ of state:

The maximum points for this tender are allocated as follows:

|   | POINTS |
|---|--------|
| PRICE                                     | 80/90  |
| SPECIFIC GOALS                            | 20/10  |
| Total points for Price and SPECIFIC GOALS | 100    |

- 2.4 Failure on the part of a tenderer to submit proof or documentation required in terms of this tender to claim points for specific goals with the tender, will be interpreted to mean that preference points for specific goals are not claimed.
- 2.5 The organ of state reserves the right to require of a tenderer, either before a tender is adjudicated or at any time subsequently, to substantiate any claim in regard to preferences, in any manner required by the organ of state



Table 1: Specific goals for the tender and points claimed are indicated per the table below.

(Note to organs of state: The 80/20 or 90/10 preference point system is applicable, corresponding points must also be indicated as such.

Note to tenderers: The tenderer must indicate how they claim points for each preference point system.)

| The specific goals allocated points in terms of this tender  |   | Number of points  allocated  (80/20 system)  (To be completed by the organ of state) | Number of points claimed<br>(80/20 system)<br>(To be completed by the<br>tenderer) | Number of points allocated (90/10 system) (To be completed by the organ of state) | Number of points claimed (90/10 system)  (To be completed by the tenderer) |
|--|---|--|--|---|--|
| Historically Disadvantaged Individuals.  | Valid B-BBEE  | 6  |  | 3   |  |
| (Means a South African citizen who, due to the apartheid policy that had been in place, had no franchise in national elections prior to the  | Certificate/Affidavit Sworn under oath,   |  |  |   |  |
| introduction of the Constitution of the Republic of South Africa, 1983 (Act No. 110 of 1983) or the Constitution of the Republic of South Africa, 1993 (Act No. 200 of 1993) ("The Interim Constitution"). | ID copy of owner/s of the business /  |  |  |   |  |
|  | Ownership certificate issued<br>by Companies and<br>Intellectual Property<br>Commission (CIPC). |  |  |   |  |
| Woman  | Valid B-BBEE  | 4  |  | 2   |  |
|  | Certificate/Affidavit Sworn under oath,   |  |  |   |  |
|  | ID copy of owner/s of the business /  |  |  |   |  |
|  | Ownership certificate issued<br>by Companies and<br>Intellectual Property<br>Commission (CIPC). |  |  |   |  |
| Disabled   | Valid B-BBEE  | 1  |  | 1   |  |



| The specific goals allocated points in terms of this tender |   | Number of points  allocated  (80/20 system)  (To be completed by the organ of state) | Number of points claimed<br>(80/20 system)<br>(To be completed by the<br>tenderer) | Number of points  allocated  (90/10 system)  (To be completed by the organ of state) | Number of points claimed (90/10 system) (To be completed by the tenderer) |
|---|---|--|--|--|---|
|   | Certificate/Affidavit Sworn under oath, ID copy of owner/s of the business / Ownership certificate issued by Companies and Intellectual Property                                    |  |  |  |   |
| Youth   | Commission (CIPC).  Valid B-BBEE  Certificate/Affidavit Sworn under oath, ID copy of owner/s of the business /  Ownership certificate issued by Companies and Intellectual Property | 4  |  | 2  |   |
| Locality  • Gauteng Province = 5  = 2  • National = 0       | Commission (CIPC).  CSD/proof of municipal account /letter from the Ward  Council confirming the business address.  Lease agreement.  | 5  |  | 2  |   |



| The specific goals allocated points in terms of this tender | Number of points allocated (80/20 system) (To be completed by the organ of state) | Number of points claimed<br>(80/20 system)<br>(To be completed by the<br>tenderer) | Number of points allocated (90/10 system) (To be completed by the organ of state) | Number of points claimed (90/10 system)  (To be completed by the tenderer) |
|---|---|--|---|--|
| Total Points  | 20  |  | 10  |  |



### **DECLARATION WITH REGARD TO COMPANY/FIRM**

| 3.1. | Name of company/firm                     |  |  |  |  |  |
|------|--|--|--|--|--|--|
| 3.2. | Company registration number:             |  |  |  |  |  |
| 3.3. | TYPE OF COMPANY/ FIRM                    |  |  |  |  |  |
|      | ☐ Partnership/Joint Venture / Consortium |  |  |  |  |  |
|      | ☐ One-person business/sole propriety     |  |  |  |  |  |
|      | ☐ Close corporation                      |  |  |  |  |  |
|      | □ Public Company                         |  |  |  |  |  |
|      | ☐ Personal Liability Company             |  |  |  |  |  |
|      | ☐ (Pty) Limited                          |  |  |  |  |  |
|      | □ Non-Profit Company                     |  |  |  |  |  |
|      | □ State Owned Company                    |  |  |  |  |  |
|      | [TICK APPLICABLE BOX]                    |  |  |  |  |  |

- 3.4. I, the undersigned, who is duly authorised to do so on behalf of the company/firm, certify that the points claimed, based on the specific goals as advised in the tender, qualifies the company/ firm for the preference(s) shown and I acknowledge that:
  - i) The information furnished is true and correct;
  - ii) The preference points claimed are in accordance with the General Conditions as indicated in paragraph 1 of this form;
  - iii) In the event of a contract being awarded as a result of points claimed as shown in paragraphs 1.4 and 4.2, the contractor may be required to furnish documentary proof to the satisfaction of the organ of state that the claims are correct;

Confidential Page **69** of **88** 



- iv) If the specific goals have been claimed or obtained on a fraudulent basis or any of the conditions of contract have not been fulfilled, the organ of state may, in addition to any other remedy it may have
  - (a) disqualify the person from the tendering process;
  - (b) recover costs, losses or damages it has incurred or suffered as a result of that person's conduct;
  - (c) cancel the contract and claim any damages which it has suffered as a result of having to make less favourable arrangements due to such cancellation;
  - (d) recommend that the tenderer or contractor, its shareholders and directors, or only the shareholders and directors who acted on a fraudulent basis, be restricted from obtaining business from any organ of state for a period not exceeding 10 years, after the audi alteram partem (hear the other side) rule has been applied; and
  - (e) forward the matter for criminal prosecution, if deemed necessary.

| SIGNATURE(S) OF TENDERER(S) |  |  |  |
|-----------------------------|--|--|--|
| SURNAME AND NAME:<br>DATE:  |  |  |  |
| ADDRESS:                    |  |  |  |
|                             |  |  |  |
|                             |  |  |  |
|                             |  |  |  |
|                             |  |  |  |
|                             |  |  |  |
|                             |  |  |  |
|                             |  |  |  |
|                             |  |  |  |



| SWORN AFFIDAVIT: B-BBEE QU      | UALIFYING SMALL ENTERPRISE: GENERAL  |
|---------------------------------|--|
| , the Undersigned               |  |
| Full Name and Surname:          |  |
| Identity Number:                |  |
| Hereby declare under oath as fo | ollows:  |
| 1. The contents                 | of this statement are to the best of my knowledge a true reflection of the facts.  |
| 2. I am a Membe                 | er / Director / Owner of the following enterprise and am duly authorised to act on |
| its behalf:                     |  |
| Enterprise Name:                |  |
| Trading (if applicable):        |  |
| Enterprise Physical             |  |
| Address:                        |  |
| Type of Entity (CC, Pty Ltd,    |  |
| Sole Prop etc.)                 |  |
| Nature of Business:             |  |
|                                 | As per the Broad-Based Black Economic Empowerment Act 53 of 2003 as                |
|                                 | Amended by Act No 46 of 2013 "Black People" is a generic term which means          |
|                                 | Africans, Coloureds and Indians –  |
| Definition of "Black People:    | a. Who are citizens of the Republic of South Africa by birth or descent; or        |
|                                 | b. Who became citizens of the Republic of South Africa by naturalization-          |
|                                 | i. Before 27 April 1994; or  |
|                                 | ii. On or after 27 April 1994 and who would have been entitled to acquire          |
|                                 | citizenship by naturalization prior to that date                                   |
| 3. I hereby decla               | are under Oath that:   |
| -                               | % Black Owned as per Amended Code Series 100 of the Amended Codes                  |
|                                 | d under section 9 (1) of B-BBEE Act No 53 of 2003 as Amended by Act No 46 of       |
| 2013.                           | d dilder section 5 (1) or b bble Act No 33 or 2003 as Amended by Act No 40 or      |
|                                 |  |
|                                 | od Practice issued under section 9 (1) of B-BBEE Act No 53 of 2003 as Amended by   |
| Act No 46 of 2013.              | · · · · · · · · · · · · · · · · · · ·  |



| <ul> <li>The Enterprise is</li> </ul>                              | % Black Designated Group Owned as per Amended Code S   | eries 100 of  |
|--|--|---------------|
| the Amended Codes of Good Pr                                       | actice issued under section 9 (1) of B-BBEE Act No 53 of 2003  | as Amended    |
| by Act No 46 of 2013.  |  |               |
| Based on the Financial Stateme                                     | nts/Management Accounts and other information available of   | on the latest |
| financial year-end of  | , the annual Total Revenue was between R10,000,  | 000.00 (Ten   |
| Million Rands) and R50,000,000                                     | 0.00 (Fifty Million Rands).  |               |
| Please confirm on the table bel                                    | ow the B-BBEE level contributor, by ticking the applicable box   | ۲.            |
| 100% Black Owned   | Level One (135% B-BBEE procurement recognition level)  |               |
| At least 51% Black Owned   | Level Two (125% B-BBEE procurement recognition level)  |               |
| prescribed oath and consider t<br>which I represent in this matter | nd the contents of this affidavit and I have no objection he oath binding on my conscience and on the Owners of the content of | e Enterprise  |
| Deponent Signature:  |  |               |
| Date:  |  |               |
|  |  |               |
| Commissioner of Oaths Signature and Stamp                          |  |               |



by Act No 46 of 2013.

# SWORN AFFIDAVIT: B-BBEE QUALIFYING MICRO ENTERPRISE: GENERAL

| ·  |   |
|--|---|
| , the Undersigned  |   |
| Full Name and Surname:   |   |
| Identity Number:   |   |
| Hereby declare under oath as fo  | ollows:   |
| 1. The contents of this staten   | nent are to the best of my knowledge a true reflection of the facts.  |
| 2. I am a Member / Director ,  | Owner of the following enterprise and am duly authorised to act on its behalf:  |
| Enterprise Name:   |   |
| Trading (if applicable):   |   |
| Enterprise Physical  |   |
| Address:   |   |
| Type of Entity (CC, Pty Ltd,   |   |
| Sole Prop etc.)  |   |
| Nature of Business:  |   |
| Definition of "Black People:   | As per the Broad-Based Black Economic Empowerment Act 53 of 2003 as Amended by Act No 46 of 2013 "Black People" is a generic term which means Africans, Coloureds and Indians —  c. Who are citizens of the Republic of South Africa by birth or descent; or d. Who became citizens of the Republic of South Africa by naturalization—  iii. Before 27 April 1994; or  iv. On or after 27 April 1994 and who would have been entitled to acquire citizenship by naturalization prior to that date |
| of Good Practice issued 2013.  The Enterprise is Amended Codes of Good | th that:% Black Owned as per Amended Code Series 100 of the Amended Codes d under section 9 (1) of B-BBEE Act No 53 of 2003 as Amended by Act No 46 of% Black Woman Owned as per Amended Code Series 100 of the od Practice issued under section 9 (1) of B-BBEE Act No 53 of 2003 as Amended by  |
|  | % Black Designated Group Owned as per Amended Code Series 100 of Good Practice issued under section 9 (1) of B-BBEE Act No 53 of 2003 as Amended  |



|        | based on the i mancial statem   | ents/Management Accounts and other information available on the latest   |  |
|--------|---|--|--|
|        | financial year-end of   | , the annual Total Revenue was between R10,000,000.00 (Ten   |  |
|        | Million Rands) or less.   |  |  |
|        |   |  |  |
|        | <ul> <li>Please confirm on the table be</li> </ul>  | low the B-BBEE level contributor, by ticking the applicable box.   |  |
|        | 100% Black Owned  | Level One (135% B-BBEE procurement recognition level)  |  |
|        | At least 51% Black Owned  | Level Two (125% B-BBEE procurement recognition level)  |  |
|        | Less than 51% Black Owned   | Level Four (100% B-BBEE procurement recognition level)   |  |
| 4.     | consider the oath binding on my omatter.  | nts of this affidavit and I have no objection to take the prescribed oath and conscience and on the Owners of the Enterprise which I represent in this |  |
| 5.     | The sworn affidavit will be valid for a period of 12 months from the date signed by commissioner. |  |  |
| Depo   | onent Signature:  |  |  |
| Date   | ::  |  |  |
| - 0.00 |   |  |  |
|        |   |  |  |
|        |   |  |  |
|        |   |  |  |
| Com    | missioner of Oaths  |  |  |
|        | ature and Stamp   |  |  |
| _      | •   |  |  |



### ANNEXURE E: Government Procurement: General Conditions of Contract - July 2011

#### **NOTES**

The purpose of this document is to:

- (i) Draw special attention to certain general conditions applicable to government Bids, contracts and orders; and
- (ii) To ensure that clients be familiar with regard to the rights and obligations of all parties involved in doing business with government.

In this document words in the singular also mean in the plural and vice versa and words in the masculine also mean in the feminine and neuter.

2 The GCC will form part of all bid documents and may not be amended.

② Special Conditions of Contract (SCC) relevant to a specific bid, should be compiled separately for every bid (if (applicable) and will supplement the GCC. Whenever there is a conflict, the provisions in the SCC shall prevail.

### **TABLE OF CLAUSES**

- 1. Definitions
- 2. Application
- 3. General
- 4. Standards
- 5. Use of contract documents and information; inspection
- 6. Patent rights
- 7. Performance security
- 8. Inspections, tests and analysis
- 9. Packing
- 10. Delivery and documents
- 11. Insurance
- 12. Transportation
- 13. Incidental services
- 14. Spare parts
- 15. Warranty
- 16. Payment
- 17. Prices
- 18. Contract amendments
- 19. Assignment



- 20. Subcontracts
- 21. Delays in the supplier's performance
- 22. Penalties
- 23. Termination for default
- 24. Dumping and countervailing duties
- 25. Force Majeure
- 26. Termination for insolvency
- 27. Settlement of disputes
- 28. Limitation of liability
- 29. Governing language
- 30. Applicable law
- 31. Notices
- 32. Taxes and duties
- 33. National Industrial Participation Programme (NIPP)
- 34. Prohibition of restrictive practices

#### **General conditions of contract**

### 1. Definitions

The following terms shall be interpreted as indicated:

- 1.1 "Closing time" means the date and hour specified in the bidding documents for the receipt of Bids.
- 1.2 "Contract" means the written agreement entered into between the purchaser and the supplier, as recorded in the contract form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.
- 1.3 "Contract price" means the price payable to the supplier under the contract for the full and proper performance of his contractual obligations.
- "Corrupt practice" means the offering, giving, receiving, or soliciting of anything of value to influence the action of a public official in the procurement process or in contract execution.
- "Countervailing duties" are imposed in cases where an enterprise abroad is subsidized by its government and encouraged to market its products internationally.
- "Country of origin" means the place where the goods were mined, grown or produced or from which the services are supplied. Goods are produced when, through manufacturing, processing or substantial and major assembly of components, a commercially recognized new product results that is substantially different in basic characteristics or in purpose or utility from its components.
- 1.7 "Day" means calendar day.



- 1.8 "Delivery" means delivery in compliance of the conditions of the contract or order.
- 1.9 "Delivery ex stock" means immediate delivery directly from stock actually on hand.
- "Delivery into consignees store or to his site" means delivered and unloaded in the specified store or depot or on the specified site in compliance with the conditions of the contract or order, the supplier bearing all risks and charges involved until the supplies are so delivered and a valid receipt is obtained.
- "Dumping" occurs when a private enterprise abroad market its goods on own initiative in the RSA at lower prices than that of the country of origin and which have the potential to harm the local industries in the RSA.
- 1.12 "Force majeure" means an event beyond the control of the supplier and not involving the supplier's fault or negligence and not foreseeable. Such events may include, but is not restricted to, acts of the purchaser in its sovereign capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions and freight embargoes.
- 1.13 "Fraudulent practice" means a misrepresentation of facts in order to influence a procurement process or the execution of a contract to the detriment of any bidder, and includes collusive practice among bidders (prior to or after bid submission) designed to establish bid prices at artificial non-competitive levels and to deprive the bidder of the benefits of free and open competition.
- 1.14 "GCC" means the General Conditions of Contract.
- "Goods" means all of the equipment, machinery, and/or other materials that the supplier is required to supply to the purchaser under the contract.
- "Imported content" means that portion of the bidding price represented by the cost of components, parts or materials which have been or are still to be imported (whether by the supplier or his subcontractors) and which costs are inclusive of the costs abroad, plus freight and other direct importation costs such as landing costs, dock dues, import duty, sales duty or other similar tax or duty at the South African place of entry as well as transportation and handling charges to the factory in the Republic where the supplies covered by the bid will be manufactured.
- 1.17 "Local content" means that portion of the bidding price which is not included in the imported content provided that local manufacture does take place.
- 1.18 "Manufacture" means the production of products in a factory using labour, materials, components and machinery and includes other related value-adding activities.
- 1.19 "Order" means an official written order issued for the supply of goods or works or the rendering of a service.
- 1.20 "Project site," where applicable, means the place indicated in bidding documents.
- 1.21 "Purchaser" means the organisation purchasing the goods.
- 1.22 "Republic" means the RSA.
- 1.23 "SCC" means the Special Conditions of Contract.



- "Services" means those functional services ancillary to the supply of the goods, such as transportation and any other incidental services, such as installation, commissioning, provision of technical assistance, training, catering, gardening, security, maintenance and other such obligations of the supplier covered under the contract.
- 1.25 "Written" or "in writing" means handwritten in ink or any form of electronic or mechanical writing.

### 2. Application

- 2.1 These general conditions are applicable to all Bids, contracts and orders including Bids for functional and professional services, sales, hiring, letting and the granting or acquiring of rights, but excluding immovable property, unless otherwise indicated in the bidding documents.
- 2.2 Where applicable, SCC are also laid down to cover specific supplies, services or works.
- 2.3 Where such SCC are in conflict with these general conditions, the special conditions shall apply.

#### 3. General

- 3.1 Unless otherwise indicated in the bidding documents, the purchaser shall not be liable for any expense incurred in the preparation and submission of a bid. Where applicable a non-refundable fee for documents may be charged.
- 3.2 With certain exceptions, invitations to bid are only published in the Government Tender Bulletin. The Government Tender Bulletin may be obtained directly from the Government Printer, Private Bag X85, Pretoria 0001, or accessed electronically from www.treasury.gov.za

### 4. Standards

4.1 The goods supplied shall conform to the standards mentioned in the bidding documents and specifications.

### 5. Use of contract documents and information; inspection

- 5.1 The supplier shall not, without the purchaser's prior written consent, disclose the contract, or any provision thereof, or any specification, plan, drawing, pattern, sample, or information furnished by or on behalf of the purchaser in connection therewith, to any person other than a person employed by the supplier in the performance of the contract. Disclosure to any such employed person shall be made in confidence and shall extend only so far as may be necessary for purposes of such performance.
- 5.2 The supplier shall not, without the purchaser's prior written consent, make use of any document or information mentioned in GCC clause 5.1 except for purposes of performing the contract.



- 5.3 Any document, other than the contract itself mentioned in GCC clause 5.1 shall remain the property of the purchaser and shall be returned (all copies) to the purchaser on completion of the supplier's performance under the contract if so required by the purchaser.
- 5.4 The supplier shall permit the purchaser to inspect the supplier's records relating to the performance of the supplier and to have them audited by auditors appointed by the purchaser, if so required by the purchaser.

#### 6. Patent rights

6.1 The supplier shall indemnify the purchaser against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the goods or any part thereof by the purchaser.

### 7. Performance security

- 7.1 Within thirty (30) days of receipt of the notification of contract award, the successful bidder shall furnish to the purchaser the performance security of the amount specified in SCC.
- 7.2 The proceeds of the performance security shall be payable to the purchaser as compensation for any loss resulting from the supplier's failure to complete his obligations under the contract.
- 7.3 The performance security shall be denominated in the currency of the contract, or in a freely convertible currency acceptable to the purchaser and shall be in one of the following forms:
- 7.3.1 a bank guarantee or an irrevocable letter of credit issued by a reputable bank located in the purchaser's country or abroad, acceptable to the purchaser, in the form provided in the bidding documents or another form acceptable to the purchaser; or
- 7.3.2 a cashier's or certified cheque
- 7.4 The performance security will be discharged by the purchaser and returned to the supplier not later than thirty (30) days following the date of completion of the supplier's performance obligations under the contract, including any warranty obligations, unless otherwise specified in SCC.

# 8. Inspections, tests and analyses

- 8.1 All pre-bidding testing will be for the account of the bidder.
- 8.2 If it is a bid condition that supplies to be produced or services to be rendered should at any stage during production or execution or on completion be subject to inspection, the premises of the bidder or contractor shall be open, at all reasonable hours, for inspection by a representative of the Department or an organisation acting on behalf of the Department.
- 8.3 If there are no inspection requirements indicated in the bidding documents and no mention is made in the contract, but during the contract period it is decided that inspections shall be carried out, the purchaser shall itself make the necessary arrangements, including payment arrangements with the testing authority concerned.



- 8.4 If the inspections, tests and analyses referred to in clauses 8.2 and 8.3 show the supplies to be in accordance with the contract requirements, the cost of the inspections, tests and analyses shall be defrayed by the purchaser.
- 8.5 Where the supplies or services referred to in clauses 8.2 and 8.3 do not comply with the contract requirements, irrespective of whether such supplies or services are accepted or not, the cost in connection with these inspections, tests or analyses shall be defrayed by the supplier.
- 8.6 Supplies and services which are referred to in clauses 8.2 and 8.3 and which do not comply with the contract requirements may be rejected.
- 8.7 Any contract supplies may on or after delivery be inspected, tested or analysed and may be rejected if found not to comply with the requirements of the contract. Such rejected supplies shall be held at the cost and risk of the supplier who shall, when called upon, remove them immediately at his own cost and forthwith substitute them with supplies which do comply with the requirements of the contract. Failing such removal, the rejected supplies shall be returned at the suppliers cost and risk. Should the supplier fail to provide the substitute supplies forthwith, the purchaser may, without giving the supplier further opportunity to substitute the rejected supplies, purchase such supplies as may be necessary at the expense of the supplier.
- 8.8 The provisions of clauses 8.4 to 8.7 shall not prejudice the right of the purchaser to cancel the contract on account of a breach of the conditions thereof, or to act in terms of Clause 23 of GCC.

## 9. Packing

- 9.1 The supplier shall provide such packing of the goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in the contract. The packing shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packing, case size and weights shall take into consideration, where appropriate, the remoteness of the goods' final destination and the absence of heavy handling facilities at all points in transit.
- 9.2 The packing, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the contract, including additional requirements, if any, specified in SCC, and in any subsequent instructions ordered by the purchaser.

### 10. Delivery and documents

- Delivery of the goods shall be made by the supplier in accordance with the terms specified in the contract.

  The details of shipping and/or other documents to be furnished by the supplier are specified in SCC.
- 10.2 Documents to be submitted by the supplier are specified in SCC.



#### 11. Insurance

11.1 The goods supplied under the contract shall be fully insured in a freely convertible currency against loss or damage incidental to manufacture or acquisition, transportation, storage and delivery in the manner specified in the SCC.

### 12. Transportation

12.1 Should a price other than an all-inclusive delivered price be required, this shall be specified in the SCC.

### 13. Incidental services

- 13.1 The supplier may be required to provide any or all of the following services, including additional services, if any, specified in SCC:
- 13.1.1 performance or supervision of on-site assembly and/or commissioning of the supplied goods;
- 13.1.2 furnishing of tools required for assembly and/or maintenance of the supplied goods;
- 13.1.3 furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied goods;
- 13.1.4 performance or supervision or maintenance and/or repair of the supplied goods, for a period of time agreed by the parties, provided that this service shall not relieve the supplier of any warranty obligations under this contract; and
- training of the purchaser's personnel, at the supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied goods.
- 13.2 Prices charged by the supplier for incidental services, if not included in the contract price for the goods, shall be agreed upon in advance by the parties and shall not exceed the prevailing rates charged to other parties by the supplier for similar services.

## 14. Spare parts

- 14.1 As specified in SCC, the supplier may be required to provide any or all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the supplier:
- such spare parts as the purchaser may elect to purchase from the supplier, provided that this election shall not relieve the supplier of any warranty obligations under the contract; and
- 14.1.2 in the event of termination of production of the spare parts:
- 14.1.2.1 Advance notification to the purchaser of the pending termination, in sufficient time to permit the purchaser to procure needed requirements; and
- 14.1.2.2 following such termination, furnishing at no cost to the purchaser, the blueprints, drawings, and specifications of the spare parts, if requested.

# 15. Warranty



- 15.1 The supplier warrants that the goods supplied under the contract are new, unused, of the most recent or current models, and that they incorporate all recent improvements in design and materials unless provided otherwise in the contract. The supplier further warrants that all goods supplied under this contract shall have no defect, arising from design, materials, or workmanship (except when the design and/or material is required by the purchaser's specifications) or from any act or omission of the supplier, that may develop under normal use of the supplied goods in the conditions prevailing in the country of final destination.
- 15.2 This warranty shall remain valid for twelve (12) months after the goods, or any portion thereof as the case may be, have been delivered to and accepted at the final destination indicated in the contract, or for eighteen (18) months after the date of shipment from the port or place of loading in the source country, whichever period concludes earlier, unless specified otherwise in SCC.
- 15.3 The purchaser shall promptly notify the supplier in writing of any claims arising under this warranty.
- 15.4 Upon receipt of such notice, the supplier shall, within the period specified in SCC and with all reasonable speed, repair or replace the defective goods or parts thereof, without costs to the purchaser.
- 15.5 If the supplier, having been notified, fails to remedy the defect(s) within the period specified in SCC, the purchaser may proceed to take such remedial action as may be necessary, at the supplier's risk and expense and without prejudice to any other rights which the purchaser may have against the supplier under the contract.

### 16. Payment

- 16.1 The method and conditions of payment to be made to the supplier under this contract shall be specified in SCC
- 16.2 The supplier shall furnish the purchaser with an invoice accompanied by a copy of the delivery note and upon fulfilment of other obligations stipulated in the contract.
- Payments shall be made promptly by the purchaser, but in no case later than thirty (30) days after submission of an invoice or claim by the supplier.
- 16.4 Payment will be made in rand unless otherwise stipulated in SCC.

### 17. Prices

17.1 Prices charged by the supplier for goods delivered and services performed under the contract shall not vary from the prices quoted by the supplier in his bid, with the exception of any price adjustments authorised in SCC or in the purchaser's request for bid validity extension, as the case may be.

### 18. Contract amendments

18.1 No variation in or modification of the terms of the contract shall be made except by written amendment signed by the parties concerned.



# 19. Assignment

19.1 The supplier shall not assign, in whole or in part, its obligations to perform under the contract, except with the purchaser's prior written consent.

#### 20. Subcontracts

20.1 The supplier shall notify the purchaser in writing of all subcontracts awarded under this contract if not already specified in the bid. Such notification, in the original bid or later, shall not relieve the supplier from any liability or obligation under the contract.

### 21. Delays in the supplier's performance

- 21.1 Delivery of the goods and performance of services shall be made by the supplier in accordance with the time schedule prescribed by the purchaser in the contract.
- 21.2 If at any time during performance of the contract, the supplier or its subcontractor(s) should encounter conditions impeding timely delivery of the goods and performance of services, the supplier shall promptly notify the purchaser in writing of the fact of the delay, its likely duration and its cause(s). As soon as practicable after receipt of the supplier's notice, the purchaser shall evaluate the situation and may at his discretion extend the supplier's time for performance, with or without the imposition of penalties, in which case the extension shall be ratified by the parties by amendment of contract.
- 21.3 No provision in a contract shall be deemed to prohibit the obtaining of supplies or services from a national department, provincial department, or a local authority.
- 21.4 The right is reserved to procure outside of the contract small quantities or to have minor essential services executed if an emergency arises, the supplier's point of supply is not situated at or near the place where the supplies are required, or the supplier's services are not readily available.
- 21.5 Except as provided under GCC Clause 25, a delay by the supplier in the performance of its delivery obligations shall render the supplier liable to the imposition of penalties, pursuant to GCC Clause 22, unless an extension of time is agreed upon pursuant to GCC Clause 21.2 without the application of penalties.
- 21.6 Upon any delay beyond the delivery period in the case of a supplies contract, the purchaser shall, without cancelling the contract, be entitled to purchase supplies of a similar quality and up to the same quantity in substitution of the goods not supplied in conformity with the contract and to return any goods delivered later at the supplier's expense and risk, or to cancel the contract and buy such goods as may be required to complete the contract and without prejudice to his other rights, be entitled to claim damages from the supplier.



#### 22. Penalties

22.1 Subject to GCC Clause 25, if the supplier fails to deliver any or all of the goods or to perform the services within the period(s) specified in the contract, the purchaser shall, without prejudice to its other remedies under the contract, deduct from the contract price, as a penalty, a sum calculated on the delivered price of the delayed goods or unperformed services using the current prime interest rate calculated for each day of the delay until actual delivery or performance. The purchaser may also consider termination of the contract pursuant to GCC Clause 23.

### 23. Termination for default

- 23.1 The purchaser, without prejudice to any other remedy for breach of contract, by written notice of default sent to the supplier, may terminate this contract in whole or in part:
- 23.1.1 if the supplier fails to deliver any or all of the goods within the period(s) specified in the contract, or within any extension thereof granted by the purchaser pursuant to GCC Clause 21.2;
- 23.1.2 if the Supplier fails to perform any other obligation(s) under the contract; or
- 23.1.3 if the supplier, in the judgment of the purchaser, has engaged in corrupt or fraudulent practices in competing for or in executing the contract.
- 23.2 In the event the purchaser terminates the contract in whole or in part, the purchaser may procure, upon such terms and in such manner as it deems appropriate, goods, works or services similar to those undelivered, and the supplier shall be liable to the purchaser for any excess costs for such similar goods, works or services. However, the supplier shall continue performance of the contract to the extent not terminated.
- 23.3 Where the purchaser terminates the contract in whole or in part, the purchaser may decide to impose a restriction penalty on the supplier by prohibiting such supplier from doing business with the public sector for a period not exceeding 10 years.
- 23.4 If a purchaser intends imposing a restriction on a supplier or any person associated with the supplier, the supplier will be allowed a time period of not more than fourteen (14) days to provide reasons why the envisaged restriction should not be imposed. Should the supplier fail to respond within the stipulated fourteen (14) days the purchaser may regard the intended penalty as not objected against and may impose it on the supplier.
- 23.5 Any restriction imposed on any person by the Accounting Officer / Authority will, at the discretion of the Accounting Officer / Authority, also be applicable to any other enterprise or any partner, manager, director or other person who wholly or partly exercises or exercised or may exercise control over the enterprise of the first-mentioned person, and with which enterprise or person the first-mentioned person, is or was in the opinion of the Accounting Officer / Authority actively associated.



- 23.6 If a restriction is imposed, the purchaser must, within five (5) working days of such imposition, furnish the National Treasury, with the following information:
- 23.6.1 the name and address of the supplier and / or person restricted by the purchaser;
- 23.6.2 the date of commencement of the restriction
- 23.6.3 the period of restriction; and
- 23.6.4 the reasons for the restriction.
- 23.7 These details will be loaded in the National Treasury's central database of suppliers or persons prohibited from doing business with the public sector.
- 23.8 If a court of law convicts a person of an offence as contemplated in sections 12 or 13 of the Prevention and Combating of Corrupt Activities Act, No. 12 of 2004, the court may also rule that such person's name be endorsed on the Register for Tender Defaulters. When a person's name has been endorsed on the Register, the person will be prohibited from doing business with the public sector for a period not lesthan five years and not more than 10 years. The National Treasury is empowered to determine the period of restriction and each case will be dealt with on its own merits. According to section 32 of the Act the Register must be open to the public. The Register can be perused on the National Treasury website.

# 24. Anti-dumping and countervailing duties and rights

When, after the date of bid, provisional payments are required, or anti-dumping or countervailing duties are imposed, or the amount of a provisional payment or anti-dumping or countervailing right is increased in respect of any dumped or subsidized import, the State is not liable for any amount so required or imposed, or for the amount of any such increase. When, after the said date, such a provisional payment is no longer required or any such anti-dumping or countervailing right is abolished, or where the amount of such provisional payment or any such right is reduced, any such favourable difference shall on demand be paid forthwith by the contractor to the State or the State may deduct such amounts from moneys (if any) which may otherwise be due to the contractor in regard to supplies or services which he delivered or rendered, or is to deliver or render in terms of the contract or any other contract or any other amount which may be due to him.

# 25. Force majeure

- 25.1 Notwithstanding the provisions of GCC Clauses 22 and 23, the supplier shall not be liable for forfeiture of its performance security, damages, or termination for default if and to the extent that his delay in performance or other failure to perform his obligations under the contract is the result of an event of force majeure.
- 25.2 If a force majeure situation arises, the supplier shall promptly notify the purchaser in writing of such condition and the cause thereof. Unless otherwise directed by the purchaser in writing, the supplier shall



continue to perform its obligations under the contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the force majeure event.

### 26. Termination for insolvency

26.1 The purchaser may at any time terminate the contract by giving written notice to the supplier if the supplier becomes bankrupt or otherwise insolvent. In this event, termination will be without compensation to the supplier, provided that such termination will not prejudice or affect any right of action or remedy which has accrued or will accrue thereafter to the purchaser.

### 27. Settlement of disputes

- 27.1 If any dispute or difference of any kind whatsoever arises between the purchaser and the supplier in connection with or arising out of the contract, the parties shall make every effort to resolve amicably such dispute or difference by mutual consultation.
- 27.2 If, after thirty (30) days, the parties have failed to resolve their dispute or difference by such mutual consultation, then either the purchaser or the supplier may give notice to the other party of his intention to commence with mediation. No mediation in respect of this matter may be commenced unless such notice is given to the other party.
- 27.3 Should it not be possible to settle a dispute by means of mediation, it may be settled in a South African court of law.
- 27.4 Mediation proceedings shall be conducted in accordance with the rules of procedure specified in the SCC.
- 27.5 Notwithstanding any reference to mediation and/or court proceedings herein,
- 27.5.1 the parties shall continue to perform their respective obligations under the contract unless they otherwise agree; and
- 27.5.2 the purchaser shall pay the supplier any monies due the supplier.

### 28. Limitation of liability

- 28.1 Except in cases of criminal negligence or wilful misconduct, and in the case of infringement pursuant to Clause 6.
- 28.1.1 the supplier shall not be liable to the purchaser, whether in contract, tort, or otherwise, for any indirect or consequential loss or damage, loss of use, loss of production, or loss of profits or interest costs, provided that this exclusion shall not apply to any obligation of the supplier to pay penalties and/or damages to the purchaser; and
- 28.1.2 the aggregate liability of the supplier to the purchaser, whether under the contract, in tort or otherwise, shall not exceed the total contract price, provided that this limitation shall not apply to the cost of repairing or replacing defective equipment.

### 29. Governing language



29.1 The contract shall be written in English. All correspondence and other documents pertaining to the contract that is exchanged by the parties shall also be written in English.

### 30. Applicable law

30.1 The contract shall be interpreted in accordance with South African laws, unless otherwise specified in SCC.

### 31. Notices

- 31.1 Every written acceptance of a bid shall be posted to the supplier concerned by registered or certified mail and any other notice to him shall be posted by ordinary mail to the address furnished in his bid or to the address notified later by him in writing and such posting shall be deemed to be proper service of such notice
- 31.2 The time mentioned in the contract documents for performing any act after such aforesaid notice has been given, shall be reckoned from the date of posting of such notice.

#### 32. Taxes and duties

- 32.1 A foreign supplier shall be entirely responsible for all taxes, stamp duties, license fees, and other such levies imposed outside the purchaser's country.
- A local supplier shall be entirely responsible for all taxes, duties, license fees, etc., incurred until delivery of the contracted goods to the purchaser.
- 32.3 No contract shall be concluded with any bidder whose tax matters are not in order. Prior to the award of a bid the Department must be in possession of a tax clearance certificate, submitted by the bidder. This certificate must be an original issued by the SARSs.

### 33. National Industrial Participation (NIP) Programme

The NIP Programme administered by the DTI shall be applicable to all contracts that are subject to the NIP obligation.

### 34. Prohibition of restrictive practices

- 34.1 In terms of section 4 (1) (b) (iii) of the Competition Act No. 89 of 1998, as amended, an agreement between, or concerted practice by, firms, or a decision by an association of firms, is prohibited if it is between parties in a horizontal relationship and if a bidder (s) is / are or a contractor(s) was / were involved in collusive bidding (or bid rigging).
- 34.2 If a bidder(s) or contractor(s), based on reasonable grounds or evidence obtained by the purchaser, has / have engaged in the restrictive practice referred to above, the purchaser may refer the matter to the Competition Commission for investigation and possible imposition of administrative penalties as contemplated in the Competition Act No. 89 of 1998.
- 34.3 If a bidder(s) or contractor(s), has / have been found guilty by the Competition Commission of the restrictive practice referred to above, the purchaser may, in addition and without prejudice to any other



remedy provided for, invalidate the bid(s) for such item(s) offered, and / or terminate the contract in whole or part, and / or restrict the bidder(s) or contractor(s) from conducting business with the public sector for a period not exceeding ten (10) years and / or claim damages from the bidder(s) or contractor(s) concerned.

# The above General Conditions of Contract (GCC) are accepted by:

| Name:        |  |
|--------------|--|
| Designation: |  |
| Bidder:      |  |
| Signature:   |  |
| Date:        |  |