

Equipment Required:	Class II Biological Safety Cabinet		
Business Unit	OR Tambo and Chris Hani		
Laboratory Name	Cradock		
Requisition Number			
Requisition Date			
Specification	Detail	Comply	Non-Compliant
Certified	SABS approved cabinet.		
Type	Class II BSC		
Work surface	Stainless steel removal floor pan 4ft		
Display	LCD		
Noise Level	<50 dB		
Front window	Sliding glass		
Light	>1000 lux illumination for work area/ UV light		
Height	Height Adjustable stand		
Installation	Installation and commissioning		
Electricity Plugs	2 Plug points in the cabinet.		
Delivery Period	6 weeks		
Operator Manual	Yes		
Warranty	1 Year		
Service	Third party acceptable		
Sibusiso Thabede		Sign: 	Date: 20/08/2024
Quality Assurance Coordinator		Sign: 	Date: 20/08/2024
Tandiswa Tywaku		Sign: 	Date: 20/08/2024
Laboratory Manager		Sign: 	Date: 20/08/2024